




Quant Time: Jul 02 07:52:34 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 23 10:10:58 2021
Response via : Initial Calibration



TIC: BN015283.D\data.ms

(70) Pentachlorophenol (C)




16.739min (-0.012) 24061.85 ng

response **206015**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.40
264.00	61.60	65.37
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070121\
Data File : BN015283.D
Acq On : 02 Jul 2021 06:07
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 02 07:52:34 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 23 10:10:58 2021
Response via : Initial Calibration



TIC: BN015283.D\data.ms

(77) Benzidine

19.345min (-0.012) 0.00 ng

response 1364779

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.70
183.00	13.20	11.63
0.00	0.00	0.00