

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070121\
 Data File : BN015282.D
 Acq On : 02 Jul 2021 04:19
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

mohammad
 7/2/2021 10:39:24 AM

Quant Time: Jul 02 04:54:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 01 09:06:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.735	152	23801	0.40	ng	# 0.00
7) Naphthalene-d8	10.495	136	103932	0.40	ng	0.00
13) Acenaphthene-d10	14.345	164	53809	0.40	ng	0.00
19) Phenanthrene-d10	17.091	188	100834	0.40	ng	0.00
29) Chrysene-d12	21.291	240	83027	0.40	ng	#-0.01
35) Perylene-d12	23.536	264	71976	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.356	112	25874	0.43	ng	0.00
5) Phenol-d6	6.905	99	31759	0.42	ng	0.00
8) Nitrobenzene-d5	8.860	82	29181	0.50	ng	0.00
11) 2-Methylnaphthalene-d10	12.083	152	67054	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.842	330	6267	0.49	ng	0.00
15) 2-Fluorobiphenyl	12.962	172	85896	0.39	ng	0.00
27) Fluoranthene-d10	19.127	212	112606	0.38	ng	0.00
31) Terphenyl-d14	19.733	244	82395	0.42	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.346	88	4093m	0.37	ng	
3) n-Nitrosodimethylamine	3.696	42	8223	0.39	ng	# 76
6) bis(2-Chloroethyl)ether	7.163	93	31361	0.42	ng	# 84
9) Naphthalene	10.546	128	119325	0.40	ng	97
10) Hexachlorobutadiene	10.838	225	22886	0.41	ng	# 99
12) 2-Methylnaphthalene	12.158	142	77443	0.40	ng	# 89
16) Acenaphthylene	14.066	152	95804	0.40	ng	98
17) Acenaphthene	14.408	154	65503	0.39	ng	99
18) Fluorene	15.398	166	79524	0.39	ng	99
20) 4,6-Dinitro-2-methylph...	15.474	198	325	0.37	ng	# 83
21) 4-Bromophenyl-phenylether	16.287	248	24466	0.38	ng	89
22) Hexachlorobenzene	16.397	284	28506	0.39	ng	# 92
23) Atrazine	16.555	200	14711	0.44	ng	95
24) Pentachlorophenol	16.750	266	3982	0.56	ng	99
25) Phenanthrene	17.127	178	128002	0.39	ng	99
26) Anthracene	17.224	178	105583	0.38	ng	99
28) Fluoranthene	19.157	202	139809	0.40	ng	# 97
30) Pyrene	19.520	202	137355	0.43	ng	99
32) Benzo(a)anthracene	21.270	228	109757	0.39	ng	98
33) Chrysene	21.323	228	122105	0.39	ng	98
34) Bis(2-ethylhexyl)phtha...	21.206	149	40606	0.54	ng	# 87
36) Indeno(1,2,3-cd)pyrene	25.791	276	84244	0.33	ng	# 89
37) Benzo(b)fluoranthene	22.861	252	110601	0.39	ng	# 92
38) Benzo(k)fluoranthene	22.906	252	121106	0.40	ng	# 94
39) Benzo(a)pyrene	23.436	252	99668	0.39	ng	# 80
40) Dibenzo(a,h)anthracene	25.809	278	63063	0.31	ng	# 91
41) Benzo(g,h,i)perylene	26.479	276	74517	0.33	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070121\
Data File : BN015282.D
Acq On : 02 Jul 2021 04:19
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
7/2/2021 10:39:24 AM


Quant Time: Jul 02 04:54:39 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 01 09:06:31 2021
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.735 min Scan# 505
 Delta R.T. 0.003 min
 Lab File: BN015282.D
 Acq: 02 Jul 2021 04:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC




Tgt Ion: 152 Resp: 23801
 Ion Ratio Lower Upper
 152 100
 150 155.3 119.0 178.6
 115 60.3 62.2 93.4#

Manual Integrations
 APPROVED



mohammad
 7/2/2021 10:39:24 AM




#2
 1,4-Dioxane
 Concen: 0.37 ng m
 RT: 3.346 min Scan# 29
 Delta R.T. -0.006 min
 Lab File: BN015282.D
 Acq: 02 Jul 2021 04:19



Tgt Ion: 88 Resp: 4093
 Ion Ratio Lower Upper
 88 100
 43 26.5 12.9 19.3#
 58 75.3 38.1 57.1#









#3
n-Nitrosodimethylamine
Concen: 0.39 ng
RT: 3.696 min Scan# 67
Delta R.T. 0.003 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion: 42 Resp: 8223
Ion Ratio Lower Upper
42 100
74 152.1 149.5 224.3
44 9.1 4.2 6.2#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
7/2/2021 10:39:24 AM



#4
2-Fluorophenol
Concen: 0.43 ng
RT: 5.356 min Scan# 247
Delta R.T. -0.006 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion: 112 Resp: 25874
Ion Ratio Lower Upper
112 100
64 56.1 39.0 58.6
63 28.2 20.6 30.8









#5
Phenol-d6
Concen: 0.42 ng
RT: 6.905 min Scan# 415
Delta R.T. 0.003 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion: 99 Resp: 31759
Ion Ratio Lower Upper
99 100
42 17.1 8.9 13.3#
71 30.2 35.6 53.4#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
7/2/2021 10:39:24 AM



#6
bis(2-Chloroethyl)ether
Concen: 0.42 ng
RT: 7.163 min Scan# 443
Delta R.T. -0.006 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19


Tgt Ion: 93 Resp: 31361
Ion Ratio Lower Upper
93 100
63 75.2 46.6 70.0#
95 32.8 28.6 42.8





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.495 min Scan# 740
Delta R.T. 0.003 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:	136	Resp:	103932
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.8	14.8
54	6.6	5.4	8.2
68	5.5	4.9	7.3




Manual Integrations
APPROVED


mohammad
7/2/2021 10:39:24 AM



#8
Nitrobenzene-d5
Concen: 0.50 ng
RT: 8.860 min Scan# 611
Delta R.T. -0.010 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19


Tgt Ion:	82	Resp:	29181
Ion	Ratio	Lower	Upper
82	100		
128	32.8	31.6	47.4
54	54.9	29.0	43.4





#9
Naphthalene
Concen: 0.40 ng
RT: 10.546 min Scan# 744
Delta R.T. 0.003 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:128 Resp: 119325
Ion Ratio Lower Upper
128 100
129 10.8 9.1 13.7
127 12.9 11.8 17.6

Manual Integrations
APPROVED



mohammad
7/2/2021 10:39:24 AM




#10
Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.838 min Scan# 767
Delta R.T. 0.003 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19



Tgt Ion:225 Resp: 22886
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 50.2 75.2









#11
2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 12.083 min Scan# 949
Delta R.T. -0.005 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion:152 Resp: 67054
Ion Ratio Lower Upper
152 100
151 20.5 16.7 25.1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
7/2/2021 10:39:24 AM




#12
2-Methylnaphthalene
Concen: 0.40 ng
RT: 12.158 min Scan# 966
Delta R.T. -0.001 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion:142 Resp: 77443
Ion Ratio Lower Upper
142 100
141 87.6 72.3 108.5
115 33.5 40.7 61.1#









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.345 min Scan# 1189
Delta R.T. -0.001 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19





Tgt Ion:164 Resp: 53809
Ion Ratio Lower Upper
164 100
162 100.0 83.8 125.6
160 45.2 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 0.49 ng
RT: 15.842 min Scan# 1307
Delta R.T. -0.001 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19




Tgt Ion:330 Resp: 6267
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 36.8 22.4 33.6#




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
7/2/2021 10:39:24 AM




#15
2-Fluorobiphenyl
Concen: 0.39 ng
RT: 12.962 min Scan# 1080
Delta R.T. -0.001 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19





Tgt Ion:172 Resp: 85896
Ion Ratio Lower Upper
172 100
171 34.8 30.6 45.8
170 23.5 20.2 30.4



#16
Acenaphthylene
Concen: 0.40 ng
RT: 14.066 min Scan# 1167
Delta R.T. -0.001 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19




Tgt Ion:152 Resp: 95804
Ion Ratio Lower Upper
152 100
151 19.4 16.7 25.1
153 13.1 10.5 15.7




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED




mohammad
7/2/2021 10:39:24 AM




#17
Acenaphthene
Concen: 0.39 ng
RT: 14.408 min Scan# 1194
Delta R.T. -0.001 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19





Tgt Ion:154 Resp: 65503
Ion Ratio Lower Upper
154 100
153 112.9 90.9 136.3
152 56.1 45.8 68.8



#18
Fluorene
Concen: 0.39 ng
RT: 15.398 min Scan# 1272
Delta R.T. -0.001 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19




Tgt Ion:166 Resp: 79524
Ion Ratio Lower Upper
166 100
165 97.6 79.2 118.8
167 13.5 10.7 16.1



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
7/2/2021 10:39:24 AM







#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.091 min Scan# 1409
Delta R.T. -0.000 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion:188 Resp: 100834
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.2 9.0 13.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
7/2/2021 10:39:24 AM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.37 ng
RT: 15.474 min Scan# 1278
Delta R.T. -0.014 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion:198 Resp: 325
Ion Ratio Lower Upper
198 100
182 32.3 34.5 51.7#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.38 ng
RT: 16.287 min Scan# 1343
Delta R.T. -0.001 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19





Tgt Ion:248 Resp: 24466
Ion Ratio Lower Upper
248 100
250 96.5 86.2 129.4
141 52.7 36.2 54.2



#22
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.397 min Scan# 1352
Delta R.T. -0.013 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19




Tgt Ion:284 Resp: 28506
Ion Ratio Lower Upper
284 100
142 34.9 22.2 33.2#
249 31.0 26.6 39.8



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Manual Integrations
APPROVED

mohammad
7/2/2021 10:39:24 AM



#23
Atrazine
Concen: 0.44 ng
RT: 16.555 min Scan# 1365
Delta R.T. -0.001 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:200 Resp: 14711
Ion Ratio Lower Upper
200 100
173 27.0 21.7 32.5
215 48.5 43.1 64.7

Manual Integrations
APPROVED



mohammad
7/2/2021 10:39:24 AM




#24
Pentachlorophenol
Concen: 0.56 ng
RT: 16.750 min Scan# 1381
Delta R.T. -0.001 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19




Tgt Ion:266 Resp: 3982
Ion Ratio Lower Upper
266 100
264 63.7 51.0 76.4
268 64.3 50.6 76.0





#25
Phenanthrene
Concen: 0.39 ng
RT: 17.127 min Scan# 1412
Delta R.T. -0.001 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:178 Resp: 128002
Ion Ratio Lower Upper
178 100
176 18.7 15.6 23.4
179 15.4 12.1 18.1




Manual Integrations
APPROVED


mohammad
7/2/2021 10:39:24 AM




#26
Anthracene
Concen: 0.38 ng
RT: 17.224 min Scan# 1420
Delta R.T. -0.001 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion:178 Resp: 105583
Ion Ratio Lower Upper
178 100
176 18.2 15.1 22.7
179 15.3 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.38 ng
RT: 19.127 min Scan# 1709
Delta R.T. -0.006 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19





Tgt Ion:212 Resp: 112606
Ion Ratio Lower Upper
212 100
106 14.1 7.0 10.6#
104 7.7 4.1 6.1#



#28
Fluoranthene
Concen: 0.40 ng
RT: 19.157 min Scan# 1715
Delta R.T. -0.006 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19




Tgt Ion:202 Resp: 139809
Ion Ratio Lower Upper
202 100
101 11.8 6.6 10.0#
203 17.4 13.9 20.9



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
7/2/2021 10:39:24 AM







#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.291 min Scan# 2007
Delta R.T. -0.012 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion:240 Resp: 83027
Ion Ratio Lower Upper
240 100
120 7.7 4.2 6.4#
236 25.5 21.4 32.2

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
7/2/2021 10:39:24 AM



#30
Pyrene
Concen: 0.43 ng
RT: 19.520 min Scan# 1788
Delta R.T. -0.010 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19


Tgt Ion:202 Resp: 137355
Ion Ratio Lower Upper
202 100
200 20.5 16.9 25.3
203 17.4 13.9 20.9





#31
Terphenyl-d14
Concen: 0.42 ng
RT: 19.733 min Scan# 1831
Delta R.T. -0.006 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:244 Resp: 82395
Ion Ratio Lower Upper
244 100
212 7.7 6.3 9.5
122 12.4 6.2 9.2#




Manual Integrations
APPROVED


mohammad
7/2/2021 10:39:24 AM



#32
Benzo(a)anthracene
Concen: 0.39 ng
RT: 21.270 min Scan# 2005
Delta R.T. -0.011 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion:228 Resp: 109757
Ion Ratio Lower Upper
228 100
226 27.1 23.0 34.6
229 19.4 15.7 23.5









#33
Chrysene
Concen: 0.39 ng
RT: 21.323 min Scan# 2010
Delta R.T. -0.012 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion:	228	Resp:	122105
Ion Ratio	Lower	Upper	
228	100		
226	29.7	25.0	37.6
229	19.2	15.9	23.9

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
7/2/2021 10:39:24 AM



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.54 ng
RT: 21.206 min Scan# 1999
Delta R.T. -0.022 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19


Tgt Ion:	149	Resp:	40606
Ion Ratio	Lower	Upper	
149	100		
167	27.7	28.9	43.3#
279	3.3	4.9	7.3#





#35
Perylene-d12
Concen: 0.40 ng
RT: 23.536 min Scan# 2442
Delta R.T. -0.014 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:264 Resp: 71976
Ion Ratio Lower Upper
264 100
260 25.0 20.6 31.0
265 23.0 21.4 32.2

Manual Integrations
APPROVED



mohammad
7/2/2021 10:39:24 AM




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.33 ng
RT: 25.791 min Scan# 3101
Delta R.T. -0.021 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19



Tgt Ion:276 Resp: 84244
Ion Ratio Lower Upper
276 100
138 24.8 10.0 15.0#
277 24.9 20.4 30.6









#37
Benzo(b)fluoranthene
Concen: 0.39 ng
RT: 22.861 min Scan# 2245
Delta R.T. -0.015 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion:252 Resp: 110601
Ion Ratio Lower Upper
252 100
253 22.6 18.5 27.7
125 17.7 5.8 8.6#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
7/2/2021 10:39:24 AM



#38
Benzo(k)fluoranthene
Concen: 0.40 ng
RT: 22.906 min Scan# 2258
Delta R.T. -0.015 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion:252 Resp: 121106
Ion Ratio Lower Upper
252 100
253 21.9 19.1 28.7
125 11.1 5.4 8.0#









#39
Benzo(a)pyrene
Concen: 0.39 ng
RT: 23.436 min Scan# 2413
Delta R.T. -0.018 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion:252 Resp: 99668
Ion Ratio Lower Upper
252 100
253 24.7 19.4 29.2
125 35.4 7.4 11.2#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Manual Integrations
APPROVED


mohammad
7/2/2021 10:39:24 AM




#40
Dibenzo(a,h)anthracene
Concen: 0.31 ng
RT: 25.809 min Scan# 3106
Delta R.T. -0.018 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19

Tgt Ion:278 Resp: 63063
Ion Ratio Lower Upper
278 100
139 19.0 9.0 13.4#
279 23.8 21.0 31.4







#41
Benzo(g,h,i)perylene
Concen: 0.33 ng
RT: 26.479 min Scan# 3302
Delta R.T. -0.018 min
Lab File: BN015282.D
Acq: 02 Jul 2021 04:19



Tgt Ion: 276 Resp: 74517
Ion Ratio Lower Upper
276 100
277 23.9 19.7 29.5
138 23.9 9.8 14.6#



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

mohammad
7/2/2021 10:39:24 AM