

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070621\
 Data File : BN015361.D
 Acq On : 07 Jul 2021 06:02
 Operator : CG/JU
 Sample : M2907-08
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-02-04-09-063021

Quant Time: Jul 07 07:04:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 01 09:06:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.735	152	15660	0.400	ng	# 0.00
7) Naphthalene-d8	10.495	136	73126	0.400	ng	0.00
13) Acenaphthene-d10	14.345	164	43154	0.400	ng	0.00
19) Phenanthrene-d10	17.091	188	89728	0.400	ng	0.00
29) Chrysene-d12	21.291	240	80306	0.400	ng	#-0.01
35) Perylene-d12	23.539	264	42705	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.365	112	6960	0.174	ng	0.00
5) Phenol-d6	6.905	99	5141	0.102	ng	0.00
8) Nitrobenzene-d5	8.873	82	18892	0.461	ng	0.00
11) 2-Methylnaphthalene-d10	12.087	152	43065	0.356	ng	0.00
14) 2,4,6-Tribromophenol	15.842	330	6201	0.553	ng	0.00
15) 2-Fluorobiphenyl	12.962	172	58687	0.335	ng	0.00
27) Fluoranthene-d10	19.128	212	111036	0.419	ng	0.00
31) Terphenyl-d14	19.733	244	86471	0.460	ng	0.00
Target Compounds						
24) Pentachlorophenol	16.738	266	3473	0.555	ng	# 36
34) Bis(2-ethylhexyl)phtha...	21.217	149	17534	0.328	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070621\
Data File : BN015361.D
Acq On : 07 Jul 2021 06:02
Operator : CG/JU
Sample : M2907-08
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-02-04-09-063021

Quant Time: Jul 07 07:04:26 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 01 09:06:31 2021
Response via : Initial Calibration

