

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070621\
 Data File : BN015362.D
 Acq On : 07 Jul 2021 06:38
 Operator : CG/JU
 Sample : M2923-06
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EB-01-070121

Quant Time: Jul 07 07:35:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 01 09:06:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.734	152	16250	0.400	ng	# 0.00
7) Naphthalene-d8	10.495	136	76065	0.400	ng	0.00
13) Acenaphthene-d10	14.345	164	44678	0.400	ng	0.00
19) Phenanthrene-d10	17.090	188	89312	0.400	ng	0.00
29) Chrysene-d12	21.291	240	81167	0.400	ng	#-0.01
35) Perylene-d12	23.543	264	44595	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.365	112	7569	0.183	ng	0.00
5) Phenol-d6	6.905	99	5573	0.107	ng	0.00
8) Nitrobenzene-d5	8.873	82	19728	0.463	ng	0.00
11) 2-Methylnaphthalene-d10	12.087	152	44645	0.354	ng	0.00
14) 2,4,6-Tribromophenol	15.842	330	6099	0.536	ng	0.00
15) 2-Fluorobiphenyl	12.962	172	60799	0.335	ng	0.00
27) Fluoranthene-d10	19.132	212	112260	0.426	ng	0.00
31) Terphenyl-d14	19.733	244	88238	0.465	ng	0.00
Target Compounds						
20) 4,6-Dinitro-2-methylph...	15.487	198	153	0.286	ng	# 84
24) Pentachlorophenol	16.750	266	131	0.255	ng	94
25) Phenanthrene	17.139	178	6312	0.022	ng	97
34) Bis(2-ethylhexyl)phtha...	21.217	149	202236	1.824	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070621\
Data File : BN015362.D
Acq On : 07 Jul 2021 06:38
Operator : CG/JU
Sample : M2923-06
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EB-01-070121

Quant Time: Jul 07 07:35:58 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 01 09:06:31 2021
Response via : Initial Calibration

