

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070621\
 Data File : BN015364.D
 Acq On : 07 Jul 2021 07:50
 Operator : CG/JU
 Sample : M2923-01
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-05-03-08-070121

Quant Time: Jul 07 08:27:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 01 09:06:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.734	152	14684	0.400	ng	# 0.00
7) Naphthalene-d8	10.495	136	68143	0.400	ng	0.00
13) Acenaphthene-d10	14.345	164	39391	0.400	ng	0.00
19) Phenanthrene-d10	17.090	188	81135	0.400	ng	0.00
29) Chrysene-d12	21.291	240	74543	0.400	ng	#-0.01
35) Perylene-d12	23.539	264	39924	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.365	112	7009	0.187	ng	0.00
5) Phenol-d6	6.904	99	5332	0.113	ng	0.00
8) Nitrobenzene-d5	8.873	82	18398	0.482	ng	0.00
11) 2-Methylnaphthalene-d10	12.087	152	43126	0.382	ng	0.00
14) 2,4,6-Tribromophenol	15.842	330	6440	0.599	ng	0.00
15) 2-Fluorobiphenyl	12.962	172	59899	0.374	ng	0.00
27) Fluoranthene-d10	19.132	212	111278	0.464	ng	0.00
31) Terphenyl-d14	19.733	244	91619	0.525	ng	0.00
Target Compounds						Qvalue
24) Pentachlorophenol	16.750	266	173	0.260	ng	86
25) Phenanthrene	17.127	178	8014	0.031	ng	# 95
28) Fluoranthene	19.162	202	7404	0.026	ng	# 87
34) Bis(2-ethylhexyl)phtha...	21.217	149	132876	1.415	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN070621\
Data File : BN015364.D
Acq On : 07 Jul 2021 07:50
Operator : CG/JU
Sample : M2923-01
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-05-03-08-070121

Quant Time: Jul 07 08:27:32 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 01 09:06:31 2021
Response via : Initial Calibration

