

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071222\
 Data File : BN020712.D
 Acq On : 12 Jul 2022 12:08
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 13 02:41:07 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 02 03:10:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	143	0.00
2	1,4-Dioxane	0.400	0.382	4.5	142	0.00
3	n-Nitrosodimethylamine	0.400	0.313	21.8	119	0.00
4 S	2-Fluorophenol	0.400	0.302	24.5	111	0.00
5 S	Phenol-d6	0.400	0.317	20.8	118	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.337	15.8	127	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	142	-0.01
8 S	Nitrobenzene-d5	0.400	0.342	14.5	133	-0.01
9	Naphthalene	0.400	0.387	3.3	143	-0.01
10	Hexachlorobutadiene	0.400	0.371	7.3	136	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.362	9.5	136	0.00
12	2-Methylnaphthalene	0.400	0.366	8.5	135	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	134	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.271	32.3#	105	0.00
15 S	2-Fluorobiphenyl	0.400	0.390	2.5	133	-0.01
16	Acenaphthylene	0.400	0.353	11.8	129	-0.01
17	Acenaphthene	0.400	0.380	5.0	132	-0.01
18	Fluorene	0.400	0.363	9.3	128	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	126	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.378	5.5	100	0.00
21	4-Bromophenyl-phenylether	0.400	0.390	2.5	133	-0.01
22	Hexachlorobenzene	0.400	0.407	-1.7	138	0.00
23	Atrazine	0.400	0.298	25.5#	104	0.00
24	Pentachlorophenol	0.400	0.428	-7.0	141	-0.01
25	Phenanthrene	0.400	0.368	8.0	124	-0.01
26	Anthracene	0.400	0.329	17.8	115	0.00
27 SURR	Fluoranthene-d10	0.400	0.338	15.5	116	0.00
28	Fluoranthene	0.400	0.350	12.5	120	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	97	0.00
30	Pyrene	0.400	0.464	-16.0	115	0.00
31 S	Terphenyl-d14	0.400	0.437	-9.2	108	0.00
32	Benzo(a)anthracene	0.400	0.328	18.0	84	0.00
33	Chrysene	0.400	0.405	-1.3	102	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.356	11.0	97	0.00
35 I	Perylene-d12	0.400	0.400	0.0	73	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.340	15.0	65	0.00
37	Benzo(b)fluoranthene	0.400	0.413	-3.2	79	0.00
38	Benzo(k)fluoranthene	0.400	0.402	-0.5	79	0.00
39 C	Benzo(a)pyrene	0.400	0.368	8.0	71	0.00
40	Dibenzo(a,h)anthracene	0.400	0.341	14.8	66	0.00
41	Benzo(g,h,i)perylene	0.400	0.373	6.8	70	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4