

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071422\
 Data File : BN020744.D
 Acq On : 14 Jul 2022 19:45
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 15 02:51:32 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN062622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 02 03:10:12 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	89	0.00
2	1,4-Dioxane	0.400	0.337	15.8	78	0.00
3	n-Nitrosodimethylamine	0.400	0.369	7.8	88	0.00
4 S	2-Fluorophenol	0.400	0.314	21.5	72	0.00
5 S	Phenol-d6	0.400	0.307	23.3	72	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.314	21.5	74	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	97	-0.01
8 S	Nitrobenzene-d5	0.400	0.345	13.8	91	-0.01
9	Naphthalene	0.400	0.382	4.5	96	-0.01
10	Hexachlorobutadiene	0.400	0.367	8.3	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.373	6.8	95	0.00
12	2-Methylnaphthalene	0.400	0.382	4.5	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.400	0.375	6.3	102	0.00
15 S	2-Fluorobiphenyl	0.400	0.347	13.3	83	-0.01
16	Acenaphthylene	0.400	0.342	14.5	88	-0.01
17	Acenaphthene	0.400	0.362	9.5	88	-0.01
18	Fluorene	0.400	0.425	-6.2	105	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	111	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.419	-4.7	111	0.00
21	4-Bromophenyl-phenylether	0.400	0.370	7.5	110	-0.01
22	Hexachlorobenzene	0.400	0.372	7.0	111	0.00
23	Atrazine	0.400	0.317	20.8	98	0.00
24	Pentachlorophenol	0.400	0.480	-20.0	151	-0.01
25	Phenanthrene	0.400	0.364	9.0	108	0.00
26	Anthracene	0.400	0.345	13.8	106	0.00
27 SURR	Fluoranthene-d10	0.400	0.376	6.0	113	0.00
28	Fluoranthene	0.400	0.388	3.0	117	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	116	0.00
30	Pyrene	0.400	0.386	3.5	115	0.00
31 S	Terphenyl-d14	0.400	0.377	5.8	112	0.00
32	Benzo(a)anthracene	0.400	0.352	12.0	109	0.00
33	Chrysene	0.400	0.396	1.0	120	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.337	15.8	110	0.00
35 I	Perylene-d12	0.400	0.400	0.0	108	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.355	11.3	101	0.00
37	Benzo(b)fluoranthene	0.400	0.399	0.3	113	0.00
38	Benzo(k)fluoranthene	0.400	0.396	1.0	114	0.00
39 C	Benzo(a)pyrene	0.400	0.377	5.8	107	0.00
40	Dibenzo(a,h)anthracene	0.400	0.344	14.0	99	-0.02
41	Benzo(g,h,i)perylene	0.400	0.356	11.0	99	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4