

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071524\
 Data File : BN032720.D
 Acq On : 15 Jul 2024 12:29
 Operator : MA/JU
 Sample : P3222-05
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-196-GW-940-942

Quant Time: Jul 15 13:03:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 10 23:45:48 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/16/2024
 Supervised By :mohammad ahmed 07/16/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.559	152	3267	0.400	ng	0.00
7) Naphthalene-d8	10.325	136	10022	0.400	ng	#-0.03
13) Acenaphthene-d10	14.199	164	5413	0.400	ng	-0.01
19) Phenanthrene-d10	16.966	188	10431	0.400	ng	-0.01
29) Chrysene-d12	21.175	240	9866	0.400	ng	-0.02
35) Perylene-d12	23.373	264	11503	0.400	ng	-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.190	112	2292	0.277	ng	0.00
5) Phenol-d6	6.801	99	1494	0.146	ng	0.00
8) Nitrobenzene-d5	8.755	82	2848	0.377	ng	-0.02
11) 2-Methylnaphthalene-d10	11.922	152	5434	0.354	ng	-0.04
14) 2,4,6-Tribromophenol	15.725	330	801	0.346	ng	-0.01
15) 2-Fluorobiphenyl	12.820	172	8372	0.416	ng	-0.03
27) Fluoranthene-d10	19.003	212	11082	0.411	ng	-0.02
31) Terphenyl-d14	19.606	244	9192	0.455	ng	-0.02
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.067	149	4862m	0.282	ng	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071524\
Data File : BN032720.D
Acq On : 15 Jul 2024 12:29
Operator : MA/JU
Sample : P3222-05
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-196-GW-940-942

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/16/2024
Supervised By :mohammad ahmed 07/16/2024

Quant Time: Jul 15 13:03:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 10 23:45:48 2024
Response via : Initial Calibration

