

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072021\
 Data File : BN015617.D
 Acq On : 20 Jul 2021 10:06
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 20 11:24:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 20 11:13:03 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	0.00
2	1,4-Dioxane	0.400	0.413	-3.2	110	0.00
3	n-Nitrosodimethylamine	0.400	0.362	9.5	102	0.00
4 S	2-Fluorophenol	0.400	0.384	4.0	102	0.00
5 S	Phenol-d6	0.400	0.379	5.3	101	0.00
6	bis(2-Chloroethyl)ether	0.400	0.382	4.5	103	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	105	0.00
8 S	Nitrobenzene-d5	0.400	0.340	15.0	101	0.00
9	Naphthalene	0.400	0.367	8.3	100	0.00
10	Hexachlorobutadiene	0.400	0.372	7.0	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.363	9.3	102	0.00
12	2-Methylnaphthalene	0.400	0.375	6.3	101	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.400	0.405	-1.3	123	0.00
15 S	2-Fluorobiphenyl	0.400	0.370	7.5	104	0.00
16	Acenaphthylene	0.400	0.343	14.2	102	0.00
17	Acenaphthene	0.400	0.364	9.0	103	0.00
18	Fluorene	0.400	0.366	8.5	103	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	111	0.00
20	4,6-Dinitro-2-methylphenol	0.400	1.138	-184.5#	1410	0.00
21	4-Bromophenyl-phenylether	0.400	0.351	12.3	105	0.00
22	Hexachlorobenzene	0.400	0.359	10.3	104	0.00
23	Atrazine	0.400	0.354	11.5	108	0.00
24	Pentachlorophenol	0.400	0.367	8.3	74	0.00
25	Phenanthrene	0.400	0.361	9.8	104	0.00
26	Anthracene	0.400	0.344	14.0	105	0.00
27 SURR	Fluoranthene-d10	0.400	0.351	12.3	109	0.00
28	Fluoranthene	0.400	0.371	7.3	108	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	119	0.00
30	Pyrene	0.400	0.345	13.8	108	0.00
31 S	Terphenyl-d14	0.400	0.352	12.0	110	0.00
32	Benzo(a)anthracene	0.400	0.358	10.5	114	0.00
33	Chrysene	0.400	0.366	8.5	113	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.335	16.3	109	0.00
35 I	Perylene-d12	0.400	0.400	0.0	133	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.451	-12.7	168	0.00
37	Benzo(b)fluoranthene	0.400	0.350	12.5	123	0.00
38	Benzo(k)fluoranthene	0.400	0.370	7.5	125	0.00
39 C	Benzo(a)pyrene	0.400	0.372	7.0	133	0.00
40	Dibenzo(a,h)anthracene	0.400	0.473	-18.2	178	0.00
41	Benzo(g,h,i)perylene	0.400	0.440	-10.0	165	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4