

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080124\
 Data File : BN033144.D
 Acq On : 31 Jul 2024 23:16
 Operator : MA/JU
 Sample : P3359-05
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E1GD5

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/01/2024
 Supervised By :mohammad ahmed 08/13/2024

Quant Time: Aug 01 00:19:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN072524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 30 14:49:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.495	152	2842	0.400	ng/ul	-0.05
4) Naphthalene-d8	10.240	136	9732	0.400	ng/ul	-0.08
9) Acenaphthene-d10	14.134	164	6140	0.400	ng/ul	#-0.03
13) Phenanthrene-d10	16.899	188	13673m	0.400	ng/ul	-0.04
17) Chrysene-d12	21.151	240	11199	0.400	ng/ul	0.00
23) Perylene-d12	23.317	264	5803m	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.052	96	14969	4.164	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.890	152	3834	0.241	ng/ul	-0.04
18) Fluoranthene-d10	18.953	212	12047	0.380	ng/ul	-0.02
Target Compounds						
5) Naphthalene	10.295	128	995	0.040	ng/ul#	Qvalue 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080124\
Data File : BN033144.D
Acq On : 31 Jul 2024 23:16
Operator : MA/JU
Sample : P3359-05
Misc :
ALS Vial : 21 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
E1GD5

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/01/2024
Supervised By :mohammad ahmed 08/13/2024


Quant Time: Aug 01 00:19:36 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN072524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 14:49:49 2024
Response via : Initial Calibration





#5
Naphthalene
Concen: 0.040 ng/ul
RT: 10.295 min Scan# 11
Delta R.T. -0.082 min
Lab File: BN033144.D
Acq: 31 Jul 2024 23:16

Instrument :
BNA_N
ClientSampleId :
E1GD5



Tgt Ion:128 Resp: 99
Ion Ratio Lower Upper
128 100
129 28.3 12.0 18.0
127 22.2 12.7 19.1

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/01/2024
Supervised By :mohammad ahmed 08/13/2024

