

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080124\
 Data File : BN033147.D
 Acq On : 01 Aug 2024 01:05
 Operator : MA/JU
 Sample : P3378-07
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0B06

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/01/2024
 Supervised By :mohammad ahmed 08/13/2024

Quant Time: Aug 01 01:58:31 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN072524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 30 14:49:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.491	152	2454	0.400	ng/ul	-0.06
4) Naphthalene-d8	10.245	136	8207	0.400	ng/ul	-0.07
9) Acenaphthene-d10	14.129	164	5300	0.400	ng/ul	#-0.03
13) Phenanthrene-d10	16.899	188	11503m	0.400	ng/ul	-0.04
17) Chrysene-d12	21.160	240	9967	0.400	ng/ul	0.00
23) Perylene-d12	23.323	264	4830m	0.400	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.052	96	14507	4.673	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.884	152	3850	0.287	ng/ul	-0.04
18) Fluoranthene-d10	18.953	212	10143	0.359	ng/ul	-0.02
Target Compounds						Qvalue
5) Naphthalene	10.295	128	1411	0.067	ng/ul#	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080124\
Data File : BN033147.D
Acq On : 01 Aug 2024 01:05
Operator : MA/JU
Sample : P3378-07
Misc :
ALS Vial : 24 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
C0B06

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/01/2024
Supervised By :mohammad ahmed 08/13/2024


Quant Time: Aug 01 01:58:31 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN072524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 14:49:49 2024
Response via : Initial Calibration





#5
Naphthalene
Concen: 0.067 ng/ul
RT: 10.295 min Scan# 1756
Delta R.T. -0.082 min
Lab File: BN033147.D
Acq: 01 Aug 2024 01:05

Instrument :
BNA_N
ClientSampleId :
C0B06



Tgt Ion:128 Resp: 141
Ion Ratio Lower Upper
128 100
129 23.9 12.0 18.0
127 20.0 12.7 19.1

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/01/2024
Supervised By :mohammad ahmed 08/13/2024

