


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080124\
Data File : BN033123.D
Acq On : 31 Jul 2024 08:44
Operator : MA/JU
Sample : DFTPP228
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP228

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN072524.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Jul 30 14:49:49 2024



AutoFind: Scans 2452, 2453, 2454; Background Corrected with Scan 2445

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.7	47968	PASS
68	69	0.00	2	2.0	1065	PASS
69	198	0.00	100	42.9	54568	PASS
70	69	0.00	2	0.7	384	PASS
127	198	10	80	52.4	66544	PASS
197	198	0.00	2	0.2	225	PASS
198	198	100	100	100.0	127112	PASS
199	198	5	9	6.7	8507	PASS
275	198	10	60	26.4	33613	PASS
365	198	1	100	4.0	5040	PASS
441	198	0.01	100	9.2	11685	PASS
442	198	50	100	55.1	70051	PASS
443	442	15	24	20.9	14652	PASS

m/z	Abundance
39.90	461.0
41.00	198.0
43.00	189.0
44.00	1818.0
48.90	158.0
55.00	171.0
57.10	163.0
73.00	655.0
77.00	199.0
95.90	275.0
132.90	203.0
147.00	200.0
190.90	304.0
193.00	161.0
207.00	1664.0
208.00	428.0
209.00	221.0
281.00	413.0

Instrument :
BNA_N
ClientSampleId :
DFTPP228

m/z	Abundance
36.90	164.0
38.00	688.0
39.00	3463.0
39.90	661.0
41.00	327.0
43.10	277.0
44.00	1309.0
44.90	254.0
49.00	601.0
50.00	14221.0
51.00	52968.0
52.00	3021.0
54.90	610.0
56.00	1646.0
57.00	3678.0
57.90	188.0
61.00	773.0
61.90	824.0
63.00	2393.0
64.00	285.0
65.00	1160.0
67.00	177.0
67.90	1038.0
68.90	59264.0
70.00	443.0
71.00	184.0
73.00	1122.0
74.00	6562.0
75.00	9780.0
76.00	3293.0
77.00	66368.0
78.00	4891.0
79.00	4740.0
79.90	3612.0
81.00	4988.0
81.90	1358.0
83.00	1245.0
83.90	300.0
85.00	894.0
86.00	1607.0
87.00	707.0
87.90	207.0
88.90	202.0
90.90	1084.0
91.90	1286.0
93.00	8357.0
93.90	627.0
94.90	208.0
96.00	528.0
97.00	339.0
98.00	7279.0
99.00	5331.0
99.90	496.0
100.90	2727.0
101.90	174.0
102.90	1014.0
103.90	1806.0
104.90	1958.0
105.90	579.0
107.00	19632.0
108.00	3453.0
108.90	642.0
110.00	37184.0
111.00	5941.0
111.90	585.0
112.90	221.0
114.90	242.0
116.00	1108.0
117.00	21000.0
117.90	1673.0
119.00	267.0
119.90	348.0
121.90	1446.0
122.90	2262.0
123.90	1063.0
124.90	978.0
127.00	68168.0
128.00	5504.0
128.90	27328.0
130.00	2624.0
130.90	561.0
132.00	312.0
132.90	311.0
133.90	752.0
135.00	2231.0
136.00	1021.0
137.00	1521.0
138.00	318.0
140.00	399.0
140.90	3509.0
141.90	1166.0
142.90	806.0

Instrument :
BNA_N
ClientSampleId :
DFTPP228

143.90	302.0
144.90	244.0
145.90	707.0
147.00	2218.0
147.90	4981.0
148.90	1124.0
150.10	300.0
150.90	524.0
151.60	269.0
153.00	1056.0
153.90	832.0
155.00	2048.0
156.00	3010.0
157.00	633.0
157.90	576.0
159.00	516.0
159.90	1302.0
161.00	1661.0
161.90	527.0
162.90	249.0
164.90	1521.0
165.90	1079.0
166.90	7652.0
168.00	4866.0
169.00	592.0
170.00	227.0
170.90	302.0
172.00	703.0
172.90	817.0
174.00	1665.0
175.00	2776.0
175.90	901.0
176.90	1226.0
177.80	376.0
179.00	5268.0
180.00	3648.0
180.90	1450.0
181.90	224.0
184.00	424.0
184.90	2343.0
186.00	17576.0
187.00	5432.0
188.00	722.0
189.00	1215.0
190.00	212.0
190.90	809.0
192.00	1722.0
193.00	2110.0
193.90	443.0
195.00	254.0
196.00	3740.0
196.90	674.0
197.90	123552.0
198.90	8412.0
199.90	692.0
201.40	515.0
203.00	1188.0
204.00	4695.0
205.00	8213.0
206.00	34272.0
207.00	5996.0
208.00	1732.0
208.90	774.0
210.90	1480.0
211.70	364.0
214.90	359.0
215.90	913.0
216.90	8890.0
218.00	1259.0
218.80	181.0
219.70	166.0
221.00	3907.0
221.90	1284.0
223.00	1966.0
224.00	17264.0
225.00	4670.0
225.90	699.0
226.90	8533.0
228.00	1090.0
228.90	1713.0
229.90	287.0
230.90	757.0
233.80	579.0
234.90	614.0
235.90	366.0
236.90	962.0
238.90	257.0
239.90	352.0
240.90	410.0
242.00	950.0
243.00	804.0
244.00	13971.0
245.00	1873.0
245.90	3088.0
246.90	622.0
248.90	487.0

Instrument :
BNA_N
ClientSampleId :
DFTPP228

250.90	191.0
251.90	208.0
252.90	412.0
254.00	468.0
254.90	67968.0
256.00	11346.0
257.00	963.0
257.90	4403.0
258.90	735.0
265.00	1826.0
265.90	285.0
270.80	212.0
272.00	214.0
272.90	1881.0
274.00	5582.0
275.00	29872.0
276.00	4292.0
277.00	2814.0
277.90	676.0
281.00	431.0
282.90	351.0
283.90	181.0
285.00	468.0
292.90	764.0
293.90	190.0
295.90	9365.0
296.90	1220.0
303.00	879.0
304.00	276.0
307.90	157.0
314.00	390.0
314.90	1162.0
315.90	462.0
320.90	308.0
323.00	2469.0
323.90	472.0
326.90	669.0
327.80	317.0
331.90	267.0
332.90	291.0
334.00	1658.0
334.90	422.0
340.80	337.0
345.90	477.0
352.00	659.0
352.90	570.0
354.00	727.0
354.90	284.0
365.00	3845.0
365.90	605.0
370.90	248.0
371.90	1234.0
372.90	384.0
382.90	314.0
401.90	452.0
403.00	709.0
403.80	233.0
421.00	512.0
421.90	517.0
422.90	3477.0
423.80	1098.0
425.00	205.0
441.00	7363.0
442.00	55400.0
443.00	10639.0
444.00	939.0

Instrument :
BNA_N
ClientSampleId :
DFTPP228

m/z	Abundance
36.90	272.0
38.00	685.0
39.00	4016.0
39.90	453.0
41.00	365.0
43.00	203.0
44.00	1540.0
44.90	225.0
49.00	532.0
50.00	15155.0
51.00	57112.0
52.00	3163.0
53.00	193.0
55.00	484.0
56.00	1889.0
57.00	4221.0
58.10	191.0
61.00	734.0
62.00	926.0
63.00	2341.0
63.90	350.0
65.00	1283.0
67.00	174.0
67.90	1234.0
69.00	64496.0
69.90	438.0
73.00	1192.0
74.00	6999.0
75.00	11038.0
76.00	3847.0
77.00	75744.0
78.00	5432.0
79.00	5571.0
79.90	3891.0
80.90	5742.0
81.90	1444.0
83.00	1328.0
83.90	283.0
85.00	1007.0
85.90	1709.0
87.00	741.0
87.80	283.0
89.00	180.0
90.90	1263.0
92.00	1449.0
93.00	9216.0
94.00	601.0
95.00	274.0
95.90	603.0
98.00	7622.0
98.90	5665.0
99.90	487.0
101.00	2910.0
102.00	207.0
103.00	975.0
103.90	1969.0
104.90	2195.0
106.00	611.0
107.00	24704.0
108.00	3819.0
109.00	695.0
110.00	41040.0
111.00	6741.0
111.90	891.0
112.90	308.0
114.90	160.0
115.90	1202.0
117.00	24512.0
118.00	1742.0
119.00	283.0
119.90	194.0
120.80	189.0
122.00	1688.0
123.00	2440.0
123.90	1259.0
124.90	1076.0
126.00	379.0
127.00	79384.0
128.00	5818.0
128.90	32880.0
130.00	2682.0
130.90	701.0
132.00	264.0
133.00	465.0
133.90	1003.0
135.00	2590.0
135.90	1111.0
137.00	1744.0
137.90	265.0
139.90	422.0
141.00	4473.0
141.90	1274.0

Instrument :
BNA_N
ClientSampleId :
DFTPP228

142.90	929.0
143.90	208.0
145.00	236.0
146.00	827.0
147.00	2368.0
147.90	5939.0
149.00	1336.0
149.90	326.0
150.90	526.0
151.90	503.0
152.90	1446.0
153.90	918.0
154.90	2377.0
156.00	3609.0
156.90	606.0
158.00	911.0
158.90	639.0
160.00	1308.0
161.00	1913.0
161.90	578.0
162.90	189.0
163.90	252.0
165.00	1564.0
166.00	1179.0
167.00	8707.0
167.90	5564.0
169.00	684.0
169.90	275.0
170.90	366.0
171.90	911.0
173.00	892.0
174.00	1808.0
175.00	3392.0
175.90	894.0
177.00	1237.0
177.80	584.0
178.90	6766.0
180.00	4839.0
180.90	2012.0
182.00	380.0
182.80	171.0
183.90	582.0
185.00	2938.0
186.00	22872.0
187.00	6895.0
188.00	717.0
189.00	1872.0
190.00	316.0
190.90	814.0
192.00	2419.0
193.00	2465.0
193.90	533.0
194.90	211.0
196.00	4815.0
197.90	152512.0
198.90	10049.0
200.00	911.0
201.40	749.0
202.90	1360.0
204.00	6612.0
205.00	11116.0
206.00	45088.0
207.00	7079.0
207.90	2056.0
209.00	709.0
210.00	856.0
211.00	1760.0
211.70	359.0
215.00	473.0
216.00	1007.0
216.90	12083.0
217.90	1568.0
218.90	159.0
220.90	5644.0
222.00	1708.0
222.90	2643.0
224.00	24104.0
225.00	6253.0
225.90	630.0
227.00	11552.0
227.90	1399.0
229.00	2269.0
229.90	373.0
231.00	720.0
232.90	204.0
233.90	790.0
234.90	844.0
235.90	522.0
237.00	1038.0
237.90	247.0
239.00	394.0
240.00	423.0
240.90	635.0
241.90	1280.0
243.00	1375.0
244.00	18704.0

Instrument :
BNA_N
ClientSampleId :
DFTPP228

245.00	2601.0
245.90	4078.0
246.90	874.0
247.90	247.0
249.00	789.0
251.00	217.0
251.90	171.0
252.90	371.0
254.00	718.0
255.00	96088.0
256.00	14522.0
257.00	1324.0
258.00	6339.0
258.80	861.0
264.00	229.0
264.90	2895.0
265.90	411.0
269.90	175.0
270.90	251.0
272.10	276.0
272.90	3204.0
274.00	7433.0
275.00	41536.0
276.00	5954.0
276.90	4035.0
277.90	753.0
278.90	196.0
281.00	374.0
281.90	227.0
282.90	497.0
283.80	333.0
285.00	716.0
286.00	188.0
288.90	159.0
291.80	188.0
292.90	915.0
293.90	271.0
295.00	335.0
296.00	13805.0
297.00	2041.0
301.90	242.0
303.00	1365.0
303.90	405.0
307.90	239.0
310.00	155.0
313.90	640.0
314.90	1652.0
316.00	750.0
321.00	338.0
321.90	208.0
323.00	4088.0
324.00	775.0
326.90	945.0
327.90	353.0
332.00	308.0
332.90	395.0
334.00	2328.0
335.00	657.0
341.00	436.0
345.90	1037.0
346.80	213.0
351.90	1153.0
352.90	823.0
354.00	1445.0
354.90	306.0
364.90	6020.0
366.00	792.0
370.90	316.0
372.00	2098.0
372.90	519.0
382.80	573.0
390.00	276.0
391.00	245.0
401.90	754.0
402.90	1097.0
403.90	338.0
420.90	747.0
422.00	843.0
423.00	5203.0
424.00	1560.0
425.00	160.0
441.00	14984.0
442.00	81808.0
443.00	17592.0
444.00	1565.0

Instrument :
BNA_N
ClientSampleId :
DFTPP228

m/z	Abundance
37.10	152.0
38.10	515.0
39.10	2582.0
40.00	481.0
41.00	301.0
42.90	190.0
44.00	1427.0
49.00	281.0
50.00	9387.0
51.00	33824.0
52.00	1702.0
55.00	443.0
55.90	1176.0
57.00	2559.0
60.90	525.0
62.00	613.0
63.00	1645.0
64.00	232.0
65.00	647.0
67.00	167.0
67.90	922.0
68.90	39944.0
69.90	272.0
73.00	1126.0
74.00	4255.0
75.00	6859.0
76.00	2282.0
77.00	46056.0
78.00	3428.0
79.00	3078.0
80.00	2533.0
81.00	3271.0
81.90	904.0
83.00	895.0
83.90	215.0
85.00	637.0
85.90	1243.0
87.00	494.0
91.00	863.0
91.90	917.0
93.00	6114.0
94.00	365.0
96.00	340.0
97.00	294.0
97.90	4896.0
99.00	3800.0
99.90	373.0
101.00	1804.0
102.90	765.0
103.90	1206.0
104.90	1315.0
105.90	389.0
107.00	14796.0
108.00	2233.0
109.00	512.0
110.00	25208.0
110.90	3725.0
111.90	383.0
112.90	189.0
114.80	152.0
116.00	853.0
117.00	15798.0
118.00	1039.0
119.00	232.0
119.90	231.0
120.90	155.0
121.90	1017.0
123.00	1682.0
123.90	809.0
124.90	656.0
127.00	52080.0
128.00	4211.0
128.90	21352.0
129.90	1725.0
131.00	462.0
131.90	164.0
132.90	299.0
133.90	587.0
135.00	2005.0
135.90	675.0
136.90	919.0
138.00	191.0
139.90	280.0
140.90	3028.0
141.90	855.0
142.90	653.0
143.90	233.0
145.00	169.0
145.90	473.0
147.00	1510.0
147.90	3936.0
148.90	852.0

Instrument :
BNA_N
ClientSampleId :
DFTPP228

149.90	212.0
151.00	423.0
151.80	217.0
153.00	880.0
153.90	643.0
154.90	1670.0
156.00	2268.0
156.90	363.0
157.90	486.0
159.00	384.0
159.90	1023.0
161.00	1416.0
161.80	332.0
162.90	178.0
163.90	250.0
164.90	1066.0
165.90	915.0
167.00	6589.0
168.00	3614.0
168.90	549.0
170.00	214.0
170.90	246.0
171.90	516.0
172.90	649.0
173.90	1083.0
175.00	2238.0
176.00	636.0
176.90	1206.0
177.90	427.0
178.90	4335.0
180.00	3167.0
180.90	1348.0
181.90	201.0
183.90	349.0
185.00	2077.0
186.00	15343.0
187.00	4817.0
187.90	501.0
189.00	1039.0
189.80	196.0
190.90	665.0
191.90	1592.0
192.90	1744.0
193.90	314.0
194.90	154.0
195.90	3232.0
197.90	105272.0
198.90	7060.0
199.90	605.0
201.20	461.0
202.90	914.0
203.90	4245.0
205.00	7469.0
206.00	31016.0
207.00	5754.0
207.90	1334.0
209.00	722.0
210.10	533.0
210.90	1264.0
211.70	223.0
215.00	486.0
215.90	702.0
216.90	8114.0
218.00	946.0
220.00	213.0
221.00	4172.0
222.90	2023.0
224.00	18208.0
225.00	4480.0
226.10	432.0
226.90	8086.0
228.00	1044.0
229.00	1518.0
230.00	268.0
231.00	773.0
233.90	484.0
234.90	641.0
236.00	385.0
237.00	894.0
238.90	285.0
239.90	283.0
240.90	392.0
242.00	970.0
242.90	1088.0
244.00	13977.0
245.00	1738.0
245.90	2770.0
246.90	557.0
249.00	498.0
250.90	163.0
251.80	207.0
252.90	356.0
254.00	543.0
255.00	68352.0
256.00	9411.0
256.90	776.0

Instrument :
BNA_N
ClientSampleId :
DFTPP228

257.90	4485.0
259.00	742.0
260.00	159.0
260.90	203.0
265.00	1800.0
265.80	307.0
270.90	229.0
271.90	177.0
273.00	2076.0
274.00	5717.0
275.00	29432.0
275.90	4275.0
276.90	3221.0
277.90	513.0
278.90	179.0
281.00	456.0
282.90	486.0
284.00	306.0
285.00	483.0
292.90	728.0
294.00	154.0
294.90	249.0
295.90	9677.0
297.00	1277.0
302.90	1217.0
303.90	270.0
308.00	176.0
309.90	214.0
314.00	456.0
314.90	1131.0
316.00	714.0
321.00	362.0
323.00	2947.0
324.00	474.0
326.90	808.0
327.90	320.0
332.00	377.0
332.90	337.0
334.00	2261.0
334.90	552.0
341.00	381.0
345.80	641.0
352.00	835.0
352.90	766.0
354.00	1026.0
355.00	266.0
365.00	5256.0
365.80	591.0
370.90	173.0
372.00	1805.0
373.00	384.0
382.80	365.0
390.00	168.0
390.90	155.0
402.00	649.0
402.90	878.0
403.90	299.0
420.90	800.0
422.00	668.0
423.00	5134.0
423.90	1212.0
441.00	12709.0
442.00	72944.0
443.00	15725.0
444.00	1308.0

Instrument :
BNA_N
ClientSampleId :
DFTPP228