

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081321\
 Data File : BN015868.D
 Acq On : 13 Aug 2021 08:38
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 13 11:06:05 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 13 10:59:04 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	97	0.00
2	1,4-Dioxane	0.400	0.418	-4.5	99	0.00
3	n-Nitrosodimethylamine	0.400	0.376	6.0	98	0.00
4 S	2-Fluorophenol	0.400	0.377	5.8	93	0.00
5 S	Phenol-d6	0.400	0.376	6.0	94	0.00
6	bis(2-Chloroethyl)ether	0.400	0.410	-2.5	97	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	0.00
8 S	Nitrobenzene-d5	0.400	0.364	9.0	99	0.00
9	Naphthalene	0.400	0.394	1.5	96	0.00
10	Hexachlorobutadiene	0.400	0.399	0.3	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.392	2.0	96	0.00
12	2-Methylnaphthalene	0.400	0.389	2.8	90	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.400	0.326	18.5	94	0.00
15 S	2-Fluorobiphenyl	0.400	0.386	3.5	98	0.00
16	Acenaphthylene	0.400	0.373	6.8	96	0.00
17	Acenaphthene	0.400	0.382	4.5	97	0.00
18	Fluorene	0.400	0.387	3.3	97	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	97	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.272	32.0#	102	0.00
21	4-Bromophenyl-phenylether	0.400	0.375	6.3	96	0.00
22	Hexachlorobenzene	0.400	0.386	3.5	97	0.00
23	Atrazine	0.400	0.337	15.8	95	0.00
24	Pentachlorophenol	0.400	0.310	22.5	97	0.00
25	Phenanthrene	0.400	0.382	4.5	97	0.00
26	Anthracene	0.400	0.366	8.5	97	0.00
27 SURR	Fluoranthene-d10	0.400	0.377	5.8	98	0.00
28	Fluoranthene	0.400	0.387	3.3	98	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	99	0.00
30	Pyrene	0.400	0.383	4.3	100	0.00
31 S	Terphenyl-d14	0.400	0.375	6.3	99	0.00
32	Benzo(a)anthracene	0.400	0.359	10.3	94	0.00
33	Chrysene	0.400	0.384	4.0	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.322	19.5	101	0.00
35 I	Perylene-d12	0.400	0.400	0.0	98	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.371	7.3	98	0.00
37	Benzo(b)fluoranthene	0.400	0.381	4.8	97	0.00
38	Benzo(k)fluoranthene	0.400	0.396	1.0	100	0.00
39 C	Benzo(a)pyrene	0.400	0.373	6.8	97	0.00
40	Dibenzo(a,h)anthracene	0.400	0.368	8.0	97	0.00
41	Benzo(g,h,i)perylene	0.400	0.372	7.0	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4