

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081321\
 Data File : BN015868.D
 Acq On : 13 Aug 2021 08:38
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

mohammad

Quant Time: Aug 13 11:06:05 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 13 10:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.919	152	8062	0.400	ng	0.00
7) Naphthalene-d8	10.723	136	39352	0.400	ng	0.00
13) Acenaphthene-d10	14.548	164	22762	0.400	ng	0.00
19) Phenanthrene-d10	17.297	188	46611	0.400	ng	0.00
29) Chrysene-d12	21.493	240	52440	0.400	ng	# 0.00
35) Perylene-d12	23.922	264	47739	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.494	112	7521	0.377	ng	0.00
5) Phenol-d6	7.071	99	10186	0.376	ng	0.00
8) Nitrobenzene-d5	9.076	82	11967	0.364	ng	0.00
11) 2-Methylnaphthalene-d10	12.309	152	24838	0.392	ng	0.00
14) 2,4,6-Tribromophenol	16.044	330	3270	0.326	ng	0.00
15) 2-Fluorobiphenyl	13.177	172	34599	0.386	ng	0.00
27) Fluoranthene-d10	19.336	212	54184	0.377	ng	0.00
31) Terphenyl-d14	19.931	244	55804	0.375	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.410	88	1328	0.418	ng	98
3) n-Nitrosodimethylamine	3.770	42	3085	0.376	ng	# 97
6) bis(2-Chloroethyl)ether	7.347	93	8818	0.410	ng	99
9) Naphthalene	10.774	128	40075	0.394	ng	100
10) Hexachlorobutadiene	11.066	225	11070	0.399	ng	# 99
12) 2-Methylnaphthalene	12.385	142	27270	0.389	ng	99
16) Acenaphthylene	14.269	152	34273	0.373	ng	100
17) Acenaphthene	14.611	154	24268	0.382	ng	99
18) Fluorene	15.601	166	30656	0.387	ng	100
20) 4,6-Dinitro-2-methylph...	15.677	198	1231	0.272	ng	98
21) 4-Bromophenyl-phenylether	16.494	248	9470	0.375	ng	96
22) Hexachlorobenzene	16.616	284	12732	0.386	ng	98
23) Atrazine	16.762	200	7029	0.337	ng	100
24) Pentachlorophenol	16.957	266	3722	0.310	ng	100
25) Phenanthrene	17.346	178	48784	0.382	ng	99
26) Anthracene	17.431	178	41662	0.366	ng	100
28) Fluoranthene	19.366	202	59906	0.387	ng	100
30) Pyrene	19.728	202	59662	0.383	ng	100
32) Benzo(a)anthracene	21.471	228	59420	0.359	ng	99
33) Chrysene	21.525	228	63439	0.384	ng	99
34) Bis(2-ethylhexyl)phtha...	21.408	149	22046m	0.322	ng	
36) Indeno(1,2,3-cd)pyrene	26.431	276	59255	0.371	ng	100
37) Benzo(b)fluoranthene	23.183	252	62780	0.381	ng	100
38) Benzo(k)fluoranthene	23.228	252	61351	0.396	ng	100
39) Benzo(a)pyrene	23.813	252	52760	0.373	ng	100
40) Dibenzo(a,h)anthracene	26.455	278	48231	0.368	ng	99
41) Benzo(g,h,i)perylene	27.208	276	51379	0.372	ng	100

8/16/2021 12:22:20 PM

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081321\
Data File : BN015868.D
Acq On : 13 Aug 2021 08:38
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED


mohammad

Quant Time: Aug 13 11:06:05 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN081221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 13 10:59:04 2021
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.919 min Scan# 525
Delta R.T. -0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Tgt Ion: 152 Resp: 8062
Ion Ratio Lower Upper
152 100
150 156.6 125.8 188.8
115 67.3 53.6 80.4




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



8/16/2021 12:22:20 PM




#2
1,4-Dioxane
Concen: 0.418 ng
RT: 3.410 min Scan# 36
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Tgt Ion: 88 Resp: 1328
Ion Ratio Lower Upper
88 100
43 37.0 29.4 44.2
58 83.9 69.3 103.9





#3
n-Nitrosodimethylamine
Concen: 0.376 ng
RT: 3.770 min Scan# 75
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38





Tgt Ion: 42 Resp: 3085
Ion Ratio Lower Upper
42 100
74 104.4 86.1 129.1
44 7.8 3.8 5.8#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


8/16/2021 12:22:20 PM



#4
2-Fluorophenol
Concen: 0.377 ng
RT: 5.494 min Scan# 262
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion: 112 Resp: 7521
Ion Ratio Lower Upper
112 100
64 66.5 53.3 79.9
63 36.0 28.6 42.8









#5
Phenol-d6
Concen: 0.376 ng
RT: 7.071 min Scan# 433
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion	Ratio	Lower	Upper
99	100		
42	23.2	18.7	28.1
71	40.5	32.4	48.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad



#6
bis(2-Chloroethyl)ether
Concen: 0.410 ng
RT: 7.347 min Scan# 463
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion	Ratio	Lower	Upper
93	100		
63	78.3	63.8	95.6
95	32.5	26.0	39.0







#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.723 min Scan# 758
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion:	136	Resp:	39352
Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.4
54	7.2	5.4	8.2
68	5.1	3.8	5.8



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







8/16/2021 12:22:20 PM



#8
Nitrobenzene-d5
Concen: 0.364 ng
RT: 9.076 min Scan# 628
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion:	82	Resp:	11967
Ion	Ratio	Lower	Upper
82	100		
128	36.1	29.2	43.8
54	53.2	41.1	61.7









#9
Naphthalene
Concen: 0.394 ng
RT: 10.774 min Scan# 762
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion:	128	Resp:	40075
Ion Ratio	Lower	Upper	
128	100		
129	11.2	8.8	13.2
127	13.1	10.5	15.7

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad



#10
Hexachlorobutadiene
Concen: 0.399 ng
RT: 11.066 min Scan# 785
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion:	225	Resp:	11070
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.6	51.4	77.0







#11
2-Methylnaphthalene-d10
Concen: 0.392 ng
RT: 12.309 min Scan# 1000
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion:152 Resp: 24838
Ion Ratio Lower Upper
152 100
151 20.5 16.5 24.7



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







8/16/2021 12:22:20 PM




#12
2-Methylnaphthalene
Concen: 0.389 ng
RT: 12.385 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion:142 Resp: 27270
Ion Ratio Lower Upper
142 100
141 88.7 70.5 105.7
115 38.3 29.8 44.8







#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.548 min Scan# 1205
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Tgt Ion:164 Resp: 22762
Ion Ratio Lower Upper
164 100
162 106.0 83.5 125.3
160 48.2 37.7 56.5




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



8/16/2021 12:22:20 PM




#14
2,4,6-Tribromophenol
Concen: 0.326 ng
RT: 16.044 min Scan# 1323
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38




Tgt Ion:330 Resp: 3270
Ion Ratio Lower Upper
330 100
332 108.2 84.1 126.1
141 30.3 24.2 36.4









#15
2-Fluorobiphenyl
Concen: 0.386 ng
RT: 13.177 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38





Tgt Ion:	172	Resp:	34599
Ion Ratio	Lower	Upper	
172	100		
171	35.6	28.7	43.1
170	23.7	19.0	28.4



#16
Acenaphthylene
Concen: 0.373 ng
RT: 14.269 min Scan# 1183
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Tgt Ion:	152	Resp:	34273
Ion Ratio	Lower	Upper	
152	100		
151	19.1	15.4	23.2
153	13.0	10.5	15.7




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



8/16/2021 12:22:20 PM



#17
Acenaphthene
Concen: 0.382 ng
RT: 14.611 min Scan# 1210
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Tgt Ion:154 Resp: 24268
Ion Ratio Lower Upper
154 100
153 111.9 90.1 135.1
152 53.4 43.5 65.3




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



8/16/2021 12:22:20 PM




#18
Fluorene
Concen: 0.387 ng
RT: 15.601 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38




Tgt Ion:166 Resp: 30656
Ion Ratio Lower Upper
166 100
165 97.6 77.8 116.6
167 13.6 11.0 16.4





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.297 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	9.8	7.2	10.8

Manual Integrations
APPROVED


mohammad





8/16/2021 12:22:20 PM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.272 ng
RT: 15.677 min Scan# 1294
Delta R.T. -0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38




Tgt Ion	Ratio	Lower	Upper
198	100		
182	14.2	11.9	17.9
77	0.0	0.0	0.0





#21
4-Bromophenyl-phenylether
Concen: 0.375 ng
RT: 16.494 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion	Ratio	Lower	Upper
248	100		
250	97.4	78.6	118.0
141	60.9	43.4	65.2

Manual Integrations
APPROVED


mohammad





8/16/2021 12:22:20 PM




#22
Hexachlorobenzene
Concen: 0.386 ng
RT: 16.616 min Scan# 1370
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38




Tgt Ion	Ratio	Lower	Upper
284	100		
142	29.4	22.6	34.0
249	32.2	25.2	37.8







#23
Atrazine
Concen: 0.337 ng
RT: 16.762 min Scan# 1382
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Tgt Ion: 200 Resp: 7029
Ion Ratio Lower Upper
200 100
173 20.7 16.4 24.6
215 51.9 41.5 62.3




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



8/16/2021 12:22:20 PM




#24
Pentachlorophenol
Concen: 0.310 ng
RT: 16.957 min Scan# 1398
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Tgt Ion: 266 Resp: 3722
Ion Ratio Lower Upper
266 100
264 62.8 50.1 75.1
268 63.9 51.0 76.6







#25
Phenanthrene
Concen: 0.382 ng
RT: 17.346 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion	178	176	179
Ratio	100	18.5	16.0
Lower		14.6	12.6
Upper		21.8	18.8



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







8/16/2021 12:22:20 PM



#26
Anthracene
Concen: 0.366 ng
RT: 17.431 min Scan# 1437
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion	178	176	179
Ratio	100	17.9	15.4
Lower		14.3	12.2
Upper		21.5	18.4







#27
Fluoranthene-d10
Concen: 0.377 ng
RT: 19.336 min Scan# 1751
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion:	212	Resp:	54184
Ion Ratio	Lower	Upper	
212	100		
106	9.0	7.4	11.2
104	5.3	4.3	6.5



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad







8/16/2021 12:22:20 PM




#28
Fluoranthene
Concen: 0.387 ng
RT: 19.366 min Scan# 1757
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion:	202	Resp:	59906
Ion Ratio	Lower	Upper	
202	100		
101	8.5	6.8	10.2
203	17.4	13.8	20.8







#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.493 min Scan# 2026
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Tgt Ion:240 Resp: 52440
Ion Ratio Lower Upper
240 100
120 6.2 3.8 5.8#
236 25.7 20.0 30.0




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



8/16/2021 12:22:20 PM




#30
Pyrene
Concen: 0.383 ng
RT: 19.728 min Scan# 1830
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38




Tgt Ion:202 Resp: 59662
Ion Ratio Lower Upper
202 100
200 20.2 16.2 24.2
203 17.5 14.1 21.1






#31
Terphenyl-d14
Concen: 0.375 ng
RT: 19.931 min Scan# 1871
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Tgt Ion:244 Resp: 55804
Ion Ratio Lower Upper
244 100
212 7.4 6.0 9.0
122 8.2 6.5 9.7




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



8/16/2021 12:22:20 PM




#32
Benzo(a)anthracene
Concen: 0.359 ng
RT: 21.471 min Scan# 2024
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Tgt Ion:228 Resp: 59420
Ion Ratio Lower Upper
228 100
226 27.1 21.1 31.7
229 19.1 15.7 23.5









#33
Chrysene
Concen: 0.384 ng
RT: 21.525 min Scan# 2029
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.9	23.4	35.2
229	19.4	15.8	23.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.322 ng m
RT: 21.408 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38


Tgt Ion	Ratio	Lower	Upper
149	100		
167	26.1	22.9	34.3
279	5.1	4.4	6.6





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.922 min Scan# 2555
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:264 Resp: 47739
Ion Ratio Lower Upper
264 100
260 23.8 19.1 28.7
265 24.6 19.8 29.8

Manual Integrations
APPROVED


mohammad





8/16/2021 12:22:20 PM




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.371 ng
RT: 26.431 min Scan# 3288
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Tgt Ion:276 Resp: 59255
Ion Ratio Lower Upper
276 100
138 17.5 13.8 20.8
277 25.3 20.3 30.5





#37
Benzo(b)fluoranthene
Concen: 0.381 ng
RT: 23.183 min Scan# 2339
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38





Tgt Ion:252 Resp: 62780
Ion Ratio Lower Upper
252 100
253 22.7 18.2 27.2
125 8.1 6.2 9.4

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad


8/16/2021 12:22:20 PM




#38
Benzo(k)fluoranthene
Concen: 0.396 ng
RT: 23.228 min Scan# 2352
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38

Tgt Ion:252 Resp: 61351
Ion Ratio Lower Upper
252 100
253 22.6 18.1 27.1
125 7.1 5.6 8.4







#39
Benzo(a)pyrene
Concen: 0.373 ng
RT: 23.813 min Scan# 2523
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Tgt Ion:252 Resp: 52760
Ion Ratio Lower Upper
252 100
253 23.2 18.6 27.8
125 9.3 7.3 10.9




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Manual Integrations
APPROVED

mohammad



8/16/2021 12:22:20 PM




#40
Dibenzo(a,h)anthracene
Concen: 0.368 ng
RT: 26.455 min Scan# 3295
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38




Tgt Ion:278 Resp: 48231
Ion Ratio Lower Upper
278 100
139 13.7 10.2 15.2
279 24.6 19.5 29.3







#41
Benzo(g,h,i)perylene
Concen: 0.372 ng
RT: 27.208 min Scan# 3515
Delta R.T. 0.000 min
Lab File: BN015868.D
Acq: 13 Aug 2021 08:38



Tgt Ion:	276	Resp:	51379
Ion Ratio	Lower	Upper	
276	100		
277	24.2	19.5	29.3
138	14.7	11.8	17.8



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad

8/16/2021 12:22:20 PM