


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081621\
 Data File : BN015874.D
 Acq On : 16 Aug 2021 11:43
 Operator : CG/JU
 Sample : DFTPP201
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP201

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN081621MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Aug 17 09:43:53 2021



AutoFind: Scans 2491, 2492, 2493; Background Corrected with Scan 2485

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.5	420053	PASS
68	69	0.00	2	1.9	8662	PASS
69	198	0.00	100	46.4	458901	PASS
70	69	0.00	2	0.6	2597	PASS
127	198	10	80	50.3	497408	PASS
197	198	0.00	2	0.3	2905	PASS
198	198	100	100	100.0	988608	PASS
199	198	5	9	6.7	66128	PASS
275	198	10	60	28.7	284032	PASS
365	198	1	100	3.9	38765	PASS
441	198	0.01	100	13.5	133589	PASS
442	198	50	100	83.2	822933	PASS
443	442	15	24	20.4	167944	PASS

m/z	Abundance
40.00	879.0
44.00	1891.0
48.90	752.0
73.00	1538.0
84.10	675.0
86.00	618.0
96.20	1131.0
119.00	322.0
133.00	794.0
147.10	789.0
191.00	543.0
192.90	841.0
207.10	7088.0
208.00	1230.0
209.10	1061.0
221.10	374.0
280.90	669.0
282.00	441.0

Instrument :
BNA_N
ClientSampleId :
DFTPP201

m/z	Abundance
35.00	394.0
35.80	331.0
36.90	1574.0
38.10	4366.0
39.10	24776.0
40.00	2392.0
40.90	818.0
43.10	321.0
44.00	2294.0
45.10	1401.0
46.90	481.0
47.90	587.0
49.00	3036.0
50.00	93368.0
51.10	340416.0
52.10	17304.0
53.00	710.0
54.90	1442.0
56.00	10927.0
57.00	23048.0
57.90	1581.0
58.80	406.0
61.00	4330.0
62.00	5697.0
63.00	15107.0
63.90	2252.0
65.00	6416.0
66.10	516.0
68.10	7236.0
69.00	371456.0
70.00	2230.0
73.00	4231.0
74.00	37688.0
75.00	58960.0
76.10	18584.0
77.00	381888.0
78.00	27456.0
79.00	29840.0
80.00	21072.0
81.00	30504.0
82.00	7317.0
83.00	5663.0
84.10	1344.0
84.90	5205.0
86.00	9682.0
87.00	3740.0
88.00	1430.0
89.00	1024.0
91.00	7114.0
92.00	8000.0
93.00	46656.0
93.90	3082.0
95.10	746.0
96.00	2952.0
97.00	1130.0
98.00	35832.0
99.00	27400.0
100.00	2184.0
101.00	15057.0
102.00	1415.0
103.00	5479.0
104.00	10389.0
105.00	8860.0
106.10	3462.0
107.00	119432.0
108.00	17944.0
108.90	3475.0
110.00	208192.0
111.00	32032.0
112.10	4115.0
113.00	1226.0
114.80	553.0
116.00	8539.0
117.00	106424.0
118.00	7639.0
119.00	1257.0
120.10	1545.0
120.90	732.0
122.00	8681.0
123.10	11543.0
124.00	5797.0
125.00	5274.0
126.10	2292.0
127.00	374144.0
128.00	28008.0
129.00	154880.0
130.00	15207.0
130.90	2549.0
132.00	1083.0
133.10	1143.0
134.00	4691.0
135.00	12137.0

Instrument :
BNA_N
ClientSampleId :
DFTPP201

135.90	5929.0
137.00	7390.0
138.10	1500.0
139.00	826.0
140.00	2029.0
141.00	23456.0
142.00	6761.0
143.00	4311.0
144.00	1268.0
145.10	857.0
145.90	3958.0
147.00	11009.0
148.00	25096.0
149.00	3770.0
150.00	1598.0
151.20	2212.0
151.80	1287.0
152.90	7310.0
154.00	4719.0
155.00	12009.0
156.00	17760.0
157.00	2789.0
158.00	3235.0
159.00	2957.0
160.00	5967.0
161.00	9896.0
162.00	2770.0
163.00	1156.0
163.60	603.0
164.10	480.0
165.00	8087.0
166.00	6425.0
167.00	38048.0
168.00	20248.0
169.00	3775.0
169.90	1830.0
171.00	2207.0
171.90	4505.0
173.10	4242.0
174.00	7396.0
175.00	15221.0
176.10	4575.0
177.00	7127.0
178.00	3217.0
179.00	29472.0
180.00	22240.0
181.00	9627.0
182.00	1264.0
183.00	749.0
184.00	2583.0
185.00	13140.0
186.00	107224.0
187.00	31480.0
188.00	3133.0
189.00	7074.0
189.90	1470.0
191.00	4372.0
192.00	8942.0
193.00	11875.0
194.00	2647.0
194.90	1242.0
196.00	21720.0
198.00	685376.0
199.00	48032.0
200.00	4077.0
201.40	2019.0
203.10	5241.0
204.00	28240.0
205.00	44584.0
206.10	205952.0
207.00	32808.0
208.10	8656.0
209.00	3694.0
210.20	3504.0
211.00	7675.0
213.10	530.0
214.90	2442.0
215.90	3954.0
217.00	54608.0
218.00	7793.0
218.90	688.0
221.00	22504.0
223.00	10215.0
224.00	110176.0
225.00	29032.0
226.00	3410.0
227.00	51704.0
228.10	7036.0
229.00	10222.0
229.90	924.0
230.20	967.0
231.00	3979.0
231.90	646.0
233.00	1407.0
234.00	3391.0
235.00	4122.0

Instrument :
BNA_N
ClientSampleId :
DFTPP201

236.10	2265.0
237.00	3423.0
237.80	463.0
239.10	2013.0
240.00	1174.0
241.00	3263.0
242.00	6557.0
243.00	7296.0
244.00	72616.0
245.00	12879.0
246.00	15052.0
247.00	4133.0
247.80	892.0
249.10	2007.0
249.90	593.0
250.90	667.0
251.90	772.0
253.00	1300.0
255.00	407232.0
256.00	64568.0
257.00	6253.0
257.90	26384.0
259.00	4251.0
260.00	614.0
260.80	664.0
263.00	477.0
263.80	579.0
265.00	11045.0
266.00	1413.0
267.20	305.0
270.00	382.0
271.00	862.0
271.90	1696.0
272.90	12045.0
274.00	32568.0
275.00	186560.0
276.00	21784.0
277.00	16664.0
278.00	3467.0
278.90	947.0
281.10	1109.0
282.30	606.0
283.10	2005.0
283.90	1387.0
285.00	3275.0
285.90	716.0
289.00	902.0
290.00	718.0
290.90	530.0
292.00	655.0
292.90	3041.0
294.10	1061.0
294.90	1245.0
296.00	55360.0
297.00	8758.0
298.00	661.0
300.90	708.0
301.90	1188.0
302.20	1222.0
303.10	6537.0
304.10	1744.0
307.90	850.0
310.00	1172.0
312.90	631.0
313.90	2653.0
315.00	5385.0
316.10	2468.0
317.00	816.0
320.90	1574.0
322.00	969.0
323.00	16289.0
324.00	3252.0
327.00	2897.0
327.90	1288.0
329.20	325.0
332.00	1123.0
333.10	1689.0
334.00	10256.0
335.00	2789.0
340.00	359.0
341.00	1984.0
341.90	704.0
346.00	3326.0
347.00	886.0
351.00	586.0
352.00	4502.0
352.90	2862.0
354.00	5371.0
355.00	987.0
358.90	477.0
364.00	637.0
365.00	24096.0
366.00	2728.0
367.00	352.0
370.20	647.0
371.10	1753.0

Instrument :
BNA_N
ClientSampleId :
DFTPP201

372.00	8943.0
373.00	2429.0
382.90	2297.0
383.90	912.0
390.10	1006.0
391.00	715.0
392.10	424.0
400.80	325.0
401.90	2831.0
403.00	4300.0
404.10	1738.0
421.00	3412.0
422.10	3361.0
423.00	26320.0
424.10	6313.0
425.20	782.0
441.00	69120.0
442.00	498304.0
443.00	103320.0
444.00	8817.0
445.10	519.0

Instrument :
BNA_N
ClientSampleId :
DFTPP201

m/z	Abundance
37.00	2232.0
38.10	6473.0
39.10	33504.0
40.00	2837.0
41.10	656.0
44.10	2161.0
45.00	820.0
48.20	467.0
49.00	4454.0
50.00	128504.0
51.00	488640.0
52.10	25384.0
53.10	806.0
55.00	2239.0
56.00	14575.0
57.00	32080.0
58.00	1624.0
58.90	554.0
60.00	357.0
61.00	6107.0
62.10	8188.0
63.00	21624.0
64.00	2777.0
65.10	9292.0
66.10	955.0
68.10	10361.0
69.00	536064.0
70.00	2480.0
71.10	497.0
73.10	5025.0
74.00	53456.0
75.00	83752.0
76.10	29584.0
77.00	547200.0
78.10	36856.0
79.00	44936.0
80.00	32848.0
81.00	45912.0
82.00	11180.0
83.00	9960.0
83.90	1213.0
85.00	8274.0
86.00	13352.0
87.00	6755.0
87.90	1807.0
89.10	954.0
90.10	386.0
91.00	10008.0
92.00	12172.0
93.00	73824.0
94.00	5107.0
95.10	1410.0
96.00	3584.0
97.10	1295.0
98.00	56552.0
99.00	41432.0
100.00	3095.0
101.00	22760.0
101.90	1688.0
103.00	7743.0
104.00	14288.0
105.00	14585.0
106.00	4752.0
107.00	177408.0
108.00	27128.0
109.10	5777.0
110.00	306368.0
111.00	45304.0
111.90	5251.0
113.00	1748.0
114.20	545.0
114.90	924.0
116.00	11296.0
117.00	158464.0
118.00	11728.0
118.90	1038.0
119.90	2504.0
120.90	779.0
122.00	10952.0
123.00	17184.0
124.00	8475.0
125.00	7269.0
127.00	579904.0
128.00	45336.0
129.00	225984.0
130.00	18784.0
131.00	3587.0
132.10	2083.0
132.80	1918.0
134.00	8174.0
135.00	19264.0
136.00	7927.0

Instrument :
BNA_N
ClientSampleId :
DFTPP201

137.00	7749.0
137.90	2019.0
139.10	1430.0
140.10	3382.0
141.00	32704.0
142.00	10870.0
143.00	6260.0
144.00	2005.0
145.10	1951.0
146.00	4928.0
147.00	16984.0
148.00	34544.0
149.00	5566.0
150.00	1831.0
151.10	3432.0
152.00	2235.0
153.00	10742.0
154.00	6503.0
155.00	15339.0
156.00	23232.0
156.90	5459.0
158.00	7124.0
159.00	4426.0
160.00	11318.0
161.00	15379.0
162.00	3983.0
163.00	1234.0
164.10	1233.0
165.00	11381.0
166.00	9521.0
167.00	59328.0
168.00	27384.0
169.00	4087.0
170.00	1971.0
170.90	2459.0
172.00	5734.0
173.00	6337.0
174.00	12748.0
175.00	23176.0
176.10	6871.0
177.00	12108.0
178.00	3863.0
179.00	46912.0
180.00	29840.0
180.90	13645.0
181.90	2116.0
183.10	992.0
184.00	3948.0
185.00	21568.0
186.00	166528.0
187.00	46280.0
188.00	4523.0
189.00	11828.0
190.10	1600.0
191.00	6000.0
192.00	14298.0
193.00	16608.0
194.00	4302.0
194.90	1131.0
196.00	33016.0
197.10	8715.0
198.00	1156608.0
199.00	76224.0
200.00	6196.0
201.50	3991.0
203.00	8925.0
204.00	45792.0
205.10	75272.0
206.00	317568.0
207.00	48584.0
208.00	12947.0
209.00	4260.0
210.00	4904.0
211.00	12148.0
212.00	1807.0
212.90	742.0
214.00	306.0
215.00	4173.0
216.00	6863.0
217.00	89784.0
218.00	11784.0
219.00	1184.0
221.00	36176.0
221.90	9891.0
223.00	18976.0
224.10	176192.0
225.10	46008.0
226.00	5160.0
227.00	76128.0
228.00	12520.0
229.00	16153.0
230.00	1980.0
231.00	6487.0
232.00	1538.0
233.00	1909.0
234.00	5425.0

Instrument :
BNA_N
ClientSampleId :
DFTPP201

235.10	5796.0
236.00	4208.0
237.00	6619.0
237.90	965.0
239.00	2773.0
240.00	2732.0
240.90	4482.0
242.00	9519.0
243.10	9654.0
244.00	125136.0
245.00	18840.0
246.00	27400.0
247.00	6536.0
247.90	1125.0
249.00	4848.0
250.00	843.0
251.00	997.0
251.90	1381.0
253.00	2485.0
254.10	5975.0
255.00	706112.0
256.00	106336.0
257.00	9197.0
258.00	48136.0
259.00	7113.0
260.00	1244.0
261.00	1317.0
263.00	449.0
263.90	1314.0
265.00	17048.0
265.90	3163.0
267.00	812.0
270.10	801.0
271.00	2059.0
272.20	1934.0
273.00	20992.0
274.00	55896.0
275.00	318912.0
276.00	39040.0
277.00	28424.0
278.00	4888.0
279.00	790.0
281.00	969.0
282.10	986.0
283.10	3233.0
284.00	2007.0
285.10	5068.0
285.90	1137.0
287.80	311.0
288.80	853.0
290.00	1156.0
290.80	743.0
292.10	1153.0
293.00	5757.0
293.90	1237.0
295.00	2758.0
296.00	93624.0
297.00	13451.0
298.00	790.0
300.90	1052.0
302.00	1564.0
303.00	11535.0
304.00	2489.0
304.80	479.0
308.10	1445.0
308.90	919.0
310.00	993.0
311.90	453.0
313.10	1022.0
314.10	4220.0
315.00	11132.0
316.00	5368.0
317.00	1739.0
320.00	332.0
321.00	3120.0
321.90	1447.0
323.00	29888.0
324.00	5478.0
325.00	557.0
327.00	5486.0
328.00	2415.0
328.90	751.0
331.90	1738.0
333.00	2389.0
334.00	19512.0
335.00	4664.0
336.10	540.0
340.00	623.0
340.90	3406.0
342.10	1109.0
346.00	5851.0
346.90	626.0
351.10	381.0
352.00	7917.0
353.10	6009.0
354.10	9925.0

Instrument :
BNA_N
ClientSampleId :
DFTPP201

355.00	1728.0
358.00	335.0
358.90	542.0
365.00	47056.0
366.00	5358.0
366.80	471.0
370.00	988.0
371.10	2253.0
372.00	15944.0
373.00	3635.0
373.80	387.0
377.10	496.0
383.00	5279.0
384.00	1041.0
385.00	622.0
390.00	2191.0
391.10	1110.0
392.00	1441.0
400.90	1279.0
402.00	6089.0
403.00	8362.0
404.00	3018.0
404.90	739.0
410.00	420.0
414.90	397.0
421.00	7637.0
422.00	7937.0
423.00	52936.0
424.00	11698.0
424.90	1494.0
441.00	149568.0
442.00	902464.0
442.90	177792.0
444.00	18192.0
445.00	1154.0

Instrument :
BNA_N
ClientSampleId :
DFTPP201

m/z	Abundance
37.10	2338.0
38.00	5001.0
39.10	32976.0
40.00	2847.0
41.10	778.0
44.00	1760.0
45.10	454.0
47.10	389.0
49.10	4417.0
50.00	121408.0
51.00	431104.0
52.00	22376.0
52.90	872.0
55.00	3281.0
56.00	13459.0
57.00	31304.0
58.00	1144.0
58.90	510.0
61.00	6436.0
62.00	7418.0
63.00	19344.0
64.00	2640.0
65.00	10059.0
66.10	541.0
67.10	431.0
68.00	8388.0
69.00	469184.0
70.00	3080.0
71.10	387.0
73.10	5293.0
74.00	55080.0
75.00	77696.0
76.10	25256.0
77.00	518400.0
78.00	32576.0
79.00	40128.0
80.00	30016.0
81.00	40656.0
82.00	9191.0
83.00	8360.0
84.00	1317.0
85.00	7476.0
86.00	14385.0
87.00	4600.0
88.00	1966.0
88.90	1192.0
91.00	9678.0
91.90	10432.0
93.00	70120.0
94.00	4257.0
95.10	864.0
96.10	3244.0
96.90	1372.0
97.10	1359.0
98.00	49952.0
99.00	36568.0
99.90	3515.0
101.00	22592.0
102.00	1141.0
102.90	7394.0
104.00	13968.0
105.00	14536.0
107.00	167808.0
108.00	25768.0
109.10	4646.0
110.00	296000.0
111.00	44888.0
112.00	6187.0
112.90	2194.0
114.30	342.0
114.90	498.0
116.00	9834.0
117.00	148096.0
118.00	10998.0
119.00	1136.0
120.00	1971.0
120.80	572.0
122.00	11302.0
123.00	19160.0
124.00	8267.0
125.00	6938.0
127.00	538176.0
128.00	39480.0
129.00	222656.0
130.00	20112.0
131.00	3974.0
132.10	1630.0
132.80	1338.0
134.00	6693.0
135.00	17008.0
136.00	6793.0
137.00	9349.0

Instrument :
BNA_N
ClientSampleId :
DFTPP201

137.80	1642.0
139.00	1608.0
139.90	2700.0
141.00	31736.0
142.00	9833.0
143.10	6736.0
144.00	2317.0
145.00	1369.0
146.00	5213.0
147.00	14947.0
148.00	39016.0
149.00	7610.0
150.10	1755.0
151.20	3238.0
152.00	2257.0
153.00	10062.0
154.00	7520.0
155.00	16125.0
156.00	25512.0
157.10	5222.0
157.90	6022.0
159.00	4530.0
159.90	9360.0
161.00	13996.0
162.00	4148.0
163.00	1060.0
164.10	1639.0
165.00	12249.0
166.00	7623.0
167.00	58128.0
168.00	26336.0
169.00	5184.0
170.00	2570.0
170.90	2785.0
172.00	7081.0
173.00	6712.0
174.00	15361.0
175.00	25344.0
176.00	6068.0
177.00	11576.0
178.00	3848.0
179.00	47320.0
180.00	31016.0
181.00	14021.0
182.00	2098.0
183.00	1133.0
184.00	3879.0
185.00	21824.0
186.00	165696.0
187.00	51592.0
188.00	5119.0
189.00	11535.0
190.10	1842.0
191.00	4641.0
192.10	13549.0
193.00	14991.0
194.00	4167.0
194.90	1950.0
196.00	30448.0
198.00	1123840.0
199.00	74128.0
200.00	5851.0
201.40	4799.0
203.00	10206.0
204.00	43472.0
205.00	81088.0
206.00	323904.0
207.00	48336.0
208.00	12517.0
209.00	5131.0
210.00	5467.0
210.20	5288.0
211.00	12076.0
211.90	3257.0
213.00	1232.0
215.10	3676.0
216.10	8614.0
217.00	89088.0
218.00	12517.0
219.00	1058.0
220.20	1556.0
221.00	38552.0
221.90	12257.0
223.00	19208.0
224.00	185472.0
225.00	47000.0
226.00	5572.0
227.00	81736.0
228.00	11656.0
229.00	15526.0
230.00	2791.0
231.00	7331.0
232.00	1339.0
233.00	1682.0
233.90	5478.0
235.00	4735.0

Instrument :
BNA_N
ClientSampleId :
DFTPP201

236.00	3457.0
237.00	6498.0
238.00	855.0
239.00	3159.0
240.00	2211.0
241.10	4403.0
242.00	11325.0
243.10	10374.0
244.00	140736.0
245.10	19952.0
246.00	33296.0
247.00	5975.0
248.00	1925.0
248.90	4352.0
250.00	1069.0
251.00	1368.0
251.90	1306.0
253.00	3511.0
255.00	730688.0
256.00	111736.0
257.00	7943.0
258.00	45680.0
259.00	5963.0
260.00	1109.0
260.90	1199.0
263.00	1024.0
263.90	1402.0
265.00	19264.0
265.90	2955.0
267.00	461.0
270.00	763.0
270.90	1325.0
272.00	1818.0
273.00	23944.0
273.90	60472.0
275.00	346624.0
276.00	43720.0
277.00	31504.0
278.00	5239.0
279.00	1031.0
281.10	913.0
282.00	1187.0
283.10	3713.0
284.10	1967.0
285.00	6459.0
285.90	716.0
288.00	326.0
289.10	910.0
290.10	1127.0
291.00	1247.0
291.90	1111.0
293.00	6657.0
294.00	1939.0
295.00	2824.0
296.00	101592.0
297.00	14048.0
297.90	869.0
300.90	1323.0
302.00	1635.0
303.00	12834.0
304.00	3735.0
304.90	518.0
308.00	1337.0
308.90	1098.0
310.00	1442.0
310.90	382.0
311.90	311.0
312.90	1228.0
314.10	4637.0
314.90	13046.0
316.00	5870.0
317.00	871.0
319.90	405.0
321.00	3551.0
322.00	1498.0
323.00	31648.0
324.00	5195.0
325.00	666.0
325.80	506.0
327.00	5468.0
328.00	2786.0
329.00	854.0
332.00	3062.0
333.00	3930.0
334.00	22576.0
335.00	5229.0
336.00	645.0
339.00	624.0
339.70	441.0
341.00	4168.0
342.00	949.0
346.00	6808.0
347.00	1117.0
352.00	10353.0
353.00	6569.0
354.00	11446.0

Instrument :
BNA_N
ClientSampleId :
DFTPP201

355.00	2068.0
356.10	334.0
357.00	323.0
358.90	566.0
360.00	351.0
364.90	45144.0
365.90	7733.0
366.90	562.0
370.10	1582.0
371.00	3199.0
372.00	16496.0
373.00	5022.0
374.00	598.0
376.90	447.0
382.00	319.0
383.00	4975.0
384.00	1650.0
385.10	751.0
389.90	2161.0
391.00	1543.0
392.00	1589.0
394.90	415.0
397.00	452.0
401.00	1239.0
402.00	8585.0
403.00	12305.0
404.00	2935.0
404.80	481.0
414.90	852.0
421.00	8875.0
422.00	10669.0
423.00	67376.0
424.00	16560.0
425.00	1395.0
439.70	452.0
441.00	182080.0
442.00	1068032.0
443.00	222720.0
444.00	18736.0
445.00	963.0

Instrument :
BNA_N
ClientSampleId :
DFTPP201