

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN082322\  
 Data File : BN021332.D  
 Acq On : 23 Aug 2022 17:13  
 Operator : CG/JU  
 Sample : PB147179BL  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 SBLK179

Quant Time: Aug 24 00:59:46 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN082222.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Aug 22 18:06:52 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|-------|----------|
| -----                       |        |      |          |       |       |          |
| Internal Standards          |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 8.097  | 152  | 2341     | 0.400 | ng/u1 | 0.00     |
| 4) Naphthalene-d8           | 10.927 | 136  | 7705     | 0.400 | ng/u1 | 0.00     |
| 9) Acenaphthene-d10         | 14.749 | 164  | 4323     | 0.400 | ng/u1 | # 0.00   |
| 13) Phenanthrene-d10        | 17.498 | 188  | 8899     | 0.400 | ng/u1 | 0.00     |
| 17) Chrysene-d12            | 21.688 | 240  | 6532     | 0.400 | ng/u1 | # 0.00   |
| 23) Perylene-d12            | 24.223 | 264  | 7264     | 0.400 | ng/u1 | 0.00     |
| System Monitoring Compounds |        |      |          |       |       |          |
| 3) 1,4-Dioxane-d8           | 3.380  | 96   | 17357    | 5.748 | ng/u1 | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.517 | 152  | 4047     | 0.347 | ng/u1 | 0.00     |
| 18) Fluoranthene-d10        | 19.524 | 212  | 8358     | 0.415 | ng/u1 | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN082322\  
Data File : BN021332.D  
Acq On : 23 Aug 2022 17:13  
Operator : CG/JU  
Sample : PB147179BL  
Misc :  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
SBLK179

Quant Time: Aug 24 00:59:46 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN082222.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Aug 22 18:06:52 2022  
Response via : Initial Calibration

