

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082620\
 Data File : BN011644.D
 Acq On : 26 Aug 2020 14:32
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 26 15:45:32 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 25 03:31:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	97	-0.02
2	1,4-Dioxane	0.400	0.434	-8.5	109	-0.02
3	n-Nitrosodimethylamine	0.400	0.474	-18.5	107	0.00
4 S	2-Fluorophenol	0.400	0.375	6.3	90	0.00
5 S	Phenol-d6	0.400	0.356	11.0	82	0.00
6	bis(2-Chloroethyl)ether	0.400	0.406	-1.5	98	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	86	-0.01
8 S	Nitrobenzene-d5	0.400	0.444	-11.0	98	-0.03
9	Naphthalene	0.400	0.401	-0.3	92	-0.01
10	Hexachlorobutadiene	0.400	0.416	-4.0	96	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.380	5.0	89	-0.01
12	2-Methylnaphthalene	0.400	0.383	4.3	88	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	77	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.350	12.5	69	0.00
15 S	2-Fluorobiphenyl	0.400	0.421	-5.2	85	-0.01
16	Acenaphthylene	0.400	0.356	11.0	75	-0.01
17	Acenaphthene	0.400	0.395	1.3	81	-0.01
18	Fluorene	0.400	0.381	4.8	80	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	74	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.503	-25.7#	112	-0.01
21	4-Bromophenyl-phenylether	0.400	0.391	2.3	78	-0.01
22	Hexachlorobenzene	0.400	0.418	-4.5	81	0.00
23	Atrazine	0.400	0.276	31.0#	72	-0.01
24	Pentachlorophenol	0.400	0.431	-7.7	84	-0.01
25	Phenanthrene	0.400	0.406	-1.5	80	0.00
26	Anthracene	0.400	0.368	8.0	72	-0.01
27 SURR	Fluoranthene-d10	0.400	0.394	1.5	80	0.00
28	Fluoranthene	0.400	0.407	-1.7	80	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	78	-0.01
30	Pyrene	0.400	0.380	5.0	79	0.00
31 S	Terphenyl-d14	0.400	0.384	4.0	78	0.00
32	Benzo(a)anthracene	0.400	0.392	2.0	80	0.00
33	Chrysene	0.400	0.414	-3.5	89	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.230	42.5#	55	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.436	-9.0	90	-0.01
36 I	Perylene-d12	0.400	0.400	0.0	81	0.00
37	Benzo(b)fluoranthene	0.400	0.411	-2.7	89	0.00
38	Benzo(k)fluoranthene	0.400	0.424	-6.0	91	0.00
39 C	Benzo(a)pyrene	0.400	0.391	2.3	84	0.00
40	Dibenzo(a,h)anthracene	0.400	0.418	-4.5	90	-0.02
41	Benzo(g,h,i)perylene	0.400	0.420	-5.0	90	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082620\
Data File : BN011644.D
Acq On : 26 Aug 2020 14:32
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 26 15:45:32 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 25 03:31:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)
----------	--------	-------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0