

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082620\
 Data File : BN011653.D
 Acq On : 26 Aug 2020 19:57
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 27 00:29:49 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 25 03:31:11 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	-0.02
2	1,4-Dioxane	0.620	0.681	-9.8	133	-0.02
3	n-Nitrosodimethylamine	0.653	0.736	-12.7	123	0.00
4 S	2-Fluorophenol	1.279	1.268	0.9	115	0.00
5 S	Phenol-d6	1.808	1.732	4.2	107	0.00
6	bis(2-Chloroethyl)ether	1.285	1.347	-4.8	122	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	108	-0.01
8 S	Nitrobenzene-d5	0.346	0.385	-11.3	124	-0.03
9	Naphthalene	1.137	1.139	-0.2	116	-0.01
10	Hexachlorobutadiene	0.238	0.243	-2.1	118	-0.03
11 SURR	2-Methylnaphthalene-d10	0.757	0.716	5.4	112	-0.01
12	2-Methylnaphthalene	0.839	0.799	4.8	110	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	-0.01
14 S	2,4,6-Tribromophenol	0.220	0.193	12.3	86	0.00
15 S	2-Fluorobiphenyl	1.557	1.613	-3.6	105	-0.03
16	Acenaphthylene	2.108	1.876	11.0	94	-0.01
17	Acenaphthene	1.337	1.306	2.3	101	-0.01
18	Fluorene	1.730	1.638	5.3	99	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	-0.01
20	4,6-Dinitro-2-methylphenol	0.046	0.060	-30.4#	143	-0.01
21	4-Bromophenyl-phenylether	0.233	0.223	4.3	95	-0.01
22	Hexachlorobenzene	0.270	0.278	-3.0	99	-0.01
23	Atrazine	0.230	0.150	34.8#	84	-0.01
24	Pentachlorophenol	0.119	0.120	-0.8	97	-0.01
25	Phenanthrene	1.243	1.238	0.4	97	0.00
26	Anthracene	1.224	1.093	10.7	87	-0.01
27 SURR	Fluoranthene-d10	1.304	1.243	4.7	97	0.00
28	Fluoranthene	1.506	1.469	2.5	95	-0.01
29 I	Chrysene-d12	1.000	1.000	0.0	91	-0.01
30	Pyrene	1.418	1.379	2.8	95	-0.01
31 S	Terphenyl-d14	1.022	1.013	0.9	94	-0.01
32	Benzo(a)anthracene	1.443	1.457	-1.0	97	0.00
33	Chrysene	1.375	1.451	-5.5	106	0.00
34	Bis(2-ethylhexyl)phthalate	1.027	0.688	33.0#	68	-0.01
35	Indeno(1,2,3-cd)pyrene	1.679	1.930	-14.9	111	-0.01
36 I	Perylene-d12	1.000	1.000	0.0	100	0.00
37	Benzo(b)fluoranthene	1.449	1.452	-0.2	108	0.00
38	Benzo(k)fluoranthene	1.441	1.496	-3.8	110	0.00
39 C	Benzo(a)pyrene	1.351	1.335	1.2	105	0.00
40	Dibenzo(a,h)anthracene	1.412	1.468	-4.0	110	-0.01
41	Benzo(g,h,i)perylene	1.370	1.429	-4.3	110	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082620\
Data File : BN011653.D
Acq On : 26 Aug 2020 19:57
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 27 00:29:49 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 25 03:31:11 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0