

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082919\
 Data File : BN007494.D
 Acq On : 29 Aug 2019 09:21
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 29 12:16:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 24 00:38:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	0.665	0.587	11.7	109	0.00
3	n-Nitrosodimethylamine	0.309	0.211	31.7#	84	0.00
4 S	2-Fluorophenol	1.490	1.166	21.7	106	0.00
5 S	Phenol-d6	1.953	1.551	20.6	117	0.00
6	bis(2-Chloroethyl)ether	1.445	1.393	3.6	124	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	148	-0.01
8 S	Nitrobenzene-d5	0.348	0.308	11.5	137	0.00
9	Naphthalene	1.143	1.124	1.7	142	0.00
10	Hexachlorobutadiene	0.234	0.209	10.7	127	0.00
11 SURR	2-Methylnaphthalene-d10	0.665	0.696	-4.7	153#	0.00
12	2-Methylnaphthalene	0.779	0.821	-5.4	152#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	175#	0.00
14 S	2,4,6-Tribromophenol	0.209	0.211	-1.0	180#	0.00
15 S	2-Fluorobiphenyl	1.613	1.495	7.3	158#	0.00
16	Acenaphthylene	2.367	2.111	10.8	154#	0.00
17	Acenaphthene	1.385	1.354	2.2	167#	0.00
18	Fluorene	1.751	1.756	-0.3	171#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	183#	0.00
20	4-Bromophenyl-phenylether	0.158	0.125	20.9	167#	0.00
21	Hexachlorobenzene	0.268	0.284	-6.0	191#	0.00
22	Pentachlorophenol	0.083	0.064	22.9	150	0.00
23	Phenanthrene	1.203	1.233	-2.5	187#	0.00
24	Anthracene	1.210	1.136	6.1	170#	0.00
25 SURR	Fluoranthene-d10	1.288	1.292	-0.3	177#	0.00
26	Fluoranthene	1.545	1.585	-2.6	182#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	189#	0.00
28	Pvrene	1.609	1.552	3.5	178#	0.00
29 S	Terphenyl-d14	1.047	1.007	3.8	176#	0.00
30	Benzo(a)anthracene	1.571	1.412	10.1	167#	0.00
31	Chrysene	1.516	1.487	1.9	178#	0.00
32	Bis(2-ethylhexyl)phthalate	1.177	0.823	30.1#	140	-0.01
33	Indeno(1,2,3-cd)pyrene	1.827	1.281	29.9#	128	0.00
34 I	Perylene-d12	1.000	1.000	0.0	154#	0.00
35	Benzo(b)fluoranthene	1.449	1.513	-4.4	157#	0.00
36	Benzo(k)fluoranthene	1.432	1.505	-5.1	156#	0.00
37 C	Benzo(a)pyrene	1.379	1.310	5.0	142	0.00
38	Dibenzo(a,h)anthracene	1.376	1.147	16.6	124	0.00
39	Benzo(a,h,i)perylene	1.371	1.183	13.7	129	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082919\
Data File : BN007494.D
Acq On : 29 Aug 2019 09:21
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 29 12:16:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 24 00:38:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			