

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082919\
 Data File : BN007504.D
 Acq On : 29 Aug 2019 16:18
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 29 17:02:34 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 29 12:18:45 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	153#	-0.02
2	1,4-Dioxane	0.665	0.644	3.2	143	0.00
3	n-Nitrosodimethylamine	0.309	0.253	18.1	121	0.00
4 S	2-Fluorophenol	1.490	1.297	13.0	140	-0.02
5 S	Phenol-d6	1.953	1.635	16.3	147	0.00
6	bis(2-Chloroethyl)ether	1.445	1.411	2.4	151#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	151#	-0.01
8 S	Nitrobenzene-d5	0.348	0.356	-2.3	162#	0.00
9	Naphthalene	1.143	1.108	3.1	143	-0.01
10	Hexachlorobutadiene	0.234	0.211	9.8	131	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.631	5.1	141	-0.01
12	2-Methylnaphthalene	0.779	0.746	4.2	141	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	150	-0.01
14 S	2,4,6-Tribromophenol	0.209	0.237	-13.4	173#	-0.01
15 S	2-Fluorobiphenyl	1.613	1.612	0.1	146	-0.01
16	Acenaphthylene	2.367	2.177	8.0	136	-0.01
17	Acenaphthene	1.385	1.324	4.4	140	-0.01
18	Fluorene	1.751	1.707	2.5	143	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	150	-0.01
20	4-Bromophenyl-phenylether	0.158	0.043	72.8#	47#	0.00
21	Hexachlorobenzene	0.268	0.270	-0.7	149	-0.01
22	Pentachlorophenol	0.083	0.099	-19.3	190#	-0.01
23	Phenanthrene	1.203	1.213	-0.8	150#	-0.01
24	Anthracene	1.210	1.184	2.1	145	0.00
25 SURR	Fluoranthene-d10	1.288	1.252	2.8	141	0.00
26	Fluoranthene	1.545	1.481	4.1	139	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	146	0.00
28	Pvrene	1.609	1.564	2.8	138	0.00
29 S	Terphenyl-d14	1.047	1.025	2.1	138	0.00
30	Benzo(a)anthracene	1.571	1.601	-1.9	146	0.00
31	Chrysene	1.516	1.516	0.0	140	-0.01
32	Bis(2-ethylhexyl)phthalate	1.177	1.191	-1.2	156#	0.00
33	Indeno(1,2,3-cd)pyrene	1.827	1.968	-7.7	151#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	156#	0.00
35	Benzo(b)fluoranthene	1.449	1.553	-7.2	163#	0.00
36	Benzo(k)fluoranthene	1.432	1.543	-7.8	162#	0.00
37 C	Benzo(a)pyrene	1.379	1.531	-11.0	168#	0.00
38	Dibenzo(a,h)anthracene	1.376	1.374	0.1	151#	0.00
39	Benzo(a,h,i)perylene	1.371	1.337	2.5	147	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN082919\
Data File : BN007504.D
Acq On : 29 Aug 2019 16:18
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 29 17:02:34 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			