

Data File : BN007506.D
Acq On : 29 Aug 2019 17:30
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 30 02:05:13 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	145	0.00
2	1,4-Dioxane	0.665	0.666	-0.2	140	0.00
3	n-Nitrosodimethylamine	0.309	0.244	21.0	110	0.00
4 S	2-Fluorophenol	1.490	1.281	14.0	132	0.00
5 S	Phenol-d6	1.953	1.561	20.1	133	0.00
6	bis(2-Chloroethyl)ether	1.445	1.409	2.5	143	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	144	-0.01
8 S	Nitrobenzene-d5	0.348	0.346	0.6	150	0.00
9	Naphthalene	1.143	1.122	1.8	138	0.00
10	Hexachlorobutadiene	0.234	0.216	7.7	128	0.00
11 SURR	2-Methylnaphthalene-d10	0.665	0.628	5.6	134	0.00
12	2-Methylnaphthalene	0.779	0.737	5.4	132	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	139	0.00
14 S	2,4,6-Tribromophenol	0.209	0.211	-1.0	142	0.00
15 S	2-Fluorobiphenyl	1.613	1.585	1.7	133	0.00
16	Acenaphthylene	2.367	2.149	9.2	124	0.00
17	Acenaphthene	1.385	1.326	4.3	130	0.00
18	Fluorene	1.751	1.651	5.7	128	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
20	4-Bromophenyl-phenylether	0.158	0.127	19.6	122	0.00
21	Hexachlorobenzene	0.268	0.286	-6.7	138	0.00
22	Pentachlorophenol	0.083	0.081	2.4	136	0.00
23	Phenanthrene	1.203	1.222	-1.6	133	0.00
24	Anthracene	1.210	1.140	5.8	122	0.00
25 SURR	Fluoranthene-d10	1.288	1.324	-2.8	130	0.00
26	Fluoranthene	1.545	1.586	-2.7	131	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	148	0.00
28	Pyrene	1.609	1.470	8.6	132	0.00
29 S	Terphenyl-d14	1.047	0.946	9.6	129	0.00
30	Benzo(a)anthracene	1.571	1.462	6.9	136	-0.01
31	Chrysene	1.516	1.462	3.6	137	-0.01
32	Bis(2-ethylhexyl)phthalate	1.177	0.946	19.6	126	-0.01
33	Indeno(1,2,3-cd)pyrene	1.827	1.807	1.1	141	0.00
34 I	Perylene-d12	1.000	1.000	0.0	151#	0.00
35	Benzo(b)fluoranthene	1.449	1.369	5.5	139	0.00
36	Benzo(k)fluoranthene	1.432	1.451	-1.3	147	0.00
37 C	Benzo(a)pyrene	1.379	1.337	3.0	142	0.00
38	Dibenzo(a,h)anthracene	1.376	1.314	4.5	139	-0.01
39	Benzo(g,h,i)perylene	1.371	1.346	1.8	144	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082919\
Evaluate Continuing Calibration Report

Data File : BN007506.D
Acq On : 29 Aug 2019 17:30
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 30 02:05:13 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			