

Data File : BN007506.D
Acq On : 29 Aug 2019 17:30
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 30 02:05:13 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	145	0.00
2	1,4-Dioxane	0.400	0.401	-0.3	140	0.00
3	n-Nitrosodimethylamine	0.400	0.316	21.0	110	0.00
4 S	2-Fluorophenol	0.400	0.367	8.3	132	0.00
5 S	Phenol-d6	0.400	0.350	12.5	133	0.00
6	bis(2-Chloroethyl)ether	0.400	0.390	2.5	143	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	144	-0.01
8 S	Nitrobenzene-d5	0.400	0.397	0.8	150	0.00
9	Naphthalene	0.400	0.393	1.8	138	0.00
10	Hexachlorobutadiene	0.400	0.369	7.8	128	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.378	5.5	134	0.00
12	2-Methylnaphthalene	0.400	0.379	5.3	132	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	139	0.00
14 S	2,4,6-Tribromophenol	0.400	0.403	-0.8	142	0.00
15 S	2-Fluorobiphenyl	0.400	0.393	1.8	133	0.00
16	Acenaphthylene	0.400	0.363	9.3	124	0.00
17	Acenaphthene	0.400	0.383	4.3	130	0.00
18	Fluorene	0.400	0.377	5.8	128	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	131	0.00
20	4-Bromophenyl-phenylether	0.400	0.322	19.5	122	0.00
21	Hexachlorobenzene	0.400	0.426	-6.5	138	0.00
22	Pentachlorophenol	0.400	0.388	3.0	136	0.00
23	Phenanthrene	0.400	0.406	-1.5	133	0.00
24	Anthracene	0.400	0.377	5.8	122	0.00
25 SURR	Fluoranthene-d10	0.400	0.411	-2.7	130	0.00
26	Fluoranthene	0.400	0.411	-2.7	131	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	148	0.00
28	Pyrene	0.400	0.366	8.5	132	0.00
29 S	Terphenyl-d14	0.400	0.362	9.5	129	0.00
30	Benzo(a)anthracene	0.400	0.372	7.0	136	-0.01
31	Chrysene	0.400	0.386	3.5	137	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.353	11.8	126	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.396	1.0	141	0.00
34 I	Perylene-d12	0.400	0.400	0.0	151	0.00
35	Benzo(b)fluoranthene	0.400	0.378	5.5	139	0.00
36	Benzo(k)fluoranthene	0.400	0.405	-1.3	147	0.00
37 C	Benzo(a)pyrene	0.400	0.388	3.0	142	0.00
38	Dibenzo(a,h)anthracene	0.400	0.382	4.5	139	-0.01
39	Benzo(g,h,i)perylene	0.400	0.393	1.8	144	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082919\
Evaluate Continuing Calibration Report

Data File : BN007506.D
Acq On : 29 Aug 2019 17:30
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 30 02:05:13 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				