

Data File : BN007518.D
Acq On : 30 Aug 2019 00:41
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 30 02:47:32 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	177#	0.00
2	1,4-Dioxane	0.665	0.564	15.2	144	0.00
3	n-Nitrosodimethylamine	0.309	0.224	27.5#	124	0.00
4 S	2-Fluorophenol	1.490	1.174	21.2	147	0.00
5 S	Phenol-d6	1.953	1.622	16.9	168#	0.00
6	bis(2-Chloroethyl)ether	1.445	1.425	1.4	176#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	207#	-0.01
8 S	Nitrobenzene-d5	0.348	0.327	6.0	203#	0.00
9	Naphthalene	1.143	1.198	-4.8	211#	0.00
10	Hexachlorobutadiene	0.234	0.211	9.8	179#	0.00
11 SURR	2-Methylnaphthalene-d10	0.665	0.662	0.5	203#	0.00
12	2-Methylnaphthalene	0.779	0.791	-1.5	204#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	222#	0.00
14 S	2,4,6-Tribromophenol	0.209	0.208	0.5	225#	0.00
15 S	2-Fluorobiphenyl	1.613	1.569	2.7	211#	0.00
16	Acenaphthylene	2.367	2.164	8.6	200#	0.00
17	Acenaphthene	1.385	1.343	3.0	210#	0.00
18	Fluorene	1.751	1.713	2.2	212#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	220#	0.00
20	4-Bromophenyl-phenylether	0.158	0.107	32.3#	173#	0.00
21	Hexachlorobenzene	0.268	0.276	-3.0	224#	0.00
22	Pentachlorophenol	0.083	0.068	18.1	193#	0.00
23	Phenanthrene	1.203	1.228	-2.1	224#	0.00
24	Anthracene	1.210	1.151	4.9	207#	0.00
25 SURR	Fluoranthene-d10	1.288	1.275	1.0	210#	0.00
26	Fluoranthene	1.545	1.531	0.9	212#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	211#	0.00
28	Pyrene	1.609	1.661	-3.2	213#	0.00
29 S	Terphenyl-d14	1.047	1.057	-1.0	206#	-0.01
30	Benzo(a)anthracene	1.571	1.535	2.3	203#	-0.01
31	Chrysene	1.516	1.450	4.4	193#	-0.01
32	Bis(2-ethylhexyl)phthalate	1.177	1.062	9.8	201#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.827	1.188	35.0#	132	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	167#	-0.02
35	Benzo(b)fluoranthene	1.449	1.449	0.0	163#	0.00
36	Benzo(k)fluoranthene	1.432	1.496	-4.5	167#	-0.01
37 C	Benzo(a)pyrene	1.379	1.312	4.9	154#	-0.02
38	Dibenzo(a,h)anthracene	1.376	1.099	20.1	129	-0.02
39	Benzo(g,h,i)perylene	1.371	1.119	18.4	132	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN082919\
Evaluate Continuing Calibration Report

Data File : BN007518.D
Acq On : 30 Aug 2019 00:41
Operator : HP/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 30 02:47:32 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 29 12:18:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			