

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083022\
 Data File : BN021461.D
 Acq On : 30 Aug 2022 18:08
 Operator : CG/JU
 Sample : N4298-11
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BPOW1-8-20220822

Quant Time: Aug 31 01:22:08 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 29 16:30:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.076	152	4603	0.400	ng	0.00
7) Naphthalene-d8	10.909	136	14369	0.400	ng	0.00
13) Acenaphthene-d10	14.729	164	7802	0.400	ng	0.00
19) Phenanthrene-d10	17.480	188	16600	0.400	ng	0.00
29) Chrysene-d12	21.664	240	11027	0.400	ng	# 0.00
35) Perylene-d12	24.194	264	7838	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.599	112	1420	0.158	ng	0.00
5) Phenol-d6	7.224	99	1017	0.089	ng	0.00
8) Nitrobenzene-d5	9.254	82	3283	0.338	ng	0.00
11) 2-Methylnaphthalene-d10	12.494	152	12659	0.391	ng	0.00
14) 2,4,6-Tribromophenol	16.219	330	692	0.270	ng	0.00
15) 2-Fluorobiphenyl	13.360	172	12210	0.358	ng	0.00
27) Fluoranthene-d10	19.510	212	17324	0.406	ng	0.00
31) Terphenyl-d14	20.097	244	11620	0.469	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.566	149	864	0.056	ng	Qvalue # 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083022\
Data File : BN021461.D
Acq On : 30 Aug 2022 18:08
Operator : CG/JU
Sample : N4298-11
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BPOW1-8-20220822

Quant Time: Aug 31 01:22:08 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 29 16:30:43 2022
Response via : Initial Calibration

