

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083022\
 Data File : BN021465.D
 Acq On : 30 Aug 2022 20:40
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 31 01:23:16 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 29 16:30:43 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.500	0.489	2.2	103	0.00
3	n-Nitrosodimethylamine	0.451	0.435	3.5	102	0.00
4 S	2-Fluorophenol	0.783	0.781	0.3	106	0.00
5 S	Phenol-d6	0.998	0.973	2.5	106	0.00
6	bis(2-Chloroethyl)ether	1.256	1.193	5.0	99	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.270	0.242	10.4	103	-0.01
9	Naphthalene	1.185	1.092	7.8	101	-0.01
10	Hexachlorobutadiene	0.205	0.188	8.3	99	-0.01
11 SURR	2-Methylnaphthalene-d10	0.902	0.886	1.8	114	0.00
12	2-Methylnaphthalene	0.150	0.143	4.7	117	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.131	0.110	16.0	118	0.00
15 S	2-Fluorobiphenyl	1.749	1.595	8.8	105	0.00
16	Acenaphthylene	1.878	1.590	15.3	106	0.00
17	Acenaphthene	1.321	1.190	9.9	109	0.00
18	Fluorene	1.630	1.478	9.3	110	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4,6-Dinitro-2-methylphenol	0.000	0.000	0.0	0#	-15.63#
21	4-Bromophenyl-phenylether	0.250	0.221	11.6	109	-0.01
22	Hexachlorobenzene	0.306	0.279	8.8	109	-0.01
23	Atrazine	0.146	0.123	15.8	116	0.00
24	Pentachlorophenol	0.065	0.051	21.5	115	0.00
25	Phenanthrene	1.381	1.253	9.3	109	0.00
26	Anthracene	1.097	0.959	12.6	113	0.00
27 SURR	Fluoranthene-d10	1.028	0.910	11.5	113	0.00
28	Fluoranthene	1.435	1.266	11.8	112	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
30	Pyrene	2.132	1.979	7.2	112	0.00
31 S	Terphenyl-d14	0.899	0.876	2.6	115	0.00
32	Benzo(a)anthracene	1.393	1.246	10.6	118	0.00
33	Chrysene	1.659	1.531	7.7	115	0.00
34	Bis(2-ethylhexyl)phthalate	0.561	0.518	7.7	122	0.00
35 I	Perylene-d12	1.000	1.000	0.0	133	0.00
36	Indeno(1,2,3-cd)pyrene	1.806	1.629	9.8	141	0.00
37	Benzo(b)fluoranthene	1.972	1.718	12.9	130	0.00
38	Benzo(k)fluoranthene	1.984	1.685	15.1	128	0.00
39 C	Benzo(a)pyrene	1.376	1.276	7.3	147	0.00
40	Dibenzo(a,h)anthracene	1.454	1.292	11.1	140	0.00
41	Benzo(g,h,i)perylene	1.595	1.299	18.6	125	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4