

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083022\
 Data File : BN021465.D
 Acq On : 30 Aug 2022 20:40
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 31 01:23:16 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 29 16:30:43 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	98	0.00
2	1,4-Dioxane	0.400	0.391	2.3	103	0.00
3	n-Nitrosodimethylamine	0.400	0.386	3.5	102	0.00
4 S	2-Fluorophenol	0.400	0.399	0.3	106	0.00
5 S	Phenol-d6	0.400	0.390	2.5	106	0.00
6	bis(2-Chloroethyl)ether	0.400	0.380	5.0	99	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	0.00
8 S	Nitrobenzene-d5	0.400	0.358	10.5	103	-0.01
9	Naphthalene	0.400	0.368	8.0	101	-0.01
10	Hexachlorobutadiene	0.400	0.367	8.3	99	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	114	0.00
12	2-Methylnaphthalene	0.400	0.417	-4.2	117	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.400	0.335	16.3	118	0.00
15 S	2-Fluorobiphenyl	0.400	0.365	8.8	105	0.00
16	Acenaphthylene	0.400	0.339	15.3	106	0.00
17	Acenaphthene	0.400	0.360	10.0	109	0.00
18	Fluorene	0.400	0.363	9.3	110	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.000	100.0#	0	-15.63#
21	4-Bromophenyl-phenylether	0.400	0.353	11.8	109	-0.01
22	Hexachlorobenzene	0.400	0.365	8.8	109	-0.01
23	Atrazine	0.400	0.337	15.8	116	0.00
24	Pentachlorophenol	0.400	0.426	-6.5	115	0.00
25	Phenanthrene	0.400	0.363	9.3	109	0.00
26	Anthracene	0.400	0.350	12.5	113	0.00
27 SURR	Fluoranthene-d10	0.400	0.354	11.5	113	0.00
28	Fluoranthene	0.400	0.353	11.8	112	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	114	0.00
30	Pyrene	0.400	0.371	7.3	112	0.00
31 S	Terphenyl-d14	0.400	0.390	2.5	115	0.00
32	Benzo(a)anthracene	0.400	0.358	10.5	118	0.00
33	Chrysene	0.400	0.369	7.8	115	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.369	7.8	122	0.00
35 I	Perylene-d12	0.400	0.400	0.0	133	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.361	9.8	141	0.00
37	Benzo(b)fluoranthene	0.400	0.349	12.8	130	0.00
38	Benzo(k)fluoranthene	0.400	0.340	15.0	128	0.00
39 C	Benzo(a)pyrene	0.400	0.371	7.3	147	0.00
40	Dibenzo(a,h)anthracene	0.400	0.356	11.0	140	0.00
41	Benzo(g,h,i)perylene	0.400	0.326	18.5	125	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4