



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083022\
 Data File : BN021450.D
 Acq On : 30 Aug 2022 09:52
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN082922.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Aug 29 16:30:43 2022






AutoFind: Scans 2485, 2486, 2487; Background Corrected with Scan 2478

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.5	101117	PASS
68	69	0.00	2	1.6	1698	PASS
69	198	0.00	100	38.4	106211	PASS
70	69	0.00	2	0.6	601	PASS
127	198	10	80	50.5	139693	PASS
197	198	0.00	2	0.6	1663	PASS
198	198	100	100	100.0	276885	PASS
199	198	5	9	6.7	18523	PASS
275	198	10	60	24.6	68176	PASS
365	198	1	100	3.1	8664	PASS
441	198	0.01	100	12.0	33205	PASS
442	442	50	100	100.0	234197	PASS
443	442	15	24	19.5	45637	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083022\
 Data File : BN021450.D
 Acq On : 30 Aug 2022 09:52
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Aug 31 02:53:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 31 02:53:53 2022
 Response via : Initial Calibration



TIC: BN021450.D\data.ms

(70) Pentachlorophenol (C)

17.122min (0.000) 21868.80 ng




response 298277

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	68.40
264.00	61.60	61.31
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN083022\
Data File : BN021450.D
Acq On : 30 Aug 2022 09:52
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Aug 31 02:53:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 31 02:53:53 2022
Response via : Initial Calibration



TIC: BN021450.D\data.ms

(77) Benzidine

19.722min (0.000) 0.00 ng

response 1949261

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.46
183.00	13.20	11.25
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/30/2022	BNA_N	<u>BN021450.D</u>
Compound Name	Response	Retention Time
DDT	930855	20.951
DDD	10517	20.51
DDE	593	20.004
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
11110	941965	1.18

Instrument :
BNA_N
ClientSampleId :
DFTPP