

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090120\
 Data File : BN011683.D
 Acq On : 01 Sep 2020 14:16
 Operator : CG/JU
 Sample : L3840-10
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AE8

Manual Integrations
 APPROVED

mohammad
 9/2/2020 1:59:47 PM

Quant Time: Sep 01 14:48:24 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN081220.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 01 13:14:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	1495	0.40	ng/ul	0.00
2) Naphthalene-d8	10.53	136	5464	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.38	164	3246	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.13	188	6939m	0.40	ng/ul	0.00
16) Chrysene-d12	21.32	240	6084	0.40	ng/ul	0.00
20) Perylene-d12	23.58	264	6855	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.12	152	2335	0.30	ng/ul	0.00
14) Fluoranthene-d10	19.16	212	6023	0.31	ng/ul	0.00

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN090120\
Data File : BN011683.D
Acq On : 01 Sep 2020 14:16
Operator : CG/JU
Sample : L3840-10
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AE8

Manual Integrations
APPROVED

mohammad
9/2/2020 1:59:47 PM

Quant Time: Sep 01 14:48:24 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN081220.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 01 13:14:23 2020
Response via : Initial Calibration

