

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092321\
 Data File : BN016413.D
 Acq On : 22 Sep 2021 17:18
 Operator : CG/JU
 Sample : SSTD01016
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD010216

Quant Time: Sep 23 02:00:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN092321MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 23 01:47:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.658	152	203170	20.000	ng/u1	0.00
20) Naphthalene-d8	10.434	136	883739	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.299	164	555728	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.051	188	1122957	20.000	ng/u1	0.00
79) Chrysene-d12	21.251	240	1128932	20.000	ng/u1	# 0.00
88) Perylene-d12	23.522	264	1096773	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	18538	4.761	ng/uL	0.00
4) Pyridine-d5	3.599	84	124185	10.943	ng/u1	0.00
7) Phenol-d5	6.834	99	145993	9.524	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.005	67	92608	11.331	ng/u1	0.00
11) 2-Chlorophenol-d4	7.187	132	121976	9.388	ng/u1	0.00
15) 4-Methylphenol-d8	8.364	113	114668	9.280	ng/u1	0.00
21) Nitrobenzene-d5	8.805	128	52209	8.097	ng/u1	0.00
24) 2-Nitrophenol-d4	9.522	143	50128	6.940	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.058	165	110502	8.383	ng/u1	0.00
31) 4-Chloroaniline-d4	10.575	131	168948	9.344	ng/u1	0.00
46) Dimethylphthalate-d6	13.716	166	346064	9.699	ng/u1	0.00
49) Acenaphthylene-d8	13.987	160	439639	9.406	ng/u1	0.00
54) 4-Nitrophenol-d4	14.504	143	54994	8.000	ng/u1	0.00
60) Fluorene-d10	15.293	176	305475	9.712	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.416	200	38823	6.338	ng/u1	0.00
73) Anthracene-d10	17.145	188	462827	9.495	ng/u1	0.00
81) Pyrene-d10	19.451	212	504778	9.430	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.369	264	503195	9.370	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	20564	4.691	ng/uL	99
5) Pyridine	3.617	79	127331	11.289	ng/u1	93
6) Benzaldehyde	6.805	77	95318	10.375	ng/u1	95
8) Phenol	6.858	94	158111	10.123	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.093	93	127758	10.781	ng/u1	99
12) 2-Chlorophenol	7.222	128	126340	9.672	ng/u1	96
13) 2-Methylphenol	8.099	108	111736	9.255	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.187	45	177761	11.358	ng/u1	99
16) Acetophenone	8.469	105	190720	10.151	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.458	70	89942	10.333	ng/u1	95
18) 4-Methylphenol	8.422	108	125324	9.602	ng/u1	94
19) Hexachloroethane	8.722	117	49290	9.555	ng/u1	95
22) Nitrobenzene	8.846	77	127016	9.665	ng/u1	93
23) Isophorone	9.369	82	239521	9.381	ng/u1	98
25) 2-Nitrophenol	9.552	139	59149	7.672	ng/u1	96
26) 2,4-Dimethylphenol	9.622	107	136222	9.482	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.863	93	167575	10.367	ng/u1	98
29) 2,4-Dichlorophenol	10.081	162	112399	8.673	ng/u1	96
30) Naphthalene	10.481	128	428931	9.946	ng/u1	99
32) 4-Chloroaniline	10.599	127	177672	9.663	ng/u1	100
33) Hexachlorobutadiene	10.775	225	72229	8.357	ng/u1	97
34) Caprolactam	11.358	113	28733	6.522	ng/u1	95
35) 4-Chloro-3-methylphenol	11.728	107	119084	8.886	ng/u1	99
36) 2-Methylnaphthalene	12.099	142	285802	9.718	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092321\
 Data File : BN016413.D
 Acq On : 22 Sep 2021 17:18
 Operator : CG/JU
 Sample : SST01016
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SST010216


Quant Time: Sep 23 02:00:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN092321MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 23 01:47:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.322	142	289370	9.625	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.475	216	145800	9.340	ng/ul	97
40) Hexachlorocyclopentadiene	12.452	237	78932	8.819	ng/ul	96
41) 2,4,6-Trichlorophenol	12.716	196	83304	8.327	ng/ul	99
42) 2,4,5-Trichlorophenol	12.787	196	91558	8.459	ng/ul	95
43) 1,1'-Biphenyl	13.128	154	384280	10.243	ng/ul	99
44) 2-Chloronaphthalene	13.163	162	297468	10.002	ng/ul	98
45) 2-Nitroaniline	13.375	65	59732	7.901	ng/ul	91
47) Dimethylphthalate	13.763	163	351345	9.849	ng/ul	100
48) 2,6-Dinitrotoluene	13.881	165	55347	7.313	ng/ul#	81
50) Acenaphthylene	14.016	152	472596	9.788	ng/ul	99
51) 3-Nitroaniline	14.204	138	62457	7.456	ng/ul	87
52) Acenaphthene	14.357	153	312764	10.149	ng/ul	99
53) 2,4-Dinitrophenol	14.410	184	24102	5.947	ng/ul	89
55) 4-Nitrophenol	14.516	109	42077	9.156	ng/ul	93
56) Dibenzofuran	14.699	168	448360	10.068	ng/ul	96
57) 2,4-Dinitrotoluene	14.669	165	82639	7.577	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	14.928	232	71382	8.045	ng/ul	92
59) Diethylphthalate	15.140	149	356900	10.114	ng/ul	97
61) Fluorene	15.351	166	357078	9.936	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.351	204	182375	9.984	ng/ul	94
63) 4-Nitroaniline	15.369	138	63690	7.310	ng/ul	89
66) 4,6-Dinitro-2-methylph...	15.428	198	41023	7.086	ng/ul	96
67) N-Nitrosodiphenylamine	15.569	169	300792	9.852	ng/ul	97
68) 4-Bromophenyl-phenylether	16.245	248	99655	9.114	ng/ul	92
69) Hexachlorobenzene	16.351	284	116888	9.182	ng/ul#	92
70) Atrazine	16.528	200	97148	8.482	ng/ul	97
71) Pentachlorophenol	16.698	266	54548	8.400	ng/ul	99
72) Phenanthrene	17.093	178	569133	10.123	ng/ul	99
74) Anthracene	17.181	178	573851	9.931	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.087	216	147458	9.302	ng/uL	98
76) Pentachlorobenzene	14.616	250	147637	9.413	ng/uL	99
77) Carbazole	17.457	167	496618	9.667	ng/ul	98
78) Di-n-butylphthalate	18.040	149	529408	8.808	ng/ul	99
80) Fluoranthene	19.116	202	631070	9.588	ng/ul	96
82) Pyrene	19.481	202	666625	9.832	ng/ul	98
83) Butylbenzylphthalate	20.410	149	191530	7.046	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.181	252	191500	8.411	ng/ul#	99
85) Benzo(a)anthracene	21.239	228	644017	9.528	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.192	149	323585	8.153	ng/ul#	96
87) Chrysene	21.286	228	642836	9.757	ng/ul	99
89) Di-n-octyl phthalate	22.086	149	468182	7.082	ng/ul	100
90) Benzo(b)fluoranthene	22.833	252	627260	9.664	ng/ul	97
91) Benzo(k)fluoranthene	22.880	252	607412	10.184	ng/ul	98
93) Benzo(a)pyrene	23.416	252	613641	9.784	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.516	276	561149	7.945	ng/ul	98
95) Dibenzo(a,h)anthracene	25.833	278	588041	9.671	ng/ul	97
96) Benzo(g,h,i)perylene	26.516	276	561149	9.373	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Sep 23 02:00:10 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN092321MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 23 01:47:54 2021
Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 4.691 ng/uL
RT: 3.240 min Scan# 26
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18


Instrument :
BNA_N
ClientSampleId :
SSTD010216





Tgt Ion: 88 Resp: 20564
Ion Ratio Lower Upper
88 100
43 29.8 23.2 34.8
58 66.7 54.3 81.5




#5
Pyridine
Concen: 11.289 ng/uL
RT: 3.617 min Scan# 90
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18



Tgt Ion: 79 Resp: 127331
Ion Ratio Lower Upper
79 100
52 77.0 66.8 100.2
51 41.9 30.7 46.1









#6
Benzaldehyde
Concen: 10.375 ng/ul
RT: 6.805 min Scan# 632
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion: 77 Resp: 95318
Ion Ratio Lower Upper
77 100
105 101.1 83.7 125.5
106 96.9 82.2 123.2



#8
Phenol
Concen: 10.123 ng/ul
RT: 6.858 min Scan# 641
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion: 94 Resp: 158111
Ion Ratio Lower Upper
94 100
65 26.0 21.5 32.3
66 33.6 28.6 42.8









#10
Bis(2-Chloroethyl)ether
Concen: 10.781 ng/ul
RT: 7.093 min Scan# 681
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:	93	Resp:	127758
Ion	Ratio	Lower	Upper
93	100		
63	75.6	59.3	88.9
95	32.7	26.3	39.5



#12
2-Chlorophenol
Concen: 9.672 ng/ul
RT: 7.222 min Scan# 703
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:	128	Resp:	126340
Ion	Ratio	Lower	Upper
128	100		
64	44.8	33.4	50.2
130	32.8	25.4	38.0









#13
2-Methylphenol
Concen: 9.255 ng/ul
RT: 8.099 min Scan# 852
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:108 Resp: 111736
Ion Ratio Lower Upper
108 100
107 93.1 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 11.358 ng/ul
RT: 8.187 min Scan# 867
Delta R.T. -0.006 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion: 45 Resp: 177761
Ion Ratio Lower Upper
45 100
77 14.1 11.0 16.4
79 11.0 8.4 12.6









#16
Acetophenone
Concen: 10.151 ng/ul
RT: 8.469 min Scan# 915
Delta R.T. -0.006 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:105 Resp: 190720
Ion Ratio Lower Upper
105 100
77 76.4 59.0 88.4
51 29.8 20.3 30.5



#17
N-Nitroso-di-n-propylamine
Concen: 10.333 ng/ul
RT: 8.458 min Scan# 913
Delta R.T. -0.006 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion: 70 Resp: 89942
Ion Ratio Lower Upper
70 100
42 60.1 45.1 67.7
101 10.2 8.9 13.3
130 22.8 19.9 29.9









#18
4-Methylphenol
Concen: 9.602 ng/ul
RT: 8.422 min Scan# 907
Delta R.T. -0.006 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:108 Resp: 125324
Ion Ratio Lower Upper
108 100
107 115.5 87.0 130.4



#19
Hexachloroethane
Concen: 9.555 ng/ul
RT: 8.722 min Scan# 958
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:117 Resp: 49290
Ion Ratio Lower Upper
117 100
201 99.0 82.6 123.8
199 60.3 52.0 78.0









#22
Nitrobenzene
Concen: 9.665 ng/ul
RT: 8.846 min Scan# 979
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion: 77 Resp: 127016
Ion Ratio Lower Upper
77 100
123 45.7 41.1 61.7
65 13.4 11.0 16.6



#23
Isophorone
Concen: 9.381 ng/ul
RT: 9.369 min Scan# 1068
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion: 82 Resp: 239521
Ion Ratio Lower Upper
82 100
95 7.3 6.1 9.1
138 18.5 15.8 23.8









#25
2-Nitrophenol
Concen: 7.672 ng/ul
RT: 9.552 min Scan# 1099
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:	139	Resp:	59149
Ion Ratio	Lower	Upper	
139	100		
65	47.2	35.8	53.6
109	26.5	23.4	35.0



#26
2,4-Dimethylphenol
Concen: 9.482 ng/ul
RT: 9.622 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:	107	Resp:	136222
Ion Ratio	Lower	Upper	
107	100		
121	53.5	41.8	62.8
122	91.0	71.1	106.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 10.367 ng/ul
 RT: 9.863 min Scan# 1152
 Delta R.T. 0.000 min
 Lab File: BN016413.D
 Acq: 22 Sep 2021 17:18




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010216


Tgt Ion: 93 Resp: 167575
 Ion Ratio Lower Upper
 93 100
 95 32.8 24.9 37.3
 123 11.8 9.4 14.0



#29
 2,4-Dichlorophenol
 Concen: 8.673 ng/ul
 RT: 10.081 min Scan# 1189
 Delta R.T. 0.000 min
 Lab File: BN016413.D
 Acq: 22 Sep 2021 17:18

Tgt Ion:162 Resp: 112399
 Ion Ratio Lower Upper
 162 100
 164 66.8 51.7 77.5
 98 35.6 26.3 39.5









#30
Naphthalene
Concen: 9.946 ng/ul
RT: 10.481 min Scan# 1257
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:128 Resp: 428931
Ion Ratio Lower Upper
128 100
129 10.8 9.0 13.4
127 12.5 10.2 15.4



#32
4-Chloroaniline
Concen: 9.663 ng/ul
RT: 10.599 min Scan# 1277
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:127 Resp: 177672
Ion Ratio Lower Upper
127 100
129 32.2 25.8 38.8









#33
Hexachlorobutadiene
Concen: 8.357 ng/ul
RT: 10.775 min Scan# 1307
Delta R.T. 0.006 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:225 Resp: 72229
Ion Ratio Lower Upper
225 100
223 60.1 49.8 74.6
227 65.1 49.8 74.6



#34
Caprolactam
Concen: 6.522 ng/ul
RT: 11.358 min Scan# 1406
Delta R.T. 0.006 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:113 Resp: 28733
Ion Ratio Lower Upper
113 100
55 163.7 123.8 185.8
56 115.6 89.4 134.2









#35
4-Chloro-3-methylphenol
Concen: 8.886 ng/ul
RT: 11.728 min Scan# 1469
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:107 Resp: 119084
Ion Ratio Lower Upper
107 100
144 27.6 21.8 32.8
142 85.0 67.3 100.9



#36
2-Methylnaphthalene
Concen: 9.718 ng/ul
RT: 12.099 min Scan# 1532
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:142 Resp: 285802
Ion Ratio Lower Upper
142 100
141 88.4 71.3 106.9





#37

1-Methylnaphthalene

Concen: 9.625 ng/ul

RT: 12.322 min Scan# 1570

Delta R.T. 0.000 min

Lab File: BN016413.D

Acq: 22 Sep 2021 17:18

Instrument :

BNA_N

ClientSampleId :

SSTD010216





Tgt Ion:142 Resp: 289370

Ion Ratio Lower Upper

142 100

141 93.6 71.9 107.9

116 3.6 2.7 4.1



#39

1,2,4,5-Tetrachlorobenzene

Concen: 9.340 ng/ul

RT: 12.475 min Scan# 1596

Delta R.T. 0.000 min

Lab File: BN016413.D

Acq: 22 Sep 2021 17:18

Tgt Ion:216 Resp: 145800




Ion Ratio Lower Upper


216 100

214 75.4 61.8 92.6

179 21.9 15.8 23.8

108 16.6 11.4 17.0









#40
Hexachlorocyclopentadiene
Concen: 8.819 ng/ul
RT: 12.452 min Scan# 1592
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:237 Resp: 78932
Ion Ratio Lower Upper
237 100
235 63.4 52.7 79.1
272 12.4 11.8 17.8



#41
2,4,6-Trichlorophenol
Concen: 8.327 ng/ul
RT: 12.716 min Scan# 1637
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:196 Resp: 83304
Ion Ratio Lower Upper
196 100
198 94.6 75.2 112.8
200 31.1 24.2 36.4









#42
2,4,5-Trichlorophenol
Concen: 8.459 ng/ul
RT: 12.787 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:196 Resp: 91558
Ion Ratio Lower Upper
196 100
198 95.8 72.5 108.7
200 28.3 24.6 37.0



#43
1,1'-Biphenyl
Concen: 10.243 ng/ul
RT: 13.128 min Scan# 1707
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:154 Resp: 384280
Ion Ratio Lower Upper
154 100
153 40.1 32.1 48.1
76 12.9 9.1 13.7









#44
2-Chloronaphthalene
Concen: 10.002 ng/ul
RT: 13.163 min Scan# 1713
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:162 Resp: 297468
Ion Ratio Lower Upper
162 100
164 31.6 26.2 39.4
127 39.0 30.1 45.1



#45
2-Nitroaniline
Concen: 7.901 ng/ul
RT: 13.375 min Scan# 1749
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion: 65 Resp: 59732
Ion Ratio Lower Upper
65 100
92 71.7 57.8 86.8
138 109.5 99.6 149.4









#47
Dimethylphthalate
Concen: 9.849 ng/ul
RT: 13.763 min Scan# 1815
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:163 Resp: 351345
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.1 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 7.313 ng/ul
RT: 13.881 min Scan# 1835
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:165 Resp: 55347
Ion Ratio Lower Upper
165 100
89 58.1 35.6 53.4#
121 24.7 14.8 22.2#









#50
Acenaphthylene
Concen: 9.788 ng/ul
RT: 14.016 min Scan# 1858
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:152 Resp: 472596
Ion Ratio Lower Upper
152 100
151 19.0 14.8 22.2
153 13.3 10.3 15.5



#51
3-Nitroaniline
Concen: 7.456 ng/ul
RT: 14.204 min Scan# 1890
Delta R.T. -0.006 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:138 Resp: 62457
Ion Ratio Lower Upper
138 100
108 11.6 9.0 13.4
92 117.3 82.3 123.5









#52
Acenaphthene
Concen: 10.149 ng/ul
RT: 14.357 min Scan# 1916
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:153 Resp: 312764
Ion Ratio Lower Upper
153 100
152 47.3 36.6 54.8
154 87.7 70.9 106.3



#53
2,4-Dinitrophenol
Concen: 5.947 ng/ul
RT: 14.410 min Scan# 1925
Delta R.T. -0.006 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:184 Resp: 24102
Ion Ratio Lower Upper
184 100
63 65.2 45.2 67.8
154 57.3 51.8 77.8









#55
4-Nitrophenol
Concen: 9.156 ng/ul
RT: 14.516 min Scan# 1943
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:109 Resp: 42077
Ion Ratio Lower Upper
109 100
139 146.2 129.5 194.3
65 136.7 109.0 163.4



#56
Dibenzofuran
Concen: 10.068 ng/ul
RT: 14.699 min Scan# 1974
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:168 Resp: 448360
Ion Ratio Lower Upper
168 100
139 36.5 27.6 41.4









#57
2,4-Dinitrotoluene
Concen: 7.577 ng/ul
RT: 14.669 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:165 Resp: 82639
Ion Ratio Lower Upper
165 100
63 42.3 30.4 45.6
182 3.1 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 8.045 ng/ul
RT: 14.928 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:232 Resp: 71382
Ion Ratio Lower Upper
232 100
131 42.8 29.1 43.7
130 2.2 1.7 2.5
166 29.2 21.6 32.4









#59
Diethylphthalate
Concen: 10.114 ng/ul
RT: 15.140 min Scan# 2049
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:149 Resp: 356900
Ion Ratio Lower Upper
149 100
177 21.5 19.0 28.4
150 12.4 10.2 15.4



#61
Fluorene
Concen: 9.936 ng/ul
RT: 15.351 min Scan# 2085
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:166 Resp: 357078
Ion Ratio Lower Upper
166 100
165 99.0 74.8 112.2
167 13.4 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 9.984 ng/ul
RT: 15.351 min Scan# 2085
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:204 Resp: 182375
Ion Ratio Lower Upper
204 100
206 32.4 27.6 41.4
141 62.5 45.5 68.3



#63
4-Nitroaniline
Concen: 7.310 ng/ul
RT: 15.369 min Scan# 2088
Delta R.T. -0.006 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:138 Resp: 63690
Ion Ratio Lower Upper
138 100
92 53.2 35.5 53.3
108 69.4 49.4 74.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 7.086 ng/ul
 RT: 15.428 min Scan# 2098
 Delta R.T. 0.000 min
 Lab File: BN016413.D
 Acq: 22 Sep 2021 17:18




Instrument :
 BNA_N
 ClientSampleId :
 SST010216


Tgt Ion	Ratio	Lower	Upper
198	100		
182	6.3	4.2	6.4
77	24.0	17.5	26.3



#67
 N-Nitrosodiphenylamine
 Concen: 9.852 ng/ul
 RT: 15.569 min Scan# 2122
 Delta R.T. 0.000 min
 Lab File: BN016413.D
 Acq: 22 Sep 2021 17:18

Tgt Ion	Ratio	Lower	Upper
169	100		
168	66.1	51.6	77.4
167	36.1	27.1	40.7









#68
4-Bromophenyl-phenylether
Concen: 9.114 ng/ul
RT: 16.245 min Scan# 2237
Delta R.T. -0.006 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:	248	Resp:	99655
Ion Ratio	Lower	Upper	
248	100		
250	98.6	76.4	114.6
141	84.1	57.0	85.6



#69
Hexachlorobenzene
Concen: 9.182 ng/ul
RT: 16.351 min Scan# 2255
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:	284	Resp:	116888
Ion Ratio	Lower	Upper	
284	100		
142	39.0	25.3	37.9#
249	31.3	24.3	36.5









#70
Atrazine
Concen: 8.482 ng/ul
RT: 16.528 min Scan# 2285
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:200 Resp: 97148
Ion Ratio Lower Upper
200 100
173 29.2 22.6 34.0
215 50.3 42.0 63.0



#71
Pentachlorophenol
Concen: 8.400 ng/ul
RT: 16.698 min Scan# 2314
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:266 Resp: 54548
Ion Ratio Lower Upper
266 100
264 63.3 50.7 76.1
268 64.4 49.8 74.6









#72
Phenanthrene
Concen: 10.123 ng/ul
RT: 17.093 min Scan# 2381
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:178 Resp: 569133
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.2
176 18.7 14.7 22.1



#74
Anthracene
Concen: 9.931 ng/ul
RT: 17.181 min Scan# 2396
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:178 Resp: 573851
Ion Ratio Lower Upper
178 100
179 15.1 12.4 18.6
176 18.1 14.2 21.4









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 9.302 ng/uL
 RT: 13.087 min Scan# 1700
 Delta R.T. 0.000 min
 Lab File: BN016413.D
 Acq: 22 Sep 2021 17:18




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010216


Tgt Ion:	216	Resp:	147458
Ion Ratio	Lower	Upper	
216	100		
214	78.2	60.6	90.8
218	50.2	40.2	60.4



#76
 Pentachlorobenzene
 Concen: 9.413 ng/uL
 RT: 14.616 min Scan# 1960
 Delta R.T. 0.000 min
 Lab File: BN016413.D
 Acq: 22 Sep 2021 17:18

Tgt Ion:	250	Resp:	147637
Ion Ratio	Lower	Upper	
250	100		
248	64.3	51.1	76.7
252	66.3	52.1	78.1









#77
 Carbazole
 Concen: 9.667 ng/ul
 RT: 17.457 min Scan# 2443
 Delta R.T. 0.000 min
 Lab File: BN016413.D
 Acq: 22 Sep 2021 17:18




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010216


Tgt Ion:	167	Resp:	496618
Ion	Ratio	Lower	Upper
167	100		
166	21.8	16.8	25.2
139	12.8	9.8	14.6



#78
 Di-n-butylphthalate
 Concen: 8.808 ng/ul
 RT: 18.040 min Scan# 2542
 Delta R.T. 0.000 min
 Lab File: BN016413.D
 Acq: 22 Sep 2021 17:18

Tgt Ion:	149	Resp:	529408
Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.3	10.9
104	4.9	3.5	5.3









#80
Fluoranthene
Concen: 9.588 ng/ul
RT: 19.116 min Scan# 2725
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:202 Resp: 631070
Ion Ratio Lower Upper
202 100
101 12.5 8.7 13.1
100 9.3 6.6 9.8



#82
Pyrene
Concen: 9.832 ng/ul
RT: 19.481 min Scan# 2787
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:202 Resp: 666625
Ion Ratio Lower Upper
202 100
101 14.3 10.5 15.7
100 11.4 8.6 13.0









#83
Butylbenzylphthalate
Concen: 7.046 ng/ul
RT: 20.410 min Scan# 2945
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:149 Resp: 191530
Ion Ratio Lower Upper
149 100
91 70.0 50.6 76.0
206 20.8 18.5 27.7



#84
3,3'-Dichlorobenzidine
Concen: 8.411 ng/ul
RT: 21.181 min Scan# 3076
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:252 Resp: 191500
Ion Ratio Lower Upper
252 100
254 61.0 48.6 72.8
126 11.9 7.9 11.9#









#85
Benzo(a)anthracene
Concen: 9.528 ng/ul
RT: 21.239 min Scan# 3086
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:228 Resp: 644017
Ion Ratio Lower Upper
228 100
229 19.8 15.7 23.5
226 27.2 20.6 31.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 8.153 ng/ul
RT: 21.192 min Scan# 3078
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:149 Resp: 323585
Ion Ratio Lower Upper
149 100
167 28.1 24.2 36.4
279 3.9 4.2 6.2#









#87
Chrysene
Concen: 9.757 ng/ul
RT: 21.286 min Scan# 3094
Delta R.T. -0.006 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:228 Resp: 642836
Ion Ratio Lower Upper
228 100
226 29.6 23.4 35.0
229 19.5 14.9 22.3



#89
Di-n-octyl phthalate
Concen: 7.082 ng/ul
RT: 22.086 min Scan# 3230
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:149 Resp: 468182









#90
Benzo(b)fluoranthene
Concen: 9.664 ng/ul
RT: 22.833 min Scan# 3357
Delta R.T. -0.006 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:252 Resp: 627260
Ion Ratio Lower Upper
252 100
253 23.1 17.6 26.4
125 11.4 7.8 11.8



#91
Benzo(k)fluoranthene
Concen: 10.184 ng/ul
RT: 22.880 min Scan# 3365
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:252 Resp: 607412
Ion Ratio Lower Upper
252 100
253 22.1 17.0 25.4
125 10.6 7.8 11.6









#93
Benzo(a)pyrene
Concen: 9.784 ng/ul
RT: 23.416 min Scan# 3456
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18




Instrument :
BNA_N
ClientSampleId :
SSTD010216


Tgt Ion:252 Resp: 613641
Ion Ratio Lower Upper
252 100
253 21.0 17.4 26.0
125 12.1 8.9 13.3



#94
Indeno(1,2,3-cd)pyrene
Concen: 7.945 ng/ul
RT: 26.516 min Scan# 3983
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:276 Resp: 561149
Ion Ratio Lower Upper
276 100
138 25.6 19.4 29.0
277 23.8 18.6 27.8









#95
Dibenzo(a,h)anthracene
Concen: 9.671 ng/ul
RT: 25.833 min Scan# 3867
Delta R.T. -0.006 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Instrument :
BNA_N
ClientSampleId :
SSTD010216

Tgt Ion:278 Resp: 588041
Ion Ratio Lower Upper
278 100
139 17.4 12.2 18.4
279 23.3 18.2 27.2



#96
Benzo(g,h,i)perylene
Concen: 9.373 ng/ul
RT: 26.516 min Scan# 3983
Delta R.T. 0.000 min
Lab File: BN016413.D
Acq: 22 Sep 2021 17:18

Tgt Ion:276 Resp: 561149
Ion Ratio Lower Upper
276 100
138 25.6 17.4 26.0
277 23.8 18.4 27.6

