

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092321\
 Data File : BN016415.D
 Acq On : 22 Sep 2021 18:32
 Operator : CG/JU
 Sample : SSTD04018
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD040218

Quant Time: Sep 23 02:00:47 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN092321MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 23 01:47:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.658	152	219961	20.000	ng/u1	0.00
20) Naphthalene-d8	10.434	136	956201	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.298	164	577765	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.051	188	1164349	20.000	ng/u1	0.00
79) Chrysene-d12	21.257	240	1120034	20.000	ng/u1	0.00
88) Perylene-d12	23.522	264	1137952	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	82037	19.462	ng/uL	0.00
4) Pyridine-d5	3.599	84	586648	47.750	ng/u1	0.00
7) Phenol-d5	6.834	99	717592	43.238	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.005	67	416001	47.013	ng/u1	0.00
11) 2-Chlorophenol-d4	7.193	132	588209	41.815	ng/u1	0.00
15) 4-Methylphenol-d8	8.369	113	566340	42.333	ng/u1	0.00
21) Nitrobenzene-d5	8.810	128	269498	38.626	ng/u1	0.00
24) 2-Nitrophenol-d4	9.522	143	278885	35.685	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.057	165	545603	38.256	ng/u1	0.00
31) 4-Chloroaniline-d4	10.575	131	796629	40.720	ng/u1	0.00
46) Dimethylphthalate-d6	13.722	166	1560834	42.078	ng/u1	0.00
49) Acenaphthylene-d8	13.987	160	2035753	41.893	ng/u1	0.00
54) 4-Nitrophenol-d4	14.510	143	298083	41.707	ng/u1	0.00
60) Fluorene-d10	15.298	176	1334747	40.819	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.422	200	230544	36.299	ng/u1	0.00
73) Anthracene-d10	17.151	188	2027192	40.111	ng/u1	0.00
81) Pyrene-d10	19.451	212	2212917	41.670	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.380	264	2358085	42.323	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	87791	18.497	ng/uL	94
5) Pyridine	3.617	79	587417	48.102	ng/u1	92
6) Benzaldehyde	6.805	77	445176	44.756	ng/u1	95
8) Phenol	6.864	94	742061	43.883	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.093	93	577687	45.028	ng/u1	97
12) 2-Chlorophenol	7.222	128	596020	42.145	ng/u1	97
13) 2-Methylphenol	8.099	108	555267	42.482	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.193	45	791381	46.705	ng/u1	99
16) Acetophenone	8.475	105	881786	43.350	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.469	70	438102	46.489	ng/u1	98
18) 4-Methylphenol	8.428	108	599565	42.429	ng/u1	96
19) Hexachloroethane	8.722	117	231156	41.391	ng/u1	96
22) Nitrobenzene	8.852	77	624355	43.910	ng/u1	97
23) Isophorone	9.375	82	1211961	43.869	ng/u1	96
25) 2-Nitrophenol	9.558	139	312960	37.517	ng/u1	97
26) 2,4-Dimethylphenol	9.622	107	651124	41.889	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.863	93	773224	44.211	ng/u1	99
29) 2,4-Dichlorophenol	10.087	162	539984	38.509	ng/u1	97
30) Naphthalene	10.481	128	1901581	40.752	ng/u1	99
32) 4-Chloroaniline	10.599	127	822939	41.366	ng/u1	100
33) Hexachlorobutadiene	10.769	225	329652	35.252	ng/u1	99
34) Caprolactam	11.363	113	97264	20.406	ng/u1	94
35) 4-Chloro-3-methylphenol	11.734	107	585578	40.384	ng/u1	99
36) 2-Methylnaphthalene	12.104	142	1281735	40.278	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092321\
 Data File : BN016415.D
 Acq On : 22 Sep 2021 18:32
 Operator : CG/JU
 Sample : SST04018
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST040218

Quant Time: Sep 23 02:00:47 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN092321MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 23 01:47:54 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.322	142	1294279	39.788	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.475	216	650906	40.107	ng/ul#	98
40) Hexachlorocyclopentadiene	12.457	237	400967	43.089	ng/ul	96
41) 2,4,6-Trichlorophenol	12.722	196	409577	39.379	ng/ul	99
42) 2,4,5-Trichlorophenol	12.787	196	450839	40.064	ng/ul	97
43) 1,1'-Biphenyl	13.128	154	1692626	43.396	ng/ul	98
44) 2-Chloronaphthalene	13.163	162	1275261	41.246	ng/ul	96
45) 2-Nitroaniline	13.381	65	341898	43.498	ng/ul	95
47) Dimethylphthalate	13.769	163	1571810	42.379	ng/ul	100
48) 2,6-Dinitrotoluene	13.887	165	309509	39.336	ng/ul#	88
50) Acenaphthylene	14.016	152	2121819	42.267	ng/ul	99
51) 3-Nitroaniline	14.210	138	346990	39.841	ng/ul	90
52) Acenaphthene	14.363	153	1358503	42.403	ng/ul	97
53) 2,4-Dinitrophenol	14.416	184	158206	37.545	ng/ul	94
55) 4-Nitrophenol	14.528	109	214294	44.850	ng/ul	95
56) Dibenzofuran	14.698	168	1909192	41.238	ng/ul	94
57) 2,4-Dinitrotoluene	14.675	165	443444	39.105	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.928	232	354330	38.411	ng/ul#	86
59) Diethylphthalate	15.145	149	1561582	42.566	ng/ul	99
61) Fluorene	15.351	166	1523891	40.785	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.357	204	743485	39.149	ng/ul	96
63) 4-Nitroaniline	15.381	138	346283	38.228	ng/ul	91
66) 4,6-Dinitro-2-methylph...	15.434	198	236198	39.349	ng/ul#	95
67) N-Nitrosodiphenylamine	15.569	169	1326619	41.906	ng/ul	98
68) 4-Bromophenyl-phenylether	16.251	248	453964	40.040	ng/ul	94
69) Hexachlorobenzene	16.351	284	512098	38.799	ng/ul	93
70) Atrazine	16.534	200	474515	39.957	ng/ul	95
71) Pentachlorophenol	16.698	266	296488	44.037	ng/ul	97
72) Phenanthrene	17.092	178	2391737	41.030	ng/ul	99
74) Anthracene	17.187	178	2454313	40.964	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.087	216	677723	41.231	ng/ul	97
76) Pentachlorobenzene	14.616	250	639906	39.349	ng/ul	96
77) Carbazole	17.457	167	2223760	41.747	ng/ul	99
78) Di-n-butylphthalate	18.045	149	2623134	42.091	ng/ul	99
80) Fluoranthene	19.122	202	2771615	42.444	ng/ul	97
82) Pyrene	19.486	202	2855208	42.447	ng/ul	99
83) Butylbenzylphthalate	20.410	149	1116152	41.388	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.186	252	963961	42.673	ng/ul#	92
85) Benzo(a)anthracene	21.239	228	2756729	41.108	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.198	149	1767107	44.877	ng/ul	95
87) Chrysene	21.292	228	2678501	40.979	ng/ul	98
89) Di-n-octyl phthalate	22.086	149	2862732	41.739	ng/ul	100
90) Benzo(b)fluoranthene	22.839	252	2774487	41.197	ng/ul	98
91) Benzo(k)fluoranthene	22.839	252	2774487	44.836	ng/ul	96
93) Benzo(a)pyrene	23.427	252	2739548	42.099	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.533	276	2621477	35.771	ng/ul	99
95) Dibenzo(a,h)anthracene	25.845	278	2658682	42.145	ng/ul#	96
96) Benzo(g,h,i)perylene	26.533	276	2621477	42.201	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092321\
 Data File : BN016415.D
 Acq On : 22 Sep 2021 18:32
 Operator : CG/JU
 Sample : SST04018
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST040218

Quant Time: Sep 23 02:00:47 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN092321MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 23 01:47:54 2021
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 18.497 ng/uL
RT: 3.240 min Scan# 26
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion: 88 Resp: 87791
Ion Ratio Lower Upper
88 100
43 27.0 23.2 34.8
58 73.9 54.3 81.5



#5
Pyridine
Concen: 48.102 ng/uL
RT: 3.617 min Scan# 90
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion: 79 Resp: 587417
Ion Ratio Lower Upper
79 100
52 74.7 66.8 100.2
51 35.0 30.7 46.1









#6
Benzaldehyde
Concen: 44.756 ng/ul
RT: 6.805 min Scan# 632
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion: 77 Resp: 445176
Ion Ratio Lower Upper
77 100
105 100.9 83.7 125.5
106 96.1 82.2 123.2



#8
Phenol
Concen: 43.883 ng/ul
RT: 6.864 min Scan# 642
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion: 94 Resp: 742061
Ion Ratio Lower Upper
94 100
65 26.5 21.5 32.3
66 34.3 28.6 42.8





#10

Bis(2-Chloroethyl)ether

Concen: 45.028 ng/ul

RT: 7.093 min Scan# 681

Delta R.T. 0.000 min

Lab File: BN016415.D

Acq: 22 Sep 2021 18:32

Instrument :

BNA_N

ClientSampleId :

SSTD040218





Tgt Ion: 93 Resp: 577687

Ion Ratio Lower Upper

93 100

63 77.2 59.3 88.9

95 32.8 26.3 39.5



#12

2-Chlorophenol

Concen: 42.145 ng/ul

RT: 7.222 min Scan# 703

Delta R.T. 0.000 min

Lab File: BN016415.D

Acq: 22 Sep 2021 18:32




Tgt Ion:128 Resp: 596020


Ion Ratio Lower Upper

128 100

64 45.0 33.4 50.2

130 32.1 25.4 38.0









#13
2-Methylphenol
Concen: 42.482 ng/ul
RT: 8.099 min Scan# 852
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:108 Resp: 555267
Ion Ratio Lower Upper
108 100
107 88.6 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 46.705 ng/ul
RT: 8.193 min Scan# 868
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion: 45 Resp: 791381
Ion Ratio Lower Upper
45 100
77 14.1 11.0 16.4
79 10.7 8.4 12.6









#16
Acetophenone
Concen: 43.350 ng/ul
RT: 8.475 min Scan# 916
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:105 Resp: 881786
Ion Ratio Lower Upper
105 100
77 76.2 59.0 88.4
51 25.9 20.3 30.5



#17
N-Nitroso-di-n-propylamine
Concen: 46.489 ng/ul
RT: 8.469 min Scan# 915
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion: 70 Resp: 438102
Ion Ratio Lower Upper
70 100
42 55.4 45.1 67.7
101 10.0 8.9 13.3
130 23.0 19.9 29.9









#18
4-Methylphenol
Concen: 42.429 ng/ul
RT: 8.428 min Scan# 908
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:108 Resp: 599565
Ion Ratio Lower Upper
108 100
107 112.7 87.0 130.4



#19
Hexachloroethane
Concen: 41.391 ng/ul
RT: 8.722 min Scan# 958
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:117 Resp: 231156
Ion Ratio Lower Upper
117 100
201 98.5 82.6 123.8
199 61.7 52.0 78.0









#22
Nitrobenzene
Concen: 43.910 ng/ul
RT: 8.852 min Scan# 980
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion: 77 Resp: 624355
Ion Ratio Lower Upper
77 100
123 48.6 41.1 61.7
65 13.2 11.0 16.6



#23
Isophorone
Concen: 43.869 ng/ul
RT: 9.375 min Scan# 1069
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion: 82 Resp: 1211961
Ion Ratio Lower Upper
82 100
95 6.6 6.1 9.1
138 18.0 15.8 23.8









#25
2-Nitrophenol
Concen: 37.517 ng/ul
RT: 9.558 min Scan# 1100
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:139 Resp: 312960
Ion Ratio Lower Upper
139 100
65 45.9 35.8 53.6
109 27.1 23.4 35.0



#26
2,4-Dimethylphenol
Concen: 41.889 ng/ul
RT: 9.622 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:107 Resp: 651124
Ion Ratio Lower Upper
107 100
121 52.2 41.8 62.8
122 88.1 71.1 106.7









#27

Bis(2-Chloroethoxy)methane
Concen: 44.211 ng/ul
RT: 9.863 min Scan# 1152
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Instrument :
BNA_N
ClientSampleId :
SSTD040218




Tgt Ion: 93 Resp: 773224
Ion Ratio Lower Upper
93 100
95 32.2 24.9 37.3
123 11.7 9.4 14.0




#29

2,4-Dichlorophenol
Concen: 38.509 ng/ul
RT: 10.087 min Scan# 1190
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:162 Resp: 539984
Ion Ratio Lower Upper
162 100
164 66.5 51.7 77.5
98 36.0 26.3 39.5









#30
Naphthalene
Concen: 40.752 ng/ul
RT: 10.481 min Scan# 1257
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:128 Resp: 1901581
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.4
127 13.0 10.2 15.4



#32
4-Chloroaniline
Concen: 41.366 ng/ul
RT: 10.599 min Scan# 1277
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:127 Resp: 822939
Ion Ratio Lower Upper
127 100
129 32.0 25.8 38.8









#33
Hexachlorobutadiene
Concen: 35.252 ng/ul
RT: 10.769 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:225 Resp: 329652
Ion Ratio Lower Upper
225 100
223 60.7 49.8 74.6
227 61.5 49.8 74.6



#34
Caprolactam
Concen: 20.406 ng/ul
RT: 11.363 min Scan# 1407
Delta R.T. 0.012 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:113 Resp: 97264
Ion Ratio Lower Upper
113 100
55 162.7 123.8 185.8
56 118.8 89.4 134.2









#35
4-Chloro-3-methylphenol
Concen: 40.384 ng/ul
RT: 11.734 min Scan# 1470
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:107 Resp: 585578
Ion Ratio Lower Upper
107 100
144 26.5 21.8 32.8
142 82.9 67.3 100.9



#36
2-Methylnaphthalene
Concen: 40.278 ng/ul
RT: 12.104 min Scan# 1533
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:142 Resp: 1281735
Ion Ratio Lower Upper
142 100
141 87.1 71.3 106.9









#37
1-Methylnaphthalene
Concen: 39.788 ng/ul
RT: 12.322 min Scan# 1570
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:142 Resp: 1294279
Ion Ratio Lower Upper
142 100
141 92.1 71.9 107.9
116 3.8 2.7 4.1



#39
1,2,4,5-Tetrachlorobenzene
Concen: 40.107 ng/ul
RT: 12.475 min Scan# 1596
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:216 Resp: 650906
Ion Ratio Lower Upper
216 100
214 77.8 61.8 92.6
179 20.9 15.8 23.8
108 17.4 11.4 17.0#









#40
Hexachlorocyclopentadiene
Concen: 43.089 ng/ul
RT: 12.457 min Scan# 1593
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:237 Resp: 400967
Ion Ratio Lower Upper
237 100
235 62.8 52.7 79.1
272 12.7 11.8 17.8



#41
2,4,6-Trichlorophenol
Concen: 39.379 ng/ul
RT: 12.722 min Scan# 1638
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:196 Resp: 409577
Ion Ratio Lower Upper
196 100
198 95.3 75.2 112.8
200 30.2 24.2 36.4









#42
2,4,5-Trichlorophenol
Concen: 40.064 ng/ul
RT: 12.787 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:196 Resp: 450839
Ion Ratio Lower Upper
196 100
198 94.5 72.5 108.7
200 30.2 24.6 37.0



#43
1,1'-Biphenyl
Concen: 43.396 ng/ul
RT: 13.128 min Scan# 1707
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:154 Resp: 1692626
Ion Ratio Lower Upper
154 100
153 39.4 32.1 48.1
76 13.1 9.1 13.7









#44
2-Chloronaphthalene
Concen: 41.246 ng/ul
RT: 13.163 min Scan# 1713
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:162 Resp: 1275261
Ion Ratio Lower Upper
162 100
164 33.3 26.2 39.4
127 41.7 30.1 45.1



#45
2-Nitroaniline
Concen: 43.498 ng/ul
RT: 13.381 min Scan# 1750
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion: 65 Resp: 341898
Ion Ratio Lower Upper
65 100
92 73.6 57.8 86.8
138 116.7 99.6 149.4









#47
Dimethylphthalate
Concen: 42.379 ng/ul
RT: 13.769 min Scan# 1816
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:163 Resp: 1571810
Ion Ratio Lower Upper
163 100
194 4.2 3.3 4.9
164 10.3 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 39.336 ng/ul
RT: 13.887 min Scan# 1836
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:165 Resp: 309509
Ion Ratio Lower Upper
165 100
89 53.9 35.6 53.4#
121 21.4 14.8 22.2









#50
Acenaphthylene
Concen: 42.267 ng/ul
RT: 14.016 min Scan# 1858
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:152 Resp: 2121819
Ion Ratio Lower Upper
152 100
151 19.5 14.8 22.2
153 12.8 10.3 15.5



#51
3-Nitroaniline
Concen: 39.841 ng/ul
RT: 14.210 min Scan# 1891
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:138 Resp: 346990
Ion Ratio Lower Upper
138 100
108 10.9 9.0 13.4
92 114.0 82.3 123.5









#52
Acenaphthene
Concen: 42.403 ng/ul
RT: 14.363 min Scan# 1917
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:153 Resp: 1358503
Ion Ratio Lower Upper
153 100
152 47.3 36.6 54.8
154 85.3 70.9 106.3



#53
2,4-Dinitrophenol
Concen: 37.545 ng/ul
RT: 14.416 min Scan# 1926
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:184 Resp: 158206
Ion Ratio Lower Upper
184 100
63 62.6 45.2 67.8
154 61.5 51.8 77.8









#55
4-Nitrophenol
Concen: 44.850 ng/ul
RT: 14.528 min Scan# 1945
Delta R.T. 0.012 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:109 Resp: 214294
Ion Ratio Lower Upper
109 100
139 151.7 129.5 194.3
65 133.2 109.0 163.4



#56
Dibenzofuran
Concen: 41.238 ng/ul
RT: 14.698 min Scan# 1974
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:168 Resp: 1909192
Ion Ratio Lower Upper
168 100
139 37.8 27.6 41.4









#57
2,4-Dinitrotoluene
Concen: 39.105 ng/ul
RT: 14.675 min Scan# 1970
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:165 Resp: 443444
Ion Ratio Lower Upper
165 100
63 40.5 30.4 45.6
182 2.8 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 38.411 ng/ul
RT: 14.928 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:232 Resp: 354330
Ion Ratio Lower Upper
232 100
131 46.5 29.1 43.7#
130 2.4 1.7 2.5
166 32.5 21.6 32.4#









#59
Diethylphthalate
Concen: 42.566 ng/ul
RT: 15.145 min Scan# 2050
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:149 Resp: 1561582
Ion Ratio Lower Upper
149 100
177 22.9 19.0 28.4
150 12.7 10.2 15.4



#61
Fluorene
Concen: 40.785 ng/ul
RT: 15.351 min Scan# 2085
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:166 Resp: 1523891
Ion Ratio Lower Upper
166 100
165 95.5 74.8 112.2
167 13.3 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 39.149 ng/ul
RT: 15.357 min Scan# 2086
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:204 Resp: 743485
Ion Ratio Lower Upper
204 100
206 33.5 27.6 41.4
141 60.4 45.5 68.3



#63
4-Nitroaniline
Concen: 38.228 ng/ul
RT: 15.381 min Scan# 2090
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:138 Resp: 346283
Ion Ratio Lower Upper
138 100
92 47.7 35.5 53.3
108 70.5 49.4 74.2









#66
4,6-Dinitro-2-methylphenol
Concen: 39.349 ng/ul
RT: 15.434 min Scan# 2099
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:198 Resp: 236198
Ion Ratio Lower Upper
198 100
182 4.0 4.2 6.4#
77 24.5 17.5 26.3



#67
N-Nitrosodiphenylamine
Concen: 41.906 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:169 Resp: 1326619
Ion Ratio Lower Upper
169 100
168 66.5 51.6 77.4
167 35.0 27.1 40.7









#68
4-Bromophenyl-phenylether
Concen: 40.040 ng/ul
RT: 16.251 min Scan# 2238
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:248 Resp: 453964
Ion Ratio Lower Upper
248 100
250 100.0 76.4 114.6
141 77.4 57.0 85.6



#69
Hexachlorobenzene
Concen: 38.799 ng/ul
RT: 16.351 min Scan# 2255
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:284 Resp: 512098
Ion Ratio Lower Upper
284 100
142 37.0 25.3 37.9
249 32.2 24.3 36.5









#70
Atrazine
Concen: 39.957 ng/ul
RT: 16.534 min Scan# 2286
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:200 Resp: 474515
Ion Ratio Lower Upper
200 100
173 27.5 22.6 34.0
215 47.7 42.0 63.0



#71
Pentachlorophenol
Concen: 44.037 ng/ul
RT: 16.698 min Scan# 2314
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:266 Resp: 296488
Ion Ratio Lower Upper
266 100
264 64.4 50.7 76.1
268 65.6 49.8 74.6









#72
Phenanthrene
Concen: 41.030 ng/ul
RT: 17.092 min Scan# 2381
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:	178	Resp:	2391737
Ion	Ratio	Lower	Upper
178	100		
179	15.4	12.2	18.2
176	18.9	14.7	22.1



#74
Anthracene
Concen: 40.964 ng/ul
RT: 17.187 min Scan# 2397
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:	178	Resp:	2454313
Ion	Ratio	Lower	Upper
178	100		
179	15.4	12.4	18.6
176	18.4	14.2	21.4









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 41.231 ng/uL
 RT: 13.087 min Scan# 1700
 Delta R.T. 0.000 min
 Lab File: BN016415.D
 Acq: 22 Sep 2021 18:32




Instrument :
 BNA_N
 ClientSampleId :
 SSTD040218


Tgt Ion:	216	Resp:	677723
Ion	Ratio	Lower	Upper
216	100		
214	78.1	60.6	90.8
218	48.5	40.2	60.4



#76
 Pentachlorobenzene
 Concen: 39.349 ng/uL
 RT: 14.616 min Scan# 1960
 Delta R.T. 0.000 min
 Lab File: BN016415.D
 Acq: 22 Sep 2021 18:32

Tgt Ion:	250	Resp:	639906
Ion	Ratio	Lower	Upper
250	100		
248	60.0	51.1	76.7
252	62.4	52.1	78.1









#77
 Carbazole
 Concen: 41.747 ng/ul
 RT: 17.457 min Scan# 2443
 Delta R.T. 0.000 min
 Lab File: BN016415.D
 Acq: 22 Sep 2021 18:32




Instrument :
 BNA_N
 ClientSampleId :
 SSTD040218


Tgt Ion:167 Resp: 2223760
 Ion Ratio Lower Upper
 167 100
 166 21.0 16.8 25.2
 139 13.4 9.8 14.6



#78
 Di-n-butylphthalate
 Concen: 42.091 ng/ul
 RT: 18.045 min Scan# 2543
 Delta R.T. 0.006 min
 Lab File: BN016415.D
 Acq: 22 Sep 2021 18:32

Tgt Ion:149 Resp: 2623134
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.3 10.9
 104 4.7 3.5 5.3









#80
Fluoranthene
Concen: 42.444 ng/ul
RT: 19.122 min Scan# 2726
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:202 Resp: 2771615
Ion Ratio Lower Upper
202 100
101 12.2 8.7 13.1
100 9.2 6.6 9.8



#82
Pyrene
Concen: 42.447 ng/ul
RT: 19.486 min Scan# 2788
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:202 Resp: 2855208
Ion Ratio Lower Upper
202 100
101 13.7 10.5 15.7
100 11.3 8.6 13.0









#83
Butylbenzylphthalate
Concen: 41.388 ng/ul
RT: 20.410 min Scan# 2945
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:149 Resp: 1116152
Ion Ratio Lower Upper
149 100
91 69.4 50.6 76.0
206 21.7 18.5 27.7



#84
3,3'-Dichlorobenzidine
Concen: 42.673 ng/ul
RT: 21.186 min Scan# 3077
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:252 Resp: 963961
Ion Ratio Lower Upper
252 100
254 66.6 48.6 72.8
126 12.2 7.9 11.9#









#85
Benzo(a)anthracene
Concen: 41.108 ng/ul
RT: 21.239 min Scan# 3086
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:228 Resp: 2756729
Ion Ratio Lower Upper
228 100
229 19.2 15.7 23.5
226 26.1 20.6 31.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 44.877 ng/ul
RT: 21.198 min Scan# 3079
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:149 Resp: 1767107
Ion Ratio Lower Upper
149 100
167 27.5 24.2 36.4
279 4.6 4.2 6.2









#87
Chrysene
Concen: 40.979 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:228 Resp: 2678501
Ion Ratio Lower Upper
228 100
226 30.1 23.4 35.0
229 19.5 14.9 22.3



#89
Di-n-octyl phthalate
Concen: 41.739 ng/ul
RT: 22.086 min Scan# 3230
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:149 Resp: 2862732









#90
Benzo(b)fluoranthene
Concen: 41.197 ng/ul
RT: 22.839 min Scan# 3358
Delta R.T. 0.000 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:252 Resp: 2774487
Ion Ratio Lower Upper
252 100
253 22.6 17.6 26.4
125 11.6 7.8 11.8



#91
Benzo(k)fluoranthene
Concen: 44.836 ng/ul
RT: 22.839 min Scan# 3358
Delta R.T. -0.041 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:252 Resp: 2774487
Ion Ratio Lower Upper
252 100
253 22.6 17.0 25.4
125 11.6 7.8 11.6









#93
Benzo(a)pyrene
Concen: 42.099 ng/ul
RT: 23.427 min Scan# 3458
Delta R.T. 0.012 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32




Instrument :
BNA_N
ClientSampleId :
SSTD040218


Tgt Ion:252 Resp: 2739548
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 12.7 8.9 13.3



#94
Indeno(1,2,3-cd)pyrene
Concen: 35.771 ng/ul
RT: 26.533 min Scan# 3986
Delta R.T. 0.018 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:276 Resp: 2621477
Ion Ratio Lower Upper
276 100
138 23.8 19.4 29.0
277 23.9 18.6 27.8









#95
Dibenzo(a,h)anthracene
Concen: 42.145 ng/ul
RT: 25.845 min Scan# 3869
Delta R.T. 0.006 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Instrument :
BNA_N
ClientSampleId :
SSTD040218

Tgt Ion:278 Resp: 2658682
Ion Ratio Lower Upper
278 100
139 18.4 12.2 18.4#
279 23.3 18.2 27.2



#96
Benzo(g,h,i)perylene
Concen: 42.201 ng/ul
RT: 26.533 min Scan# 3986
Delta R.T. 0.018 min
Lab File: BN016415.D
Acq: 22 Sep 2021 18:32

Tgt Ion:276 Resp: 2621477
Ion Ratio Lower Upper
276 100
138 23.8 17.4 26.0
277 23.9 18.4 27.6

