

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092321\
 Data File : BN016416.D
 Acq On : 22 Sep 2021 19:09
 Operator : CG/JU
 Sample : SST08019
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST080219

Quant Time: Sep 23 02:01:05 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN092321MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 23 01:47:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.658	152	227732	20.000	ng/u1	0.00
20) Naphthalene-d8	10.434	136	992980	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.298	164	595430	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.051	188	1216491	20.000	ng/u1	0.00
79) Chrysene-d12	21.263	240	1115385	20.000	ng/u1	# 0.00
88) Perylene-d12	23.527	264	1182200	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	170069	38.969	ng/uL	0.00
4) Pyridine-d5	3.599	84	1251653	98.400	ng/u1	0.00
7) Phenol-d5	6.840	99	1526064	88.814	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.005	67	870916	95.065	ng/u1	0.00
11) 2-Chlorophenol-d4	7.193	132	1236651	84.912	ng/u1	0.00
15) 4-Methylphenol-d8	8.375	113	1195793	86.333	ng/u1	0.01
21) Nitrobenzene-d5	8.810	128	584839	80.719	ng/u1	0.00
24) 2-Nitrophenol-d4	9.528	143	620975	76.514	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.063	165	1161837	78.448	ng/u1	0.00
31) 4-Chloroaniline-d4	10.581	131	1673165	82.356	ng/u1	0.00
46) Dimethylphthalate-d6	13.728	166	3214765	84.094	ng/u1	0.01
49) Acenaphthylene-d8	13.992	160	4185149	83.570	ng/u1	0.00
54) 4-Nitrophenol-d4	14.522	143	655240	88.959	ng/u1	0.02
60) Fluorene-d10	15.298	176	2698650	80.081	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.428	200	527975	79.565	ng/u1	0.01
73) Anthracene-d10	17.157	188	4120117	78.028	ng/u1	0.01
81) Pyrene-d10	19.457	212	4478520	84.683	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.386	264	4872092	84.171	ng/u1	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	185693	37.790	ng/uL	94
5) Pyridine	3.617	79	1240724	98.134	ng/u1	89
6) Benzaldehyde	6.811	77	806106	78.278	ng/u1	95
8) Phenol	6.869	94	1570541	89.708	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.099	93	1206194	90.808	ng/u1	99
12) 2-Chlorophenol	7.228	128	1256776	85.836	ng/u1	98
13) 2-Methylphenol	8.105	108	1190521	87.976	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.199	45	1619602	92.323	ng/u1	98
16) Acetophenone	8.481	105	1791944	85.088	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.481	70	897797	92.018	ng/u1	98
18) 4-Methylphenol	8.440	108	1263078	86.333	ng/u1	97
19) Hexachloroethane	8.722	117	486133	84.077	ng/u1	94
22) Nitrobenzene	8.857	77	1318917	89.323	ng/u1	96
23) Isophorone	9.381	82	2556814	89.121	ng/u1	97
25) 2-Nitrophenol	9.557	139	688022	79.424	ng/u1	96
26) 2,4-Dimethylphenol	9.628	107	1357479	84.097	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.869	93	1590805	87.589	ng/u1	97
29) 2,4-Dichlorophenol	10.087	162	1127747	77.447	ng/u1	96
30) Naphthalene	10.487	128	3860777	79.674	ng/u1	99
32) 4-Chloroaniline	10.604	127	1715522	83.040	ng/u1	99
33) Hexachlorobutadiene	10.769	225	683434	70.378	ng/u1	99
34) Caprolactam	11.387	113	211688	42.766	ng/u1	96
35) 4-Chloro-3-methylphenol	11.740	107	1244189	82.627	ng/u1	95
36) 2-Methylnaphthalene	12.104	142	2613517	79.086	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092321\
 Data File : BN016416.D
 Acq On : 22 Sep 2021 19:09
 Operator : CG/JU
 Sample : SST08019
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST080219

Quant Time: Sep 23 02:01:05 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN092321MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 23 01:47:54 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.328	142	2650388	78.459	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.481	216	1326538	79.312	ng/ul	98
40) Hexachlorocyclopentadiene	12.457	237	874651	91.203	ng/ul	95
41) 2,4,6-Trichlorophenol	12.722	196	868335	81.010	ng/ul	99
42) 2,4,5-Trichlorophenol	12.793	196	961879	82.942	ng/ul	99
43) 1,1'-Biphenyl	13.134	154	3380099	84.089	ng/ul#	98
44) 2-Chloronaphthalene	13.169	162	2601775	81.652	ng/ul	97
45) 2-Nitroaniline	13.387	65	763190	94.216	ng/ul	96
47) Dimethylphthalate	13.775	163	3197481	83.653	ng/ul	99
48) 2,6-Dinitrotoluene	13.892	165	692997	85.461	ng/ul	90
50) Acenaphthylene	14.022	152	4238777	81.932	ng/ul	98
51) 3-Nitroaniline	14.216	138	723581	80.616	ng/ul	90
52) Acenaphthene	14.363	153	2726710	82.584	ng/ul	98
53) 2,4-Dinitrophenol	14.422	184	383521	88.317	ng/ul	94
55) 4-Nitrophenol	14.540	109	455432	92.490	ng/ul	94
56) Dibenzofuran	14.704	168	3845287	80.593	ng/ul	95
57) 2,4-Dinitrotoluene	14.681	165	972422	83.209	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.934	232	760501	79.997	ng/ul	89
59) Diethylphthalate	15.151	149	3259137	86.203	ng/ul	98
61) Fluorene	15.357	166	3009039	78.143	ng/ul	94
62) 4-Chlorophenyl-phenyle...	15.357	204	1486000	75.926	ng/ul	97
63) 4-Nitroaniline	15.392	138	680251	72.868	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.445	198	529888	84.491	ng/ul#	98
67) N-Nitrosodiphenylamine	15.575	169	2693440	81.436	ng/ul	96
68) 4-Bromophenyl-phenylether	16.251	248	939904	79.348	ng/ul	91
69) Hexachlorobenzene	16.357	284	1074552	77.924	ng/ul	96
70) Atrazine	16.539	200	1011786	81.546	ng/ul	98
71) Pentachlorophenol	16.704	266	670000	95.248	ng/ul	98
72) Phenanthrene	17.098	178	4785605	78.578	ng/ul	97
74) Anthracene	17.186	178	4870354	77.804	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.087	216	1377961	80.238	ng/ul	95
76) Pentachlorobenzene	14.622	250	1307940	76.979	ng/ul	99
77) Carbazole	17.463	167	4527223	81.347	ng/ul	98
78) Di-n-butylphthalate	18.045	149	5455196	83.783	ng/ul	99
80) Fluoranthene	19.122	202	5640382	86.736	ng/ul	96
82) Pyrene	19.486	202	5631583	84.070	ng/ul	95
83) Butylbenzylphthalate	20.410	149	2463621	91.734	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.186	252	2024019	89.975	ng/ul#	94
85) Benzo(a)anthracene	21.245	228	5419242	81.148	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.198	149	3546823	90.449	ng/ul	97
87) Chrysene	21.298	228	5168132	79.399	ng/ul	97
89) Di-n-octyl phthalate	22.092	149	6195171	86.945	ng/ul	100
90) Benzo(b)fluoranthene	22.845	252	5742579	82.077	ng/ul	97
91) Benzo(k)fluoranthene	22.845	252	5742579	89.328	ng/ul	98
93) Benzo(a)pyrene	23.433	252	5578710	82.520	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.551	276	5550054	72.898	ng/ul	98
95) Dibenzo(a,h)anthracene	25.857	278	5672435	86.553	ng/ul#	95
96) Benzo(g,h,i)perylene	26.551	276	5550054	86.002	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#2
1,4-Dioxane
Concen: 37.790 ng/uL
RT: 3.240 min Scan# 26
Delta R.T. -0.000 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Instrument :
BNA_N
ClientSampleId :
SSTD080219




Tgt Ion: 88 Resp: 185693

Ion	Ratio	Lower	Upper
88	100		
43	26.3	23.2	34.8
58	73.0	54.3	81.5






#5
Pyridine
Concen: 98.134 ng/uL
RT: 3.617 min Scan# 90
Delta R.T. -0.000 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09



Tgt Ion: 79 Resp: 1240724

Ion	Ratio	Lower	Upper
79	100		
52	72.5	66.8	100.2
51	33.2	30.7	46.1









#6
Benzaldehyde
Concen: 78.278 ng/ul
RT: 6.811 min Scan# 633
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion: 77 Resp: 806106
Ion Ratio Lower Upper
77 100
105 100.8 83.7 125.5
106 95.6 82.2 123.2



#8
Phenol
Concen: 89.708 ng/ul
RT: 6.869 min Scan# 643
Delta R.T. 0.012 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion: 94 Resp: 1570541
Ion Ratio Lower Upper
94 100
65 25.7 21.5 32.3
66 34.1 28.6 42.8





#10

Bis(2-Chloroethyl)ether

Concen: 90.808 ng/ul

RT: 7.099 min Scan# 682





Delta R.T. 0.006 min

Lab File: BN016416.D

Acq: 22 Sep 2021 19:09

Tgt Ion: 93 Resp: 1206194

Ion	Ratio	Lower	Upper
93	100		
63	75.1	59.3	88.9
95	32.3	26.3	39.5



#12

2-Chlorophenol

Concen: 85.836 ng/ul

RT: 7.228 min Scan# 704




Delta R.T. 0.006 min


Lab File: BN016416.D

Acq: 22 Sep 2021 19:09

Tgt Ion: 128 Resp: 1256776

Ion	Ratio	Lower	Upper
128	100		
64	43.4	33.4	50.2
130	32.2	25.4	38.0









#13
2-Methylphenol
Concen: 87.976 ng/ul
RT: 8.105 min Scan# 853
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:108 Resp: 1190521
Ion Ratio Lower Upper
108 100
107 87.2 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 92.323 ng/ul
RT: 8.199 min Scan# 869
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion: 45 Resp: 1619602
Ion Ratio Lower Upper
45 100
77 14.3 11.0 16.4
79 11.6 8.4 12.6









#16
Acetophenone
Concen: 85.088 ng/ul
RT: 8.481 min Scan# 917
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:105 Resp: 1791944
Ion Ratio Lower Upper
105 100
77 74.3 59.0 88.4
51 25.0 20.3 30.5



#17
N-Nitroso-di-n-propylamine
Concen: 92.018 ng/ul
RT: 8.481 min Scan# 917
Delta R.T. 0.018 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion: 70 Resp: 897797
Ion Ratio Lower Upper
70 100
42 55.2 45.1 67.7
101 10.4 8.9 13.3
130 22.7 19.9 29.9









#18
4-Methylphenol
Concen: 86.333 ng/ul
RT: 8.440 min Scan# 910
Delta R.T. 0.012 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:108 Resp: 1263078
Ion Ratio Lower Upper
108 100
107 111.7 87.0 130.4



#19
Hexachloroethane
Concen: 84.077 ng/ul
RT: 8.722 min Scan# 958
Delta R.T. -0.000 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:117 Resp: 486133
Ion Ratio Lower Upper
117 100
201 96.6 82.6 123.8
199 60.3 52.0 78.0









#22
Nitrobenzene
Concen: 89.323 ng/ul
RT: 8.857 min Scan# 981
Delta R.T. 0.012 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion: 77 Resp: 1318917
Ion Ratio Lower Upper
77 100
123 48.6 41.1 61.7
65 12.9 11.0 16.6



#23
Isophorone
Concen: 89.121 ng/ul
RT: 9.381 min Scan# 1070
Delta R.T. 0.012 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion: 82 Resp: 2556814
Ion Ratio Lower Upper
82 100
95 7.0 6.1 9.1
138 18.3 15.8 23.8









#25
2-Nitrophenol
Concen: 79.424 ng/ul
RT: 9.557 min Scan# 1100
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:139 Resp: 688022
Ion Ratio Lower Upper
139 100
65 46.9 35.8 53.6
109 26.8 23.4 35.0



#26
2,4-Dimethylphenol
Concen: 84.097 ng/ul
RT: 9.628 min Scan# 1112
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:107 Resp: 1357479
Ion Ratio Lower Upper
107 100
121 51.6 41.8 62.8
122 88.4 71.1 106.7









#27

Bis(2-Chloroethoxy)methane
Concen: 87.589 ng/ul
RT: 9.869 min Scan# 1153
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Instrument :
BNA_N
ClientSampleId :
SSTD080219




Tgt Ion: 93 Resp: 1590805
Ion Ratio Lower Upper
93 100
95 33.3 24.9 37.3
123 11.8 9.4 14.0




#29

2,4-Dichlorophenol
Concen: 77.447 ng/ul
RT: 10.087 min Scan# 1190
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:162 Resp: 1127747
Ion Ratio Lower Upper
162 100
164 65.8 51.7 77.5
98 37.8 26.3 39.5









#30
Naphthalene
Concen: 79.674 ng/ul
RT: 10.487 min Scan# 1258
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:128 Resp: 3860777
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.4
127 13.5 10.2 15.4



#32
4-Chloroaniline
Concen: 83.040 ng/ul
RT: 10.604 min Scan# 1278
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:127 Resp: 1715522
Ion Ratio Lower Upper
127 100
129 31.8 25.8 38.8









#33
Hexachlorobutadiene
Concen: 70.378 ng/ul
RT: 10.769 min Scan# 1306
Delta R.T. -0.000 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:	225	Resp:	683434
Ion	Ratio	Lower	Upper
225	100		
223	61.4	49.8	74.6
227	62.1	49.8	74.6



#34
Caprolactam
Concen: 42.766 ng/ul
RT: 11.387 min Scan# 1411
Delta R.T. 0.035 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:	113	Resp:	211688
Ion	Ratio	Lower	Upper
113	100		
55	158.4	123.8	185.8
56	119.1	89.4	134.2









#35
4-Chloro-3-methylphenol
Concen: 82.627 ng/ul
RT: 11.740 min Scan# 1471
Delta R.T. 0.012 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:107 Resp: 1244189
Ion Ratio Lower Upper
107 100
144 25.9 21.8 32.8
142 79.3 67.3 100.9



#36
2-Methylnaphthalene
Concen: 79.086 ng/ul
RT: 12.104 min Scan# 1533
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:142 Resp: 2613517
Ion Ratio Lower Upper
142 100
141 90.1 71.3 106.9









#37
1-Methylnaphthalene
Concen: 78.459 ng/ul
RT: 12.328 min Scan# 1571
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:142 Resp: 2650388
Ion Ratio Lower Upper
142 100
141 92.8 71.9 107.9
116 3.6 2.7 4.1



#39
1,2,4,5-Tetrachlorobenzene
Concen: 79.312 ng/ul
RT: 12.481 min Scan# 1597
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:216 Resp: 1326538
Ion Ratio Lower Upper
216 100
214 78.6 61.8 92.6
179 21.0 15.8 23.8
108 16.9 11.4 17.0









#40
Hexachlorocyclopentadiene
Concen: 91.203 ng/ul
RT: 12.457 min Scan# 1593
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:237 Resp: 874651
Ion Ratio Lower Upper
237 100
235 62.0 52.7 79.1
272 13.1 11.8 17.8



#41
2,4,6-Trichlorophenol
Concen: 81.010 ng/ul
RT: 12.722 min Scan# 1638
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:196 Resp: 868335
Ion Ratio Lower Upper
196 100
198 95.3 75.2 112.8
200 29.4 24.2 36.4









#42
2,4,5-Trichlorophenol
Concen: 82.942 ng/ul
RT: 12.793 min Scan# 1650
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:196 Resp: 961879
Ion Ratio Lower Upper
196 100
198 92.4 72.5 108.7
200 30.7 24.6 37.0



#43
1,1'-Biphenyl
Concen: 84.089 ng/ul
RT: 13.134 min Scan# 1708
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:154 Resp: 3380099
Ion Ratio Lower Upper
154 100
153 41.0 32.1 48.1
76 13.7 9.1 13.7#









#44
2-Chloronaphthalene
Concen: 81.652 ng/ul
RT: 13.169 min Scan# 1714
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:162 Resp: 2601775
Ion Ratio Lower Upper
162 100
164 33.2 26.2 39.4
127 40.7 30.1 45.1



#45
2-Nitroaniline
Concen: 94.216 ng/ul
RT: 13.387 min Scan# 1751
Delta R.T. 0.012 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion: 65 Resp: 763190
Ion Ratio Lower Upper
65 100
92 71.7 57.8 86.8
138 117.6 99.6 149.4









#47
Dimethylphthalate
Concen: 83.653 ng/ul
RT: 13.775 min Scan# 1817
Delta R.T. 0.012 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:163 Resp: 3197481
Ion Ratio Lower Upper
163 100
194 4.3 3.3 4.9
164 9.9 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 85.461 ng/ul
RT: 13.892 min Scan# 1837
Delta R.T. 0.012 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:165 Resp: 692997
Ion Ratio Lower Upper
165 100
89 52.6 35.6 53.4
121 19.8 14.8 22.2









#50
Acenaphthylene
Concen: 81.932 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:152 Resp: 4238777
Ion Ratio Lower Upper
152 100
151 20.1 14.8 22.2
153 13.0 10.3 15.5



#51
3-Nitroaniline
Concen: 80.616 ng/ul
RT: 14.216 min Scan# 1892
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:138 Resp: 723581
Ion Ratio Lower Upper
138 100
108 10.9 9.0 13.4
92 113.9 82.3 123.5









#52
Acenaphthene
Concen: 82.584 ng/ul
RT: 14.363 min Scan# 1917
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:153 Resp: 2726710
Ion Ratio Lower Upper
153 100
152 48.1 36.6 54.8
154 89.7 70.9 106.3



#53
2,4-Dinitrophenol
Concen: 88.317 ng/ul
RT: 14.422 min Scan# 1927
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:184 Resp: 383521
Ion Ratio Lower Upper
184 100
63 61.3 45.2 67.8
154 60.0 51.8 77.8









#55
4-Nitrophenol
Concen: 92.490 ng/ul
RT: 14.540 min Scan# 1947
Delta R.T. 0.024 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:109 Resp: 455432
Ion Ratio Lower Upper
109 100
139 152.3 129.5 194.3
65 131.9 109.0 163.4



#56
Dibenzofuran
Concen: 80.593 ng/ul
RT: 14.704 min Scan# 1975
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:168 Resp: 3845287
Ion Ratio Lower Upper
168 100
139 37.3 27.6 41.4









#57
2,4-Dinitrotoluene
Concen: 83.209 ng/ul
RT: 14.681 min Scan# 1971
Delta R.T. 0.012 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:165 Resp: 972422
Ion Ratio Lower Upper
165 100
63 40.3 30.4 45.6
182 2.9 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 79.997 ng/ul
RT: 14.934 min Scan# 2014
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:232 Resp: 760501
Ion Ratio Lower Upper
232 100
131 43.6 29.1 43.7
130 2.3 1.7 2.5
166 31.7 21.6 32.4









#59
Diethylphthalate
Concen: 86.203 ng/ul
RT: 15.151 min Scan# 2051
Delta R.T. 0.012 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:149 Resp: 3259137
Ion Ratio Lower Upper
149 100
177 22.4 19.0 28.4
150 12.8 10.2 15.4



#61
Fluorene
Concen: 78.143 ng/ul
RT: 15.357 min Scan# 2086
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:166 Resp: 3009039
Ion Ratio Lower Upper
166 100
165 99.8 74.8 112.2
167 14.0 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 75.926 ng/ul
RT: 15.357 min Scan# 2086
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:204 Resp: 1486000
Ion Ratio Lower Upper
204 100
206 33.2 27.6 41.4
141 59.2 45.5 68.3



#63
4-Nitroaniline
Concen: 72.868 ng/ul
RT: 15.392 min Scan# 2092
Delta R.T. 0.018 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:138 Resp: 680251
Ion Ratio Lower Upper
138 100
92 48.1 35.5 53.3
108 68.7 49.4 74.2









#66
 4,6-Dinitro-2-methylphenol
 Concen: 84.491 ng/ul
 RT: 15.445 min Scan# 2101
 Delta R.T. 0.018 min
 Lab File: BN016416.D
 Acq: 22 Sep 2021 19:09




Instrument :
 BNA_N
 ClientSampleId :
 SSTD080219


Tgt Ion:198 Resp: 529888
 Ion Ratio Lower Upper
 198 100
 182 3.5 4.2 6.4#
 77 22.4 17.5 26.3



#67
 N-Nitrosodiphenylamine
 Concen: 81.436 ng/ul
 RT: 15.575 min Scan# 2123
 Delta R.T. 0.006 min
 Lab File: BN016416.D
 Acq: 22 Sep 2021 19:09

Tgt Ion:169 Resp: 2693440
 Ion Ratio Lower Upper
 169 100
 168 67.4 51.6 77.4
 167 36.6 27.1 40.7









#68
4-Bromophenyl-phenylether
Concen: 79.348 ng/ul
RT: 16.251 min Scan# 2238
Delta R.T. -0.000 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:	248	Resp:	939904
Ion Ratio	Lower	Upper	
248	100		
250	103.6	76.4	114.6
141	79.6	57.0	85.6



#69
Hexachlorobenzene
Concen: 77.924 ng/ul
RT: 16.357 min Scan# 2256
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:	284	Resp:	1074552
Ion Ratio	Lower	Upper	
284	100		
142	35.0	25.3	37.9
249	31.0	24.3	36.5









#70
Atrazine
Concen: 81.546 ng/ul
RT: 16.539 min Scan# 2287
Delta R.T. 0.012 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:200 Resp: 1011786
Ion Ratio Lower Upper
200 100
173 28.3 22.6 34.0
215 49.8 42.0 63.0



#71
Pentachlorophenol
Concen: 95.248 ng/ul
RT: 16.704 min Scan# 2315
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:266 Resp: 670000
Ion Ratio Lower Upper
266 100
264 61.6 50.7 76.1
268 61.5 49.8 74.6









#72
Phenanthrene
Concen: 78.578 ng/ul
RT: 17.098 min Scan# 2382
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:178 Resp: 4785605
Ion Ratio Lower Upper
178 100
179 15.9 12.2 18.2
176 20.0 14.7 22.1



#74
Anthracene
Concen: 77.804 ng/ul
RT: 17.186 min Scan# 2397
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:178 Resp: 4870354
Ion Ratio Lower Upper
178 100
179 16.0 12.4 18.6
176 18.8 14.2 21.4









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 80.238 ng/uL
 RT: 13.087 min Scan# 1700
 Delta R.T. -0.000 min
 Lab File: BN016416.D
 Acq: 22 Sep 2021 19:09

Instrument :
 BNA_N
 ClientSampleId :
 SSTD080219

Tgt Ion:216 Resp: 1377961




Ion	Ratio	Lower	Upper
216	100		
214	79.8	60.6	90.8
218	47.4	40.2	60.4




#76
 Pentachlorobenzene
 Concen: 76.979 ng/uL
 RT: 14.622 min Scan# 1961
 Delta R.T. 0.006 min
 Lab File: BN016416.D
 Acq: 22 Sep 2021 19:09

Tgt Ion:250 Resp: 1307940


Ion	Ratio	Lower	Upper
250	100		
248	64.2	51.1	76.7
252	64.2	52.1	78.1








#77
Carbazole
Concen: 81.347 ng/ul
RT: 17.463 min Scan# 2444
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Instrument :
BNA_N
ClientSampleId :
SSTD080219







Tgt Ion:167 Resp: 4527223
Ion Ratio Lower Upper
167 100
166 21.7 16.8 25.2
139 13.8 9.8 14.6



#78
Di-n-butylphthalate
Concen: 83.783 ng/ul
RT: 18.045 min Scan# 2543
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:149 Resp: 5455196
Ion Ratio Lower Upper
149 100
150 9.6 7.3 10.9
104 4.9 3.5 5.3









#80
Fluoranthene
Concen: 86.736 ng/u1
RT: 19.122 min Scan# 2726
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:202 Resp: 5640382
Ion Ratio Lower Upper
202 100
101 12.6 8.7 13.1
100 9.5 6.6 9.8



#82
Pyrene
Concen: 84.070 ng/u1
RT: 19.486 min Scan# 2788
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:202 Resp: 5631583
Ion Ratio Lower Upper
202 100
101 15.2 10.5 15.7
100 12.4 8.6 13.0









#83
Butylbenzylphthalate
Concen: 91.734 ng/ul
RT: 20.410 min Scan# 2945
Delta R.T. -0.000 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:149 Resp: 2463621
Ion Ratio Lower Upper
149 100
91 68.9 50.6 76.0
206 20.9 18.5 27.7



#84
3,3'-Dichlorobenzidine
Concen: 89.975 ng/ul
RT: 21.186 min Scan# 3077
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:252 Resp: 2024019
Ion Ratio Lower Upper
252 100
254 65.2 48.6 72.8
126 11.9 7.9 11.9#









#85
Benzo(a)anthracene
Concen: 81.148 ng/ul
RT: 21.245 min Scan# 3087
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:228 Resp: 5419242
Ion Ratio Lower Upper
228 100
229 20.7 15.7 23.5
226 28.0 20.6 31.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 90.449 ng/ul
RT: 21.198 min Scan# 3079
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:149 Resp: 3546823
Ion Ratio Lower Upper
149 100
167 28.4 24.2 36.4
279 4.5 4.2 6.2









#87
Chrysene
Concen: 79.399 ng/ul
RT: 21.298 min Scan# 3096
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:228 Resp: 5168132
Ion Ratio Lower Upper
228 100
226 30.7 23.4 35.0
229 20.6 14.9 22.3



#89
Di-n-octyl phthalate
Concen: 86.945 ng/ul
RT: 22.092 min Scan# 3231
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:149 Resp: 6195171









#90
Benzo(b)fluoranthene
Concen: 82.077 ng/ul
RT: 22.845 min Scan# 3359
Delta R.T. 0.006 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:252 Resp: 5742579
Ion Ratio Lower Upper
252 100
253 21.2 17.6 26.4
125 11.6 7.8 11.8



#91
Benzo(k)fluoranthene
Concen: 89.328 ng/ul
RT: 22.845 min Scan# 3359
Delta R.T. -0.035 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:252 Resp: 5742579
Ion Ratio Lower Upper
252 100
253 21.2 17.0 25.4
125 11.6 7.8 11.6









#93
Benzo(a)pyrene
Concen: 82.520 ng/ul
RT: 23.433 min Scan# 3459
Delta R.T. 0.018 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09




Instrument :
BNA_N
ClientSampleId :
SSTD080219


Tgt Ion:252 Resp: 5578710
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.0
125 13.3 8.9 13.3



#94
Indeno(1,2,3-cd)pyrene
Concen: 72.898 ng/ul
RT: 26.551 min Scan# 3989
Delta R.T. 0.035 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:276 Resp: 5550054
Ion Ratio Lower Upper
276 100
138 24.7 19.4 29.0
277 24.2 18.6 27.8









#95
Dibenzo(a,h)anthracene
Concen: 86.553 ng/ul
RT: 25.857 min Scan# 3871
Delta R.T. 0.018 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Instrument :
BNA_N
ClientSampleId :
SSTD080219

Tgt Ion:278 Resp: 5672435
Ion Ratio Lower Upper
278 100
139 18.5 12.2 18.4#
279 24.4 18.2 27.2



#96
Benzo(g,h,i)perylene
Concen: 86.002 ng/ul
RT: 26.551 min Scan# 3989
Delta R.T. 0.035 min
Lab File: BN016416.D
Acq: 22 Sep 2021 19:09

Tgt Ion:276 Resp: 5550054
Ion Ratio Lower Upper
276 100
138 24.7 17.4 26.0
277 24.2 18.4 27.6

