

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092821\
 Data File : BN016559.D
 Acq On : 28 Sep 2021 16:34
 Operator : CG/JU
 Sample : SSTDIC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.1

Quant Time: Sep 28 18:32:05 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 28 18:30:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.652	152	25136	0.400	ng	0.00
7) Naphthalene-d8	10.419	136	106660	0.400	ng	# 0.00
13) Acenaphthene-d10	14.281	164	52895	0.400	ng	0.00
19) Phenanthrene-d10	17.030	188	106217	0.400	ng	#-0.01
29) Chrysene-d12	21.249	240	90428	0.400	ng	# 0.00
35) Perylene-d12	23.505	264	86020	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.282	112	6520	0.127	ng	0.00
5) Phenol-d6	6.831	99	6304	0.100	ng	0.00
8) Nitrobenzene-d5	8.797	82	4372	0.071	ng	0.00
11) 2-Methylnaphthalene-d10	12.016	152	14472	0.096	ng	0.00
14) 2,4,6-Tribromophenol	15.791	330	1995	0.153	ng	0.01
15) 2-Fluorobiphenyl	12.911	172	22195	0.090	ng	0.00
27) Fluoranthene-d10	19.078	212	27326	0.092	ng	0.00
31) Terphenyl-d14	19.688	244	21808	0.119	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.263	88	861	0.074	ng	# 63
3) n-Nitrosodimethylamine	3.613	42	1880	0.084	ng	# 64
6) bis(2-Chloroethyl)ether	7.089	93	6597	0.099	ng	91
9) Naphthalene	10.470	128	28521	0.096	ng	98
10) Hexachlorobutadiene	10.762	225	6014	0.096	ng	# 99
12) 2-Methylnaphthalene	12.092	142	17369	0.100	ng	98
16) Acenaphthylene	14.002	152	21651	0.089	ng	100
17) Acenaphthene	14.345	154	15447	0.094	ng	98
18) Fluorene	15.335	166	18957	0.096	ng	99
20) 4,6-Dinitro-2-methylph...	15.449	198	183	0.030	ng	# 1
21) 4-Bromophenyl-phenylether	16.239	248	5414	0.082	ng	# 73
22) Hexachlorobenzene	16.336	284	7443	0.094	ng	95
23) Atrazine	16.519	200	3144	0.119	ng	97
24) Pentachlorophenol	16.713	266	1809	0.087	ng	98
25) Phenanthrene	17.078	178	30864	0.089	ng	99
26) Anthracene	17.164	178	24415	0.085	ng	100
28) Fluoranthene	19.108	202	33477	0.089	ng	98
30) Pyrene	19.475	202	33746	0.129	ng	100
32) Benzo(a)anthracene	21.227	228	28189	0.118	ng	98
33) Chrysene	21.281	228	30004	0.102	ng	99
34) Bis(2-ethylhexyl)phtha...	21.185	149	10214	0.228	ng	100
36) Indeno(1,2,3-cd)pyrene	25.798	276	23550	0.087	ng	# 94
37) Benzo(b)fluoranthene	22.824	252	27876	0.105	ng	# 92
38) Benzo(k)fluoranthene	22.872	252	26741	0.086	ng	# 86
39) Benzo(a)pyrene	23.406	252	27807	0.131	ng	# 87
40) Dibenzo(a,h)anthracene	25.822	278	16633	0.069	ng	# 80
41) Benzo(g,h,i)perylene	26.493	276	23827	0.107	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092821\
Data File : BN016559.D
Acq On : 28 Sep 2021 16:34
Operator : CG/JU
Sample : SSTDIC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Quant Time: Sep 28 18:32:05 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 28 18:30:42 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.652 min Scan# 496
 Delta R.T. -0.009 min
 Lab File: BN016559.D
 Acq: 28 Sep 2021 16:34




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:152 Resp: 25136
 Ion Ratio Lower Upper
 152 100
 150 153.7 125.1 187.7
 115 60.1 45.8 68.8



#2
 1,4-Dioxane
 Concen: 0.074 ng
 RT: 3.263 min Scan# 20
 Delta R.T. -0.009 min
 Lab File: BN016559.D
 Acq: 28 Sep 2021 16:34

Tgt Ion: 88 Resp: 861
 Ion Ratio Lower Upper
 88 100
 43 24.3 75.8 113.8#
 58 97.8 81.6 122.4









#3
n-Nitrosodimethylamine
Concen: 0.084 ng
RT: 3.613 min Scan# 58
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion: 42 Resp: 1880
Ion Ratio Lower Upper
42 100
74 132.1 77.5 116.3#
44 0.0 0.0 0.0



#4
2-Fluorophenol
Concen: 0.127 ng
RT: 5.282 min Scan# 239
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion: 112 Resp: 6520
Ion Ratio Lower Upper
112 100
64 55.6 48.4 72.6
63 29.4 25.4 38.2









#5
Phenol-d6
Concen: 0.100 ng
RT: 6.831 min Scan# 407
Delta R.T. -0.009 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	99	Resp:	6304
Ion	Ratio	Lower	Upper
99	100		
42	20.4	19.6	29.4
71	31.2	24.2	36.2



#6
bis(2-Chloroethyl)ether
Concen: 0.099 ng
RT: 7.089 min Scan# 435
Delta R.T. -0.009 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:	93	Resp:	6597
Ion	Ratio	Lower	Upper
93	100		
63	77.6	70.0	105.0
95	35.3	26.3	39.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.419 min Scan# 734
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	136	Resp:	106660
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.8	13.2
54	8.0	5.8	8.6
68	6.6	3.6	5.4



#8
Nitrobenzene-d5
Concen: 0.071 ng
RT: 8.797 min Scan# 606
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:	82	Resp:	4372
Ion	Ratio	Lower	Upper
82	100		
128	42.4	27.4	41.2
54	60.7	58.5	87.7









#9
Naphthalene
Concen: 0.096 ng
RT: 10.470 min Scan# 738
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	128	Resp:	28521
Ion Ratio	Lower	Upper	
128	100		
129	12.4	8.9	13.3
127	13.3	10.1	15.1



#10
Hexachlorobutadiene
Concen: 0.096 ng
RT: 10.762 min Scan# 761
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:	225	Resp:	6014
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.8	50.4	75.6









#11
2-Methylnaphthalene-d10
Concen: 0.096 ng
RT: 12.016 min Scan# 934
Delta R.T. -0.005 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:152 Resp: 14472
Ion Ratio Lower Upper
152 100
151 22.2 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.100 ng
RT: 12.092 min Scan# 951
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:142 Resp: 17369
Ion Ratio Lower Upper
142 100
141 88.5 69.9 104.9
115 36.8 27.0 40.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.281 min Scan# 1184
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:164 Resp: 52895
Ion Ratio Lower Upper
164 100
162 102.3 79.4 119.0
160 55.2 46.6 70.0



#14
2,4,6-Tribromophenol
Concen: 0.153 ng
RT: 15.791 min Scan# 1303
Delta R.T. 0.013 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:330 Resp: 1995
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 33.1 34.4 51.6#









#15
2-Fluorobiphenyl
Concen: 0.090 ng
RT: 12.911 min Scan# 1076
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:172	Resp:	22195
Ion Ratio	Lower	Upper
172	100	
171	33.7	27.4 41.2
170	23.1	18.9 28.3



#16
Acenaphthylene
Concen: 0.089 ng
RT: 14.002 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:152	Resp:	21651
Ion Ratio	Lower	Upper
152	100	
151	19.1	15.4 23.0
153	13.3	10.4 15.6









#17
Acenaphthene
Concen: 0.094 ng
RT: 14.345 min Scan# 1189
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:154 Resp: 15447
Ion Ratio Lower Upper
154 100
153 113.9 91.0 136.4
152 65.8 55.8 83.6



#18
Fluorene
Concen: 0.096 ng
RT: 15.335 min Scan# 1267
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:166 Resp: 18957
Ion Ratio Lower Upper
166 100
165 98.2 77.6 116.4
167 13.8 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.030 min Scan# 1404
Delta R.T. -0.012 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:188 Resp: 106217
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.7 6.9 10.3#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.030 ng
RT: 15.449 min Scan# 1276
Delta R.T. 0.013 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:198 Resp: 183
Ion Ratio Lower Upper
198 100
182 82.9 17.5 26.3#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.082 ng
RT: 16.239 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:248 Resp: 5414
Ion Ratio Lower Upper
248 100
250 101.7 71.8 107.6
141 48.2 69.7 104.5#



#22
Hexachlorobenzene
Concen: 0.094 ng
RT: 16.336 min Scan# 1347
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:284 Resp: 7443
Ion Ratio Lower Upper
284 100
142 35.5 30.7 46.1
249 29.7 25.8 38.8









#23
Atrazine
Concen: 0.119 ng
RT: 16.519 min Scan# 1362
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:200 Resp: 3144
Ion Ratio Lower Upper
200 100
173 27.1 22.6 34.0
215 50.3 38.2 57.4



#24
Pentachlorophenol
Concen: 0.087 ng
RT: 16.713 min Scan# 1378
Delta R.T. 0.012 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:266 Resp: 1809
Ion Ratio Lower Upper
266 100
264 62.3 50.2 75.2
268 62.6 51.8 77.8









#25
Phenanthrene
Concen: 0.089 ng
RT: 17.078 min Scan# 1408
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:178 Resp: 30864
Ion Ratio Lower Upper
178 100
176 19.0 15.0 22.6
179 15.8 12.3 18.5



#26
Anthracene
Concen: 0.085 ng
RT: 17.164 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:178 Resp: 24415
Ion Ratio Lower Upper
178 100
176 18.5 14.6 21.8
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.092 ng
RT: 19.078 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:212 Resp: 27326
Ion Ratio Lower Upper
212 100
106 14.0 8.8 13.2#
104 7.9 5.0 7.4#



#28
Fluoranthene
Concen: 0.089 ng
RT: 19.108 min Scan# 1705
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:202 Resp: 33477
Ion Ratio Lower Upper
202 100
101 11.1 7.5 11.3
203 16.9 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.249 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:240 Resp: 90428
Ion Ratio Lower Upper
240 100
120 8.0 9.0 13.4#
236 26.1 21.7 32.5



#30
Pyrene
Concen: 0.129 ng
RT: 19.475 min Scan# 1779
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:202 Resp: 33746
Ion Ratio Lower Upper
202 100
200 20.5 16.3 24.5
203 17.5 13.9 20.9









#31
Terphenyl-d14
Concen: 0.119 ng
RT: 19.688 min Scan# 1822
Delta R.T. -0.005 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:244 Resp: 21808
Ion Ratio Lower Upper
244 100
212 8.1 6.1 9.1
122 15.2 8.4 12.6#



#32
Benzo(a)anthracene
Concen: 0.118 ng
RT: 21.227 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:228 Resp: 28189
Ion Ratio Lower Upper
228 100
226 27.6 21.2 31.8
229 20.0 15.6 23.4









#33
Chrysene
Concen: 0.102 ng
RT: 21.281 min Scan# 2006
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	228	Resp:	30004
Ion Ratio	Lower	Upper	
228	100		
226	30.2	23.1	34.7
229	19.7	15.8	23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.228 ng
RT: 21.185 min Scan# 1997
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:	149	Resp:	10214
Ion Ratio	Lower	Upper	
149	100		
167	26.8	21.4	32.2
279	3.7	3.1	4.7









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.505 min Scan# 2433
 Delta R.T. -0.004 min
 Lab File: BN016559.D
 Acq: 28 Sep 2021 16:34




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:264 Resp: 86020
 Ion Ratio Lower Upper
 264 100
 260 25.2 19.9 29.9
 265 25.6 20.7 31.1



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.087 ng
 RT: 25.798 min Scan# 3103
 Delta R.T. 0.003 min
 Lab File: BN016559.D
 Acq: 28 Sep 2021 16:34

Tgt Ion:276 Resp: 23550
 Ion Ratio Lower Upper
 276 100
 138 28.8 18.2 27.2#
 277 24.8 19.8 29.8









#37
Benzo(b)fluoranthene
Concen: 0.105 ng
RT: 22.824 min Scan# 2234
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 27876
Ion Ratio Lower Upper
252 100
253 24.5 17.8 26.6
125 15.4 7.3 10.9#



#38
Benzo(k)fluoranthene
Concen: 0.086 ng
RT: 22.872 min Scan# 2248
Delta R.T. 0.003 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:252 Resp: 26741
Ion Ratio Lower Upper
252 100
253 25.1 17.8 26.6
125 20.6 7.4 11.0#









#39
Benzo(a)pyrene
Concen: 0.131 ng
RT: 23.406 min Scan# 2404
Delta R.T. -0.000 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 27807
Ion Ratio Lower Upper
252 100
253 25.7 17.9 26.9
125 20.2 8.2 12.4#



#40
Dibenzo(a,h)anthracene
Concen: 0.069 ng
RT: 25.822 min Scan# 3110
Delta R.T. 0.003 min
Lab File: BN016559.D
Acq: 28 Sep 2021 16:34

Tgt Ion:278 Resp: 16633
Ion Ratio Lower Upper
278 100
139 32.2 12.1 18.1#
279 27.4 19.2 28.8





#41
 Benzo(g,h,i)perylene
 Concen: 0.107 ng
 RT: 26.493 min Scan# 3306
 Delta R.T. -0.000 min
 Lab File: BN016559.D
 Acq: 28 Sep 2021 16:34

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	276	Resp:	23827
Ion Ratio	Lower	Upper	
276	100		
277	24.7	18.9	28.3
138	29.9	15.7	23.5#

