

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092821\
 Data File : BN016563.D
 Acq On : 28 Sep 2021 19:00
 Operator : CG/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Sep 29 01:18:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 28 18:30:42 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.660	152	30896	0.400	ng	0.00
7) Naphthalene-d8	10.419	136	130830	0.400	ng	# 0.00
13) Acenaphthene-d10	14.281	164	66337	0.400	ng	0.00
19) Phenanthrene-d10	17.042	188	123267	0.400	ng	# 0.00
29) Chrysene-d12	21.248	240	223314	0.400	ng	# 0.00
35) Perylene-d12	23.508	264	201759	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.272	112	132382	2.105	ng	0.00
5) Phenol-d6	6.840	99	136244	1.751	ng	0.00
8) Nitrobenzene-d5	8.797	82	121084	1.609	ng	0.00
11) 2-Methylnaphthalene-d10	12.020	152	325654	1.768	ng	0.00
14) 2,4,6-Tribromophenol	15.778	330	52129	3.181	ng	0.00
15) 2-Fluorobiphenyl	12.911	172	458333	1.489	ng	0.00
27) Fluoranthene-d10	19.077	212	573925	1.666	ng	0.00
31) Terphenyl-d14	19.688	244	439592	0.974	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.271	88	25098	1.760	ng	# 54
3) n-Nitrosodimethylamine	3.594	42	44382	1.620	ng	# 60
6) bis(2-Chloroethyl)ether	7.098	93	142766	1.738	ng	91
9) Naphthalene	10.470	128	603018	1.654	ng	100
10) Hexachlorobutadiene	10.761	225	117001	1.527	ng	# 99
12) 2-Methylnaphthalene	12.091	142	380766	1.779	ng	100
16) Acenaphthylene	14.002	152	529379	1.733	ng	100
17) Acenaphthene	14.345	154	343698	1.672	ng	# 92
18) Fluorene	15.334	166	416007	1.686	ng	100
20) 4,6-Dinitro-2-methylph...	15.423	198	15808	2.235	ng	# 81
21) 4-Bromophenyl-phenylether	16.238	248	128577	1.682	ng	# 72
22) Hexachlorobenzene	16.348	284	146573	1.593	ng	94
23) Atrazine	16.518	200	95995	3.134	ng	96
24) Pentachlorophenol	16.689	266	53943	2.244	ng	99
25) Phenanthrene	17.078	178	634525	1.583	ng	100
26) Anthracene	17.163	178	580203	1.750	ng	100
28) Fluoranthene	19.107	202	701657	1.603	ng	# 98
30) Pyrene	19.475	202	716881	1.109	ng	100
32) Benzo(a)anthracene	21.227	228	727735	1.228	ng	99
33) Chrysene	21.280	228	725755	1.003	ng	98
34) Bis(2-ethylhexyl)phtha...	21.185	149	237695	2.145	ng	100
36) Indeno(1,2,3-cd)pyrene	25.795	276	737668	1.163	ng	97
37) Benzo(b)fluoranthene	22.827	252	714974	1.146	ng	99
38) Benzo(k)fluoranthene	22.871	252	717737	0.986	ng	99
39) Benzo(a)pyrene	23.405	252	648838	1.307	ng	98
40) Dibenzo(a,h)anthracene	25.819	278	588828	1.045	ng	97
41) Benzo(g,h,i)perylene	26.493	276	633660	1.218	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092821\
Data File : BN016563.D
Acq On : 28 Sep 2021 19:00
Operator : CG/JU
Sample : SSTDIC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC1.6

Quant Time: Sep 29 01:18:10 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 28 18:30:42 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.660 min Scan# 497
 Delta R.T. -0.000 min
 Lab File: BN016563.D
 Acq: 28 Sep 2021 19:00




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:152 Resp: 30896
 Ion Ratio Lower Upper
 152 100
 150 157.0 125.1 187.7
 115 57.9 45.8 68.8



#2
 1,4-Dioxane
 Concen: 1.760 ng
 RT: 3.271 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN016563.D
 Acq: 28 Sep 2021 19:00

Tgt Ion: 88 Resp: 25098
 Ion Ratio Lower Upper
 88 100
 43 28.9 75.8 113.8#
 58 76.8 81.6 122.4#









#3
n-Nitrosodimethylamine
Concen: 1.620 ng
RT: 3.594 min Scan# 56
Delta R.T. -0.019 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 44382
Ion Ratio Lower Upper
42 100
74 136.2 77.5 116.3#
44 7.9 0.0 0.0#



#4
2-Fluorophenol
Concen: 2.105 ng
RT: 5.272 min Scan# 238
Delta R.T. -0.010 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion: 112 Resp: 132382
Ion Ratio Lower Upper
112 100
64 55.1 48.4 72.6
63 28.0 25.4 38.2









#5
Phenol-d6
Concen: 1.751 ng
RT: 6.840 min Scan# 408
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 136244
Ion Ratio Lower Upper
99 100
42 17.7 19.6 29.4#
71 29.6 24.2 36.2



#6
bis(2-Chloroethyl)ether
Concen: 1.738 ng
RT: 7.098 min Scan# 436
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion: 93 Resp: 142766
Ion Ratio Lower Upper
93 100
63 75.8 70.0 105.0
95 32.7 26.3 39.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.419 min Scan# 734
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	130830
Ion Ratio	Lower	Upper	
136	100		
137	10.9	8.8	13.2
54	8.8	5.8	8.6#
68	7.1	3.6	5.4#



#8
Nitrobenzene-d5
Concen: 1.609 ng
RT: 8.797 min Scan# 606
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion:	82	Resp:	121084
Ion Ratio	Lower	Upper	
82	100		
128	39.6	27.4	41.2
54	49.6	58.5	87.7#









#9
Naphthalene
Concen: 1.654 ng
RT: 10.470 min Scan# 738
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 603018
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 12.8 10.1 15.1



#10
Hexachlorobutadiene
Concen: 1.527 ng
RT: 10.761 min Scan# 761
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion:225 Resp: 117001
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 50.4 75.6









#11
2-Methylnaphthalene-d10
Concen: 1.768 ng
RT: 12.020 min Scan# 935
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 325654
Ion Ratio Lower Upper
152 100
151 20.5 16.6 25.0



#12
2-Methylnaphthalene
Concen: 1.779 ng
RT: 12.091 min Scan# 951
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion:142 Resp: 380766
Ion Ratio Lower Upper
142 100
141 87.8 69.9 104.9
115 33.4 27.0 40.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.281 min Scan# 1184
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 66337
Ion Ratio Lower Upper
164 100
162 102.1 79.4 119.0
160 48.8 46.6 70.0



#14
2,4,6-Tribromophenol
Concen: 3.181 ng
RT: 15.778 min Scan# 1302
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion:330 Resp: 52129
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 34.8 34.4 51.6









#15
2-Fluorobiphenyl
Concen: 1.489 ng
RT: 12.911 min Scan# 1076
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 458333
Ion Ratio Lower Upper
172 100
171 32.9 27.4 41.2
170 21.5 18.9 28.3



#16
Acenaphthylene
Concen: 1.733 ng
RT: 14.002 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion:152 Resp: 529379
Ion Ratio Lower Upper
152 100
151 19.3 15.4 23.0
153 13.1 10.4 15.6









#17
Acenaphthene
Concen: 1.672 ng
RT: 14.345 min Scan# 1189
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 343698
Ion Ratio Lower Upper
154 100
153 111.5 91.0 136.4
152 55.2 55.8 83.6#



#18
Fluorene
Concen: 1.686 ng
RT: 15.334 min Scan# 1267
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion:166 Resp: 416007
Ion Ratio Lower Upper
166 100
165 97.3 77.6 116.4
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.042 min Scan# 1405
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:188 Resp: 123267
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.6 6.9 10.3#



#20
4,6-Dinitro-2-methylphenol
Concen: 2.235 ng
RT: 15.423 min Scan# 1274
Delta R.T. -0.013 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion:198 Resp: 15808
Ion Ratio Lower Upper
198 100
182 12.7 17.5 26.3#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 1.682 ng
RT: 16.238 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 128577
Ion Ratio Lower Upper
248 100
250 100.3 71.8 107.6
141 45.0 69.7 104.5#



#22
Hexachlorobenzene
Concen: 1.593 ng
RT: 16.348 min Scan# 1348
Delta R.T. 0.012 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00


Tgt Ion:284 Resp: 146573
Ion Ratio Lower Upper
284 100
142 34.8 30.7 46.1
249 29.4 25.8 38.8








#23
Atrazine
Concen: 3.134 ng
RT: 16.518 min Scan# 1362
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6





Tgt Ion:200 Resp: 95995
Ion Ratio Lower Upper
200 100
173 23.4 22.6 34.0
215 48.5 38.2 57.4




#24
Pentachlorophenol
Concen: 2.244 ng
RT: 16.689 min Scan# 1376
Delta R.T. -0.013 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00



Tgt Ion:266 Resp: 53943
Ion Ratio Lower Upper
266 100
264 62.8 50.2 75.2
268 64.0 51.8 77.8









#25
Phenanthrene
Concen: 1.583 ng
RT: 17.078 min Scan# 1408
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 634525
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.3 12.3 18.5



#26
Anthracene
Concen: 1.750 ng
RT: 17.163 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion:178 Resp: 580203
Ion Ratio Lower Upper
178 100
176 18.2 14.6 21.8
179 15.4 12.2 18.4









#27
Fluoranthene-d10
Concen: 1.666 ng
RT: 19.077 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:212 Resp: 573925
Ion Ratio Lower Upper
212 100
106 13.9 8.8 13.2#
104 7.6 5.0 7.4#



#28
Fluoranthene
Concen: 1.603 ng
RT: 19.107 min Scan# 1705
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion:202 Resp: 701657
Ion Ratio Lower Upper
202 100
101 11.5 7.5 11.3#
203 17.4 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.248 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:240 Resp: 223314
Ion Ratio Lower Upper
240 100
120 7.4 9.0 13.4#
236 25.9 21.7 32.5



#30
Pyrene
Concen: 1.109 ng
RT: 19.475 min Scan# 1779
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion:202 Resp: 716881
Ion Ratio Lower Upper
202 100
200 20.7 16.3 24.5
203 17.5 13.9 20.9









#31
Terphenyl-d14
Concen: 0.974 ng
RT: 19.688 min Scan# 1822
Delta R.T. -0.005 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:244 Resp: 439592
Ion Ratio Lower Upper
244 100
212 7.7 6.1 9.1
122 14.1 8.4 12.6#



#32
Benzo(a)anthracene
Concen: 1.228 ng
RT: 21.227 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion:228 Resp: 727735
Ion Ratio Lower Upper
228 100
226 27.5 21.2 31.8
229 19.6 15.6 23.4









#33
Chrysene
Concen: 1.003 ng
RT: 21.280 min Scan# 2006
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:228 Resp: 725755
Ion Ratio Lower Upper
228 100
226 29.9 23.1 34.7
229 19.3 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 2.145 ng
RT: 21.185 min Scan# 1997
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Tgt Ion:149 Resp: 237695
Ion Ratio Lower Upper
149 100
167 26.9 21.4 32.2
279 3.6 3.1 4.7









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.508 min Scan# 2434
 Delta R.T. -0.000 min
 Lab File: BN016563.D
 Acq: 28 Sep 2021 19:00




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:264	Resp:	201759
Ion Ratio	Lower	Upper
264	100	
260	24.8	19.9 29.9
265	22.4	20.7 31.1



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 1.163 ng
 RT: 25.795 min Scan# 3102
 Delta R.T. -0.000 min
 Lab File: BN016563.D
 Acq: 28 Sep 2021 19:00

Tgt Ion:276	Resp:	737668
Ion Ratio	Lower	Upper
276	100	
138	25.6	18.2 27.2
277	24.9	19.8 29.8









#37
Benzo(b)fluoranthene
Concen: 1.146 ng
RT: 22.827 min Scan# 2235
Delta R.T. 0.003 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:252 Resp: 714974
Ion Ratio Lower Upper
252 100
253 22.2 17.8 26.6
125 10.2 7.3 10.9



#38
Benzo(k)fluoranthene
Concen: 0.986 ng
RT: 22.871 min Scan# 2248
Delta R.T. 0.003 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00


Tgt Ion:252 Resp: 717737
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 10.7 7.4 11.0








#39
Benzo(a)pyrene
Concen: 1.307 ng
RT: 23.405 min Scan# 2404
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00


Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6





Tgt Ion:252 Resp: 648838
Ion Ratio Lower Upper
252 100
253 21.9 17.9 26.9
125 11.3 8.2 12.4




#40
Dibenzo(a,h)anthracene
Concen: 1.045 ng
RT: 25.819 min Scan# 3109
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00




Tgt Ion:278 Resp: 588828
Ion Ratio Lower Upper
278 100
139 17.3 12.1 18.1
279 23.5 19.2 28.8





#41
Benzo(g,h,i)perylene
Concen: 1.218 ng
RT: 26.493 min Scan# 3306
Delta R.T. -0.000 min
Lab File: BN016563.D
Acq: 28 Sep 2021 19:00

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6



Tgt Ion:276 Resp: 633660
Ion Ratio Lower Upper
276 100
277 23.9 18.9 28.3
138 22.0 15.7 23.5

