

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092821\
 Data File : BN016564.D
 Acq On : 28 Sep 2021 19:36
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Sep 29 01:18:21 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 28 18:30:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.661	152	30371	0.400	ng	0.00
7) Naphthalene-d8	10.419	136	123949	0.400	ng	# 0.00
13) Acenaphthene-d10	14.281	164	64728	0.400	ng	0.00
19) Phenanthrene-d10	17.042	188	122565	0.400	ng	# 0.00
29) Chrysene-d12	21.249	240	119572	0.400	ng	# 0.00
35) Perylene-d12	23.508	264	114780	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.282	112	242979	3.931	ng	0.00
5) Phenol-d6	6.840	99	258741	3.384	ng	0.00
8) Nitrobenzene-d5	8.797	82	247598	3.473	ng	0.00
11) 2-Methylnaphthalene-d10	12.016	152	591409	3.388	ng	0.00
14) 2,4,6-Tribromophenol	15.779	330	112859	7.059	ng	0.00
15) 2-Fluorobiphenyl	12.911	172	836215	2.784	ng	0.00
27) Fluoranthene-d10	19.078	212	1138289	3.323	ng	0.00
31) Terphenyl-d14	19.693	244	900648	3.726	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.263	88	43604	3.110	ng	# 45
3) n-Nitrosodimethylamine	3.604	42	82094	3.049	ng	# 56
6) bis(2-Chloroethyl)ether	7.098	93	259262	3.211	ng	91
9) Naphthalene	10.470	128	1075620	3.114	ng	100
10) Hexachlorobutadiene	10.762	225	203711	2.806	ng	# 99
12) 2-Methylnaphthalene	12.092	142	683024	3.368	ng	99
16) Acenaphthylene	14.002	152	1023465	3.434	ng	100
17) Acenaphthene	14.345	154	649240	3.238	ng	# 90
18) Fluorene	15.335	166	791631	3.288	ng	99
20) 4,6-Dinitro-2-methylph...	15.423	198	42760	6.081	ng	# 74
21) 4-Bromophenyl-phenylether	16.239	248	255517	3.362	ng	# 73
22) Hexachlorobenzene	16.348	284	276628	3.024	ng	95
23) Atrazine	16.519	200	202609	6.653	ng	97
24) Pentachlorophenol	16.689	266	132187	5.530	ng	99
25) Phenanthrene	17.078	178	1236135	3.102	ng	100
26) Anthracene	17.164	178	1137617	3.452	ng	100
28) Fluoranthene	19.113	202	1345946	3.093	ng	# 98
30) Pyrene	19.475	202	1395172	4.029	ng	99
32) Benzo(a)anthracene	21.227	228	1371091	4.322	ng	99
33) Chrysene	21.281	228	1330983	3.435	ng	98
34) Bis(2-ethylhexyl)phtha...	21.185	149	516819	8.710	ng	99
36) Indeno(1,2,3-cd)pyrene	25.798	276	1436624	3.982	ng	97
37) Benzo(b)fluoranthene	22.827	252	1383383	3.898	ng	99
38) Benzo(k)fluoranthene	22.872	252	1321755	3.193	ng	99
39) Benzo(a)pyrene	23.406	252	1258274	4.457	ng	98
40) Dibenzo(a,h)anthracene	25.822	278	1166295	3.638	ng	98
41) Benzo(g,h,i)perylene	26.497	276	1236499	4.179	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092821\
Data File : BN016564.D
Acq On : 28 Sep 2021 19:36
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Sep 29 01:18:21 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 28 18:30:42 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.661 min Scan# 497
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 30371
Ion Ratio Lower Upper
152 100
150 160.2 125.1 187.7
115 57.6 45.8 68.8



#2
1,4-Dioxane
Concen: 3.110 ng
RT: 3.263 min Scan# 20
Delta R.T. -0.009 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion: 88 Resp: 43604
Ion Ratio Lower Upper
88 100
43 23.6 75.8 113.8#
58 63.6 81.6 122.4#









#3
n-Nitrosodimethylamine
Concen: 3.049 ng
RT: 3.604 min Scan# 57
Delta R.T. -0.009 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 82094
Ion Ratio Lower Upper
42 100
74 139.8 77.5 116.3#
44 7.1 0.0 0.0#



#4
2-Fluorophenol
Concen: 3.931 ng
RT: 5.282 min Scan# 239
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion: 112 Resp: 242979
Ion Ratio Lower Upper
112 100
64 54.9 48.4 72.6
63 27.9 25.4 38.2









#5
Phenol-d6
Concen: 3.384 ng
RT: 6.840 min Scan# 408
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 258741
Ion Ratio Lower Upper
99 100
42 17.5 19.6 29.4#
71 29.4 24.2 36.2



#6
bis(2-Chloroethyl)ether
Concen: 3.211 ng
RT: 7.098 min Scan# 436
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion: 93 Resp: 259262
Ion Ratio Lower Upper
93 100
63 75.5 70.0 105.0
95 32.8 26.3 39.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.419 min Scan# 734
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	123949
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.8	13.2
54	8.4	5.8	8.6
68	6.7	3.6	5.4



#8
Nitrobenzene-d5
Concen: 3.473 ng
RT: 8.797 min Scan# 606
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:	82	Resp:	247598
Ion Ratio	Lower	Upper	
82	100		
128	41.2	27.4	41.2
54	48.9	58.5	87.7









#9
Naphthalene
Concen: 3.114 ng
RT: 10.470 min Scan# 738
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 1075620
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 12.8 10.1 15.1



#10
Hexachlorobutadiene
Concen: 2.806 ng
RT: 10.762 min Scan# 761
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:225 Resp: 203711
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 50.4 75.6









#11
2-Methylnaphthalene-d10
Concen: 3.388 ng
RT: 12.016 min Scan# 934
Delta R.T. -0.005 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 591409
Ion Ratio Lower Upper
152 100
151 20.3 16.6 25.0



#12
2-Methylnaphthalene
Concen: 3.368 ng
RT: 12.092 min Scan# 951
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:142 Resp: 683024
Ion Ratio Lower Upper
142 100
141 87.8 69.9 104.9
115 33.3 27.0 40.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.281 min Scan# 1184
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 64728
Ion Ratio Lower Upper
164 100
162 102.1 79.4 119.0
160 48.1 46.6 70.0



#14
2,4,6-Tribromophenol
Concen: 7.059 ng
RT: 15.779 min Scan# 1302
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:330 Resp: 112859
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 34.1 34.4 51.6#









#15
2-Fluorobiphenyl
Concen: 2.784 ng
RT: 12.911 min Scan# 1076
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 836215
Ion Ratio Lower Upper
172 100
171 33.1 27.4 41.2
170 21.6 18.9 28.3



#16
Acenaphthylene
Concen: 3.434 ng
RT: 14.002 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:152 Resp: 1023465
Ion Ratio Lower Upper
152 100
151 19.2 15.4 23.0
153 13.1 10.4 15.6









#17
Acenaphthene
Concen: 3.238 ng
RT: 14.345 min Scan# 1189
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 649240
Ion Ratio Lower Upper
154 100
153 110.0 91.0 136.4
152 53.2 55.8 83.6#



#18
Fluorene
Concen: 3.288 ng
RT: 15.335 min Scan# 1267
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:166 Resp: 791631
Ion Ratio Lower Upper
166 100
165 96.2 77.6 116.4
167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.042 min Scan# 1405
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 122565
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.7 6.9 10.3#



#20
4,6-Dinitro-2-methylphenol
Concen: 6.081 ng
RT: 15.423 min Scan# 1274
Delta R.T. -0.013 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:198 Resp: 42760
Ion Ratio Lower Upper
198 100
182 9.6 17.5 26.3#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 3.362 ng
RT: 16.239 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 255517
Ion Ratio Lower Upper
248 100
250 99.3 71.8 107.6
141 45.3 69.7 104.5#



#22
Hexachlorobenzene
Concen: 3.024 ng
RT: 16.348 min Scan# 1348
Delta R.T. 0.012 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:284 Resp: 276628
Ion Ratio Lower Upper
284 100
142 34.8 30.7 46.1
249 29.8 25.8 38.8









#23
 Atrazine
 Concen: 6.653 ng
 RT: 16.519 min Scan# 1362
 Delta R.T. -0.000 min
 Lab File: BN016564.D
 Acq: 28 Sep 2021 19:36




Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC3.2


Tgt Ion:200 Resp: 202609
 Ion Ratio Lower Upper
 200 100
 173 24.1 22.6 34.0
 215 48.2 38.2 57.4



#24
 Pentachlorophenol
 Concen: 5.530 ng
 RT: 16.689 min Scan# 1376
 Delta R.T. -0.012 min
 Lab File: BN016564.D
 Acq: 28 Sep 2021 19:36

Tgt Ion:266 Resp: 132187
 Ion Ratio Lower Upper
 266 100
 264 62.7 50.2 75.2
 268 64.0 51.8 77.8









#25
Phenanthrene
Concen: 3.102 ng
RT: 17.078 min Scan# 1408
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 1236135
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.3 12.3 18.5



#26
Anthracene
Concen: 3.452 ng
RT: 17.164 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:178 Resp: 1137617
Ion Ratio Lower Upper
178 100
176 18.2 14.6 21.8
179 15.6 12.2 18.4









#27
Fluoranthene-d10
Concen: 3.323 ng
RT: 19.078 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 1138289
Ion Ratio Lower Upper
212 100
106 14.1 8.8 13.2#
104 7.7 5.0 7.4#



#28
Fluoranthene
Concen: 3.093 ng
RT: 19.113 min Scan# 1706
Delta R.T. 0.005 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:202 Resp: 1345946
Ion Ratio Lower Upper
202 100
101 11.6 7.5 11.3#
203 17.6 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.249 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 119572
Ion Ratio Lower Upper
240 100
120 7.6 9.0 13.4#
236 26.1 21.7 32.5



#30
Pyrene
Concen: 4.029 ng
RT: 19.475 min Scan# 1779
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:202 Resp: 1395172
Ion Ratio Lower Upper
202 100
200 20.6 16.3 24.5
203 17.7 13.9 20.9









#31
Terphenyl-d14
Concen: 3.726 ng
RT: 19.693 min Scan# 1823
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:244 Resp: 900648
Ion Ratio Lower Upper
244 100
212 7.3 6.1 9.1
122 11.9 8.4 12.6



#32
Benzo(a)anthracene
Concen: 4.322 ng
RT: 21.227 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:228 Resp: 1371091
Ion Ratio Lower Upper
228 100
226 27.4 21.2 31.8
229 19.7 15.6 23.4









#33
Chrysene
Concen: 3.435 ng
RT: 21.281 min Scan# 2006
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228 Resp: 1330983
Ion Ratio Lower Upper
228 100
226 30.1 23.1 34.7
229 19.3 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 8.710 ng
RT: 21.185 min Scan# 1997
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:149 Resp: 516819
Ion Ratio Lower Upper
149 100
167 27.2 21.4 32.2
279 3.7 3.1 4.7









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.508 min Scan# 2434
 Delta R.T. -0.000 min
 Lab File: BN016564.D
 Acq: 28 Sep 2021 19:36




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:264 Resp: 114780
 Ion Ratio Lower Upper
 264 100
 260 24.7 19.9 29.9
 265 23.9 20.7 31.1



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 3.982 ng
 RT: 25.798 min Scan# 3103
 Delta R.T. 0.003 min
 Lab File: BN016564.D
 Acq: 28 Sep 2021 19:36

Tgt Ion:276 Resp: 1436624
 Ion Ratio Lower Upper
 276 100
 138 25.7 18.2 27.2
 277 25.0 19.8 29.8









#37
Benzo(b)fluoranthene
Concen: 3.898 ng
RT: 22.827 min Scan# 2235
Delta R.T. 0.003 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 1383383
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 10.1 7.3 10.9



#38
Benzo(k)fluoranthene
Concen: 3.193 ng
RT: 22.872 min Scan# 2248
Delta R.T. 0.003 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:252 Resp: 1321755
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 10.5 7.4 11.0









#39
Benzo(a)pyrene
Concen: 4.457 ng
RT: 23.406 min Scan# 2404
Delta R.T. -0.000 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 1258274
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 11.2 8.2 12.4



#40
Dibenzo(a,h)anthracene
Concen: 3.638 ng
RT: 25.822 min Scan# 3110
Delta R.T. 0.003 min
Lab File: BN016564.D
Acq: 28 Sep 2021 19:36

Tgt Ion:278 Resp: 1166295
Ion Ratio Lower Upper
278 100
139 16.9 12.1 18.1
279 23.6 19.2 28.8





#41
 Benzo(g,h,i)perylene
 Concen: 4.179 ng
 RT: 26.497 min Scan# 3307
 Delta R.T. 0.003 min
 Lab File: BN016564.D
 Acq: 28 Sep 2021 19:36

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Tgt Ion:276 Resp: 1236499

Ion	Ratio	Lower	Upper
276	100		
277	23.8	18.9	28.3
138	22.0	15.7	23.5

