

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092821\
 Data File : BN016569.D
 Acq On : 28 Sep 2021 23:51
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Sep 29 01:56:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 29 01:39:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.651	152	32757	0.400	ng	0.00
7) Naphthalene-d8	10.419	136	141784	0.400	ng	# 0.00
13) Acenaphthene-d10	14.281	164	71143	0.400	ng	0.00
19) Phenanthrene-d10	17.042	188	133420	0.400	ng	# 0.00
29) Chrysene-d12	21.249	240	120163	0.400	ng	# 0.00
35) Perylene-d12	23.505	264	117298	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.282	112	29712	0.365	ng	0.00
5) Phenol-d6	6.831	99	29090	0.355	ng	0.00
8) Nitrobenzene-d5	8.797	82	24344	0.323	ng	0.00
11) 2-Methylnaphthalene-d10	12.016	152	78169	0.388	ng	0.00
14) 2,4,6-Tribromophenol	15.778	330	9452	0.304	ng	0.00
15) 2-Fluorobiphenyl	12.911	172	110685	0.378	ng	0.00
27) Fluoranthene-d10	19.078	212	138466	0.384	ng	0.00
31) Terphenyl-d14	19.693	244	105059	0.406	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.272	88	4720	0.335	ng	# 58
3) n-Nitrosodimethylamine	3.613	42	9230	0.355	ng	# 54
6) bis(2-Chloroethyl)ether	7.098	93	33660	0.390	ng	90
9) Naphthalene	10.470	128	146550	0.383	ng	99
10) Hexachlorobutadiene	10.761	225	28650	0.379	ng	# 98
12) 2-Methylnaphthalene	12.092	142	93008	0.391	ng	99
16) Acenaphthylene	14.002	152	107829	0.342	ng	100
17) Acenaphthene	14.345	154	80753	0.377	ng	93
18) Fluorene	15.334	166	97124	0.372	ng	99
20) 4,6-Dinitro-2-methylph...	15.436	198	1462	0.348	ng	99
21) 4-Bromophenyl-phenylether	16.239	248	29948	0.380	ng	# 74
22) Hexachlorobenzene	16.348	284	36436	0.387	ng	94
23) Atrazine	16.518	200	17261	0.356	ng	96
24) Pentachlorophenol	16.701	266	7500	0.371	ng	100
25) Phenanthrene	17.078	178	154118	0.384	ng	100
26) Anthracene	17.163	178	124119	0.356	ng	100
28) Fluoranthene	19.107	202	172198	0.392	ng	98
30) Pyrene	19.475	202	166898	0.407	ng	100
32) Benzo(a)anthracene	21.227	228	139890	0.376	ng	99
33) Chrysene	21.280	228	157025	0.413	ng	99
34) Bis(2-ethylhexyl)phtha...	21.185	149	39453	0.356	ng	99
36) Indeno(1,2,3-cd)pyrene	25.798	276	145890	0.390	ng	97
37) Benzo(b)fluoranthene	22.824	252	148087	0.378	ng	99
38) Benzo(k)fluoranthene	22.868	252	150911	0.395	ng	# 97
39) Benzo(a)pyrene	23.406	252	132825	0.370	ng	99
40) Dibenzo(a,h)anthracene	25.819	278	113000	0.435	ng	# 96
41) Benzo(g,h,i)perylene	26.493	276	131144	0.386	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092821\
 Data File : BN016569.D
 Acq On : 28 Sep 2021 23:51
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Sep 29 01:56:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 29 01:39:43 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.651 min Scan# 496
 Delta R.T. -0.010 min
 Lab File: BN016569.D
 Acq: 28 Sep 2021 23:51




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:152 Resp: 32757
 Ion Ratio Lower Upper
 152 100
 150 157.2 125.1 187.7
 115 59.9 45.8 68.8



#2
 1,4-Dioxane
 Concen: 0.335 ng
 RT: 3.272 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN016569.D
 Acq: 28 Sep 2021 23:51

Tgt Ion: 88 Resp: 4720
 Ion Ratio Lower Upper
 88 100
 43 31.0 75.8 113.8#
 58 81.5 81.6 122.4#









#3
n-Nitrosodimethylamine
Concen: 0.355 ng
RT: 3.613 min Scan# 58
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 9230
Ion Ratio Lower Upper
42 100
74 141.7 77.5 116.3#
44 8.5 0.0 0.0#



#4
2-Fluorophenol
Concen: 0.365 ng
RT: 5.282 min Scan# 239
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion: 112 Resp: 29712
Ion Ratio Lower Upper
112 100
64 54.7 48.4 72.6
63 27.7 25.4 38.2









#5
Phenol-d6
Concen: 0.355 ng
RT: 6.831 min Scan# 407
Delta R.T. -0.009 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 29090
Ion Ratio Lower Upper
99 100
42 17.5 19.6 29.4#
71 29.7 24.2 36.2



#6
bis(2-Chloroethyl)ether
Concen: 0.390 ng
RT: 7.098 min Scan# 436
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion: 93 Resp: 33660
Ion Ratio Lower Upper
93 100
63 75.3 70.0 105.0
95 33.1 26.3 39.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.419 min Scan# 734
Delta R.T. 0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	141784
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.8	13.2
54	7.5	5.8	8.6
68	6.2	3.6	5.4#



#8
Nitrobenzene-d5
Concen: 0.323 ng
RT: 8.797 min Scan# 606
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:	82	Resp:	24344
Ion Ratio	Lower	Upper	
82	100		
128	41.2	27.4	41.2#
54	49.4	58.5	87.7#









#9
Naphthalene
Concen: 0.383 ng
RT: 10.470 min Scan# 738
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 146550
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 12.9 10.1 15.1



#10
Hexachlorobutadiene
Concen: 0.379 ng
RT: 10.761 min Scan# 761
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:225 Resp: 28650
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 50.4 75.6









#11
2-Methylnaphthalene-d10
Concen: 0.388 ng
RT: 12.016 min Scan# 934
Delta R.T. -0.005 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 78169
Ion Ratio Lower Upper
152 100
151 20.6 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.391 ng
RT: 12.092 min Scan# 951
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:142 Resp: 93008
Ion Ratio Lower Upper
142 100
141 87.9 69.9 104.9
115 33.2 27.0 40.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.281 min Scan# 1184
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 71143
Ion Ratio Lower Upper
164 100
162 103.1 79.4 119.0
160 47.9 46.6 70.0



#14
2,4,6-Tribromophenol
Concen: 0.304 ng
RT: 15.778 min Scan# 1302
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:330 Resp: 9452
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 34.7 34.4 51.6









#15
2-Fluorobiphenyl
Concen: 0.378 ng
RT: 12.911 min Scan# 1076
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172 Resp: 110685
Ion Ratio Lower Upper
172 100
171 33.1 27.4 41.2
170 21.6 18.9 28.3



#16
Acenaphthylene
Concen: 0.342 ng
RT: 14.002 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:152 Resp: 107829
Ion Ratio Lower Upper
152 100
151 19.4 15.4 23.0
153 13.0 10.4 15.6









#17
Acenaphthene
Concen: 0.377 ng
RT: 14.345 min Scan# 1189
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 80753
Ion Ratio Lower Upper
154 100
153 112.9 91.0 136.4
152 56.4 55.8 83.6



#18
Fluorene
Concen: 0.372 ng
RT: 15.334 min Scan# 1267
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:166 Resp: 97124
Ion Ratio Lower Upper
166 100
165 98.0 77.6 116.4
167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.042 min Scan# 1405
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 133420
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.2 6.9 10.3#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.348 ng
RT: 15.436 min Scan# 1275
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:198 Resp: 1462
Ion Ratio Lower Upper
198 100
182 21.5 17.5 26.3
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.380 ng
RT: 16.239 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 29948
Ion Ratio Lower Upper
248 100
250 97.8 71.8 107.6
141 45.7 69.7 104.5#



#22
Hexachlorobenzene
Concen: 0.387 ng
RT: 16.348 min Scan# 1348
Delta R.T. 0.012 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:284 Resp: 36436
Ion Ratio Lower Upper
284 100
142 34.6 30.7 46.1
249 29.6 25.8 38.8









#23
Atrazine
Concen: 0.356 ng
RT: 16.518 min Scan# 1362
Delta R.T. -0.001 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 17261
Ion Ratio Lower Upper
200 100
173 23.8 22.6 34.0
215 48.6 38.2 57.4



#24
Pentachlorophenol
Concen: 0.371 ng
RT: 16.701 min Scan# 1377
Delta R.T. 0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:266 Resp: 7500
Ion Ratio Lower Upper
266 100
264 62.6 50.2 75.2
268 64.2 51.8 77.8









#25
Phenanthrene
Concen: 0.384 ng
RT: 17.078 min Scan# 1408
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 154118
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.3 12.3 18.5



#26
Anthracene
Concen: 0.356 ng
RT: 17.163 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:178 Resp: 124119
Ion Ratio Lower Upper
178 100
176 18.3 14.6 21.8
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.384 ng
RT: 19.078 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 138466
Ion Ratio Lower Upper
212 100
106 13.6 8.8 13.2#
104 7.3 5.0 7.4



#28
Fluoranthene
Concen: 0.392 ng
RT: 19.107 min Scan# 1705
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:202 Resp: 172198
Ion Ratio Lower Upper
202 100
101 11.2 7.5 11.3
203 17.2 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.249 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 120163
Ion Ratio Lower Upper
240 100
120 7.1 9.0 13.4#
236 25.8 21.7 32.5



#30
Pyrene
Concen: 0.407 ng
RT: 19.475 min Scan# 1779
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:202 Resp: 166898
Ion Ratio Lower Upper
202 100
200 20.6 16.3 24.5
203 17.4 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.406 ng
 RT: 19.693 min Scan# 1823
 Delta R.T. -0.000 min
 Lab File: BN016569.D
 Acq: 28 Sep 2021 23:51




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:244	Resp:	105059
Ion Ratio	Lower	Upper
244	100	
212	7.4	6.1 9.1
122	11.7	8.4 12.6



#32
 Benzo(a)anthracene
 Concen: 0.376 ng
 RT: 21.227 min Scan# 2001
 Delta R.T. -0.000 min
 Lab File: BN016569.D
 Acq: 28 Sep 2021 23:51

Tgt Ion:228	Resp:	139890
Ion Ratio	Lower	Upper
228	100	
226	27.3	21.2 31.8
229	19.4	15.6 23.4









#33
Chrysene
Concen: 0.413 ng
RT: 21.280 min Scan# 2006
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:228 Resp: 157025
Ion Ratio Lower Upper
228 100
226 29.7 23.1 34.7
229 19.3 15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.356 ng
RT: 21.185 min Scan# 1997
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:149 Resp: 39453
Ion Ratio Lower Upper
149 100
167 27.0 21.4 32.2
279 3.5 3.1 4.7









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.505 min Scan# 2433
 Delta R.T. -0.004 min
 Lab File: BN016569.D
 Acq: 28 Sep 2021 23:51




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:	264	Resp:	117298
Ion Ratio	Lower	Upper	
264	100		
260	24.7	19.9	29.9
265	23.8	20.7	31.1



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.390 ng
 RT: 25.798 min Scan# 3103
 Delta R.T. 0.003 min
 Lab File: BN016569.D
 Acq: 28 Sep 2021 23:51

Tgt Ion:	276	Resp:	145890
Ion Ratio	Lower	Upper	
276	100		
138	25.9	18.2	27.2
277	24.9	19.8	29.8









#37
Benzo(b)fluoranthene
Concen: 0.378 ng
RT: 22.824 min Scan# 2234
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 148087
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 10.4 7.3 10.9



#38
Benzo(k)fluoranthene
Concen: 0.395 ng
RT: 22.868 min Scan# 2247
Delta R.T. -0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Tgt Ion:252 Resp: 150911
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.6
125 12.6 7.4 11.0#





#39

Benzo(a)pyrene

Concen: 0.370 ng

RT: 23.406 min Scan# 2404

Delta R.T. -0.000 min

Lab File: BN016569.D

Acq: 28 Sep 2021 23:51





Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:252	Resp: 132825
Ion Ratio	Lower Upper
252 100	
253 22.2	17.9 26.9
125 11.4	8.2 12.4



#40

Dibenzo(a,h)anthracene

Concen: 0.435 ng




RT: 25.819 min Scan# 3109


Delta R.T. -0.000 min

Lab File: BN016569.D

Acq: 28 Sep 2021 23:51

Tgt Ion:278	Resp: 113000
Ion Ratio	Lower Upper
278 100	
139 18.8	12.1 18.1#
279 23.8	19.2 28.8





#41
Benzo(g,h,i)perylene
Concen: 0.386 ng
RT: 26.493 min Scan# 3306
Delta R.T. 0.000 min
Lab File: BN016569.D
Acq: 28 Sep 2021 23:51

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:276 Resp: 131144
Ion Ratio Lower Upper
276 100
277 24.0 18.9 28.3
138 21.6 15.7 23.5

