

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092821\
 Data File : BN016570.D
 Acq On : 29 Sep 2021 01:03
 Operator : CG/JU
 Sample : SP5710
 Misc : 8270 SIM SPIKE
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Sep 29 02:03:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 29 01:39:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.651	152	29384	0.400	ng	0.00
7) Naphthalene-d8	10.419	136	154267	0.400	ng	# 0.00
13) Acenaphthene-d10	14.281	164	78516	0.400	ng	0.00
19) Phenanthrene-d10	17.042	188	146577	0.400	ng	# 0.00
29) Chrysene-d12	21.248	240	134772	0.400	ng	# 0.00
35) Perylene-d12	23.505	264	119008	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.309	112	82	0.001	ng	0.03
5) Phenol-d6	6.849	99	150	0.002	ng	0.00
8) Nitrobenzene-d5	8.708	82	4700	0.057	ng	-0.09
11) 2-Methylnaphthalene-d10	11.985	152	5113	0.023	ng	-0.04
14) 2,4,6-Tribromophenol	15.778	330	39	0.001	ng	0.00
15) 2-Fluorobiphenyl	12.911	172	144	0.000	ng	0.00
27) Fluoranthene-d10	19.082	212	21	0.000	ng	0.00
31) Terphenyl-d14	19.693	244	37	0.000	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.281	88	10972	0.867	ng	# 29
3) n-Nitrosodimethylamine	3.594	42	10905	0.467	ng	# 44
6) bis(2-Chloroethyl)ether	7.098	93	29517	0.381	ng	90
9) Naphthalene	10.470	128	164299	0.395	ng	100
10) Hexachlorobutadiene	10.761	225	30805	0.375	ng	# 99
12) 2-Methylnaphthalene	12.091	142	105060	0.406	ng	99
16) Acenaphthylene	14.002	152	127205	0.365	ng	100
17) Acenaphthene	14.345	154	88676	0.375	ng	94
18) Fluorene	15.334	166	109344	0.379	ng	100
20) 4,6-Dinitro-2-methylph...	15.423	198	1514	0.340	ng	100
21) 4-Bromophenyl-phenylether	16.238	248	33625	0.388	ng	# 74
22) Hexachlorobenzene	16.336	284	41047	0.397	ng	94
23) Atrazine	16.518	200	20079	0.370	ng	95
24) Pentachlorophenol	16.689	266	17992	0.559	ng	99
25) Phenanthrene	17.078	178	174038	0.395	ng	100
26) Anthracene	17.163	178	146897	0.383	ng	100
28) Fluoranthene	19.107	202	182531	0.379	ng	98
30) Pyrene	19.475	202	185601	0.404	ng	100
32) Benzo(a)anthracene	21.227	228	163153	0.391	ng	99
33) Chrysene	21.280	228	174251	0.409	ng	99
34) Bis(2-ethylhexyl)phtha...	21.174	149	47234	0.380	ng	100
36) Indeno(1,2,3-cd)pyrene	25.795	276	141031	0.372	ng	98
37) Benzo(b)fluoranthene	22.824	252	162570	0.409	ng	99
38) Benzo(k)fluoranthene	22.868	252	157341	0.406	ng	99
39) Benzo(a)pyrene	23.405	252	144575	0.397	ng	99
40) Dibenzo(a,h)anthracene	25.815	278	108256	0.411	ng	98
41) Benzo(g,h,i)perylene	26.489	276	133914	0.388	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092821\
 Data File : BN016570.D
 Acq On : 29 Sep 2021 01:03
 Operator : CG/JU
 Sample : SP5710
 Misc : 8270 SIM SPIKE
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Sep 29 02:03:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 29 01:39:43 2021
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.651 min Scan# 496
 Delta R.T. -0.010 min
 Lab File: BN016570.D
 Acq: 29 Sep 2021 01:03




Instrument :
 BNA_N
 ClientSampleId :
 DFTPP


Tgt Ion:152 Resp: 29384
 Ion Ratio Lower Upper
 152 100
 150 159.0 125.1 187.7
 115 60.2 45.8 68.8



#2
 1,4-Dioxane
 Concen: 0.867 ng
 RT: 3.281 min Scan# 22
 Delta R.T. 0.009 min
 Lab File: BN016570.D
 Acq: 29 Sep 2021 01:03

Tgt Ion: 88 Resp: 10972
 Ion Ratio Lower Upper
 88 100
 43 0.0 75.8 113.8#
 58 56.0 81.6 122.4#









#3
n-Nitrosodimethylamine
Concen: 0.467 ng
RT: 3.594 min Scan# 56
Delta R.T. -0.019 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion: 42 Resp: 10905
Ion Ratio Lower Upper
42 100
74 151.9 77.5 116.3#
44 13.5 0.0 0.0#



#4
2-Fluorophenol
Concen: 0.001 ng
RT: 5.309 min Scan# 242
Delta R.T. 0.027 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion: 112 Resp: 82
Ion Ratio Lower Upper
112 100
64 54.9 48.4 72.6
63 25.6 25.4 38.2









#5
Phenol-d6
Concen: 0.002 ng
RT: 6.849 min Scan# 409
Delta R.T. 0.009 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion: 99 Resp: 150
Ion Ratio Lower Upper
99 100
42 0.0 19.6 29.4#
71 51.3 24.2 36.2#



#6
bis(2-Chloroethyl)ether
Concen: 0.381 ng
RT: 7.098 min Scan# 436
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion: 93 Resp: 29517
Ion Ratio Lower Upper
93 100
63 75.2 70.0 105.0
95 34.1 26.3 39.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.419 min Scan# 734
Delta R.T. 0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion:	136	Resp:	154267
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.8	13.2
54	6.9	5.8	8.6
68	5.7	3.6	5.4#



#8
Nitrobenzene-d5
Concen: 0.057 ng
RT: 8.708 min Scan# 599
Delta R.T. -0.089 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:	82	Resp:	4700
Ion Ratio	Lower	Upper	
82	100		
128	5.2	27.4	41.2#
54	6.2	58.5	87.7#









#9
Naphthalene
Concen: 0.395 ng
RT: 10.470 min Scan# 738
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion:128 Resp: 164299
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 12.8 10.1 15.1



#10
Hexachlorobutadiene
Concen: 0.375 ng
RT: 10.761 min Scan# 761
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:225 Resp: 30805
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 50.4 75.6









#11
2-Methylnaphthalene-d10
Concen: 0.023 ng
RT: 11.985 min Scan# 927
Delta R.T. -0.036 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion:152 Resp: 5113
Ion Ratio Lower Upper
152 100
151 1.0 16.6 25.0#



#12
2-Methylnaphthalene
Concen: 0.406 ng
RT: 12.091 min Scan# 951
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:142 Resp: 105060
Ion Ratio Lower Upper
142 100
141 88.0 69.9 104.9
115 33.1 27.0 40.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.281 min Scan# 1184
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion:164 Resp: 78516
Ion Ratio Lower Upper
164 100
162 101.0 79.4 119.0
160 46.7 46.6 70.0



#14
2,4,6-Tribromophenol
Concen: 0.001 ng
RT: 15.778 min Scan# 1302
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:330 Resp: 39
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 169.2 34.4 51.6#









#15
2-Fluorobiphenyl
Concen: 0.000 ng
RT: 12.911 min Scan# 1076
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampled :
DFTPP


Tgt Ion:	172	Resp:	144
Ion Ratio	Lower	Upper	
172	100		
171	63.8	27.4	41.2#
170	54.6	18.9	28.3#



#16
Acenaphthylene
Concen: 0.365 ng
RT: 14.002 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:	152	Resp:	127205
Ion Ratio	Lower	Upper	
152	100		
151	19.4	15.4	23.0
153	13.0	10.4	15.6









#17
Acenaphthene
Concen: 0.375 ng
RT: 14.345 min Scan# 1189
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion:154 Resp: 88676
Ion Ratio Lower Upper
154 100
153 113.0 91.0 136.4
152 56.7 55.8 83.6



#18
Fluorene
Concen: 0.379 ng
RT: 15.334 min Scan# 1267
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:166 Resp: 109344
Ion Ratio Lower Upper
166 100
165 97.3 77.6 116.4
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.042 min Scan# 1405
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion:188 Resp: 146577
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.1 6.9 10.3#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.340 ng
RT: 15.423 min Scan# 1274
Delta R.T. -0.013 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:198 Resp: 1514
Ion Ratio Lower Upper
198 100
182 21.8 17.5 26.3
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.388 ng
RT: 16.238 min Scan# 1339
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampled :
DFTPP


Tgt Ion:248 Resp: 33625
Ion Ratio Lower Upper
248 100
250 98.7 71.8 107.6
141 46.9 69.7 104.5#



#22
Hexachlorobenzene
Concen: 0.397 ng
RT: 16.336 min Scan# 1347
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:284 Resp: 41047
Ion Ratio Lower Upper
284 100
142 34.1 30.7 46.1
249 29.4 25.8 38.8









#23
Atrazine
Concen: 0.370 ng
RT: 16.518 min Scan# 1362
Delta R.T. -0.001 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion:200 Resp: 20079
Ion Ratio Lower Upper
200 100
173 23.5 22.6 34.0
215 49.3 38.2 57.4



#24
Pentachlorophenol
Concen: 0.559 ng
RT: 16.689 min Scan# 1376
Delta R.T. -0.012 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:266 Resp: 17992
Ion Ratio Lower Upper
266 100
264 62.8 50.2 75.2
268 64.1 51.8 77.8









#25
Phenanthrene
Concen: 0.395 ng
RT: 17.078 min Scan# 1408
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion:178 Resp: 174038
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.6
179 15.3 12.3 18.5



#26
Anthracene
Concen: 0.383 ng
RT: 17.163 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:178 Resp: 146897
Ion Ratio Lower Upper
178 100
176 18.4 14.6 21.8
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.000 ng
RT: 19.082 min Scan# 1700
Delta R.T. 0.004 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion:212 Resp: 21
Ion Ratio Lower Upper
212 100
106 28.6 8.8 13.2#
104 42.9 5.0 7.4#



#28
Fluoranthene
Concen: 0.379 ng
RT: 19.107 min Scan# 1705
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:202 Resp: 182531
Ion Ratio Lower Upper
202 100
101 10.9 7.5 11.3
203 17.1 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.248 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion:240 Resp: 134772
Ion Ratio Lower Upper
240 100
120 6.1 9.0 13.4#
236 25.4 21.7 32.5



#30
Pyrene
Concen: 0.404 ng
RT: 19.475 min Scan# 1779
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:202 Resp: 185601
Ion Ratio Lower Upper
202 100
200 20.6 16.3 24.5
203 17.6 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.000 ng
 RT: 19.693 min Scan# 1823
 Delta R.T. -0.000 min
 Lab File: BN016570.D
 Acq: 29 Sep 2021 01:03




Instrument :
 BNA_N
 ClientSampleId :
 DFTPP


Tgt Ion:244	Resp:	37
Ion Ratio	Lower	Upper
244	100	
212	76.4	6.1 9.1#
122	123.6	8.4 12.6#



#32
 Benzo(a)anthracene
 Concen: 0.391 ng
 RT: 21.227 min Scan# 2001
 Delta R.T. -0.000 min
 Lab File: BN016570.D
 Acq: 29 Sep 2021 01:03

Tgt Ion:228	Resp:	163153
Ion Ratio	Lower	Upper
228	100	
226	27.0	21.2 31.8
229	19.3	15.6 23.4









#33
Chrysene
Concen: 0.409 ng
RT: 21.280 min Scan# 2006
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion:228	Resp: 174251
Ion Ratio	Lower Upper
228 100	
226 29.8	23.1 34.7
229 19.4	15.8 23.8



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.380 ng
RT: 21.174 min Scan# 1996
Delta R.T. -0.011 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03


Tgt Ion:149	Resp: 47234
Ion Ratio	Lower Upper
149 100	
167 27.0	21.4 32.2
279 3.6	3.1 4.7








#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.505 min Scan# 2433
 Delta R.T. -0.004 min
 Lab File: BN016570.D
 Acq: 29 Sep 2021 01:03

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP



Tgt Ion:264 Resp: 119008




Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.9	29.9
265	23.4	20.7	31.1




#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.372 ng
 RT: 25.795 min Scan# 3102
 Delta R.T. -0.000 min
 Lab File: BN016570.D
 Acq: 29 Sep 2021 01:03

Tgt Ion:276 Resp: 141031

Ion	Ratio	Lower	Upper
276	100		
138	24.7	18.2	27.2
277	24.8	19.8	29.8









#37
Benzo(b)fluoranthene
Concen: 0.409 ng
RT: 22.824 min Scan# 2234
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion:252 Resp: 162570
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 10.0 7.3 10.9



#38
Benzo(k)fluoranthene
Concen: 0.406 ng
RT: 22.868 min Scan# 2247
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:252 Resp: 157341
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 10.6 7.4 11.0









#39
Benzo(a)pyrene
Concen: 0.397 ng
RT: 23.405 min Scan# 2404
Delta R.T. -0.000 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03




Instrument :
BNA_N
ClientSampleId :
DFTPP


Tgt Ion:252 Resp: 144575
Ion Ratio Lower Upper
252 100
253 22.2 17.9 26.9
125 11.2 8.2 12.4



#40
Dibenzo(a,h)anthracene
Concen: 0.411 ng
RT: 25.815 min Scan# 3108
Delta R.T. -0.004 min
Lab File: BN016570.D
Acq: 29 Sep 2021 01:03

Tgt Ion:278 Resp: 108256
Ion Ratio Lower Upper
278 100
139 17.3 12.1 18.1
279 23.9 19.2 28.8





#41
 Benzo(g,h,i)perylene
 Concen: 0.388 ng
 RT: 26.489 min Scan# 3305
 Delta R.T. -0.004 min
 Lab File: BN016570.D
 Acq: 29 Sep 2021 01:03

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Tgt Ion: 276 Resp: 133914

Ion	Ratio	Lower	Upper
276	100		
277	23.8	18.9	28.3
138	21.7	15.7	23.5

