

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092821\
 Data File : BN016569.D
 Acq On : 28 Sep 2021 23:51
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 29 01:56:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN092821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 29 01:39:43 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	114	0.00
2	1,4-Dioxane	0.400	0.335	16.3	96	0.00
3	n-Nitrosodimethylamine	0.400	0.355	11.3	109	0.00
4 S	2-Fluorophenol	0.400	0.365	8.8	110	0.00
5 S	Phenol-d6	0.400	0.355	11.3	111	0.00
6	bis(2-Chloroethyl)ether	0.400	0.390	2.5	115	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	113	0.00
8 S	Nitrobenzene-d5	0.400	0.323	19.3	119	0.00
9	Naphthalene	0.400	0.383	4.3	113	0.00
10	Hexachlorobutadiene	0.400	0.379	5.3	111	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.388	3.0	116	0.00
12	2-Methylnaphthalene	0.400	0.391	2.3	115	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	117	0.00
14 S	2,4,6-Tribromophenol	0.400	0.304	24.0	108	0.00
15 S	2-Fluorobiphenyl	0.400	0.378	5.5	115	0.01
16	Acenaphthylene	0.400	0.342	14.5	111	0.00
17	Acenaphthene	0.400	0.377	5.8	115	0.00
18	Fluorene	0.400	0.372	7.0	111	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	115	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.348	13.0	127	0.01
21	4-Bromophenyl-phenylether	0.400	0.380	5.0	116	0.00
22	Hexachlorobenzene	0.400	0.387	3.3	114	0.01
23	Atrazine	0.400	0.356	11.0	115	0.00
24	Pentachlorophenol	0.400	0.371	7.3	122	0.01
25	Phenanthrene	0.400	0.384	4.0	114	0.00
26	Anthracene	0.400	0.356	11.0	112	0.00
27 SURR	Fluoranthene-d10	0.400	0.384	4.0	117	0.00
28	Fluoranthene	0.400	0.392	2.0	116	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	118	0.00
30	Pyrene	0.400	0.407	-1.7	115	0.00
31 S	Terphenyl-d14	0.400	0.406	-1.5	118	0.00
32	Benzo(a)anthracene	0.400	0.376	6.0	111	0.00
33	Chrysene	0.400	0.413	-3.2	119	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.356	11.0	107	0.00
35 I	Perylene-d12	0.400	0.400	0.0	130	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.390	2.5	134	0.00
37	Benzo(b)fluoranthene	0.400	0.378	5.5	117	0.00
38	Benzo(k)fluoranthene	0.400	0.395	1.3	125	0.00
39 C	Benzo(a)pyrene	0.400	0.370	7.5	120	0.00
40	Dibenzo(a,h)anthracene	0.400	0.435	-8.7	138	0.00
41	Benzo(g,h,i)perylene	0.400	0.386	3.5	126	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4