

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
 Data File : BN016600.D
 Acq On : 30 Sep 2021 14:55
 Operator : CG/JU
 Sample : SSTDICC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4

Quant Time: Sep 30 15:27:03 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 30 15:26:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.642	152	30679	0.400	ng	0.00
7) Naphthalene-d8	10.406	136	137151	0.400	ng	0.00
13) Acenaphthene-d10	14.268	164	70325	0.400	ng	0.00
19) Phenanthrene-d10	17.029	188	125659	0.400	ng	0.00
29) Chrysene-d12	21.238	240	112614	0.400	ng	0.00
35) Perylene-d12	23.488	264	105768	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.254	112	29133	0.370	ng	0.00
5) Phenol-d6	6.821	99	31261	0.371	ng	0.00
8) Nitrobenzene-d5	8.784	82	29307	0.340	ng	0.00
11) 2-Methylnaphthalene-d10	12.007	152	75242	0.375	ng	0.00
14) 2,4,6-Tribromophenol	15.766	330	10334	0.322	ng	0.00
15) 2-Fluorobiphenyl	12.898	172	104790	0.363	ng	0.00
27) Fluoranthene-d10	19.068	212	120314	0.349	ng	0.00
31) Terphenyl-d14	19.678	244	98108	0.401	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.253	88	5235	0.345	ng	100
3) n-Nitrosodimethylamine	3.576	42	8381	0.363	ng	99
6) bis(2-Chloroethyl)ether	7.080	93	31548	0.378	ng	100
9) Naphthalene	10.457	128	139454	0.368	ng	100
10) Hexachlorobutadiene	10.749	225	25908	0.368	ng	# 100
12) 2-Methylnaphthalene	12.078	142	89478	0.379	ng	100
16) Acenaphthylene	13.989	152	110133	0.353	ng	100
17) Acenaphthene	14.332	154	77014	0.372	ng	100
18) Fluorene	15.322	166	89872	0.365	ng	100
20) 4,6-Dinitro-2-methylph...	15.410	198	3586	0.251	ng	100
21) 4-Bromophenyl-phenylether	16.226	248	27940	0.360	ng	100
22) Hexachlorobenzene	16.336	284	33596	0.371	ng	100
23) Atrazine	16.506	200	17739	0.312	ng	100
24) Pentachlorophenol	16.677	266	9921	0.324	ng	100
25) Phenanthrene	17.066	178	139659	0.368	ng	100
26) Anthracene	17.151	178	115061	0.352	ng	100
28) Fluoranthene	19.097	202	150682	0.361	ng	100
30) Pyrene	19.460	202	148122	0.404	ng	100
32) Benzo(a)anthracene	21.217	228	133386	0.389	ng	100
33) Chrysene	21.270	228	142044	0.395	ng	100
34) Bis(2-ethylhexyl)phtha...	21.174	149	50171	0.396	ng	100
36) Indeno(1,2,3-cd)pyrene	25.764	276	146658	0.372	ng	100
37) Benzo(b)fluoranthene	22.810	252	137984	0.422	ng	100
38) Benzo(k)fluoranthene	22.854	252	136068	0.411	ng	100
39) Benzo(a)pyrene	23.388	252	121790	0.398	ng	100
40) Dibenzo(a,h)anthracene	25.788	278	117860	0.370	ng	100
41) Benzo(g,h,i)perylene	26.459	276	133821	0.386	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
Data File : BN016600.D
Acq On : 30 Sep 2021 14:55
Operator : CG/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Sep 30 15:27:03 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 30 15:26:45 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.642 min Scan# 495
 Delta R.T. 0.000 min
 Lab File: BN016600.D
 Acq: 30 Sep 2021 14:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:	152	Resp:	30679
Ion	Ratio	Lower	Upper
152	100		
150	155.7	124.6	186.8
115	58.1	46.5	69.7



1,4-Dioxane
 Concen: 0.345 ng
 RT: 3.253 min Scan# 19
 Delta R.T. 0.000 min
 Lab File: BN016600.D
 Acq: 30 Sep 2021 14:55

Tgt Ion:	88	Resp:	5235
Ion	Ratio	Lower	Upper
88	100		
43	47.3	37.8	56.8
58	76.6	61.3	91.9









Scan 54 (3.576 min): BN016600.D\data.ms (-51) (#3)
 n-Nitrosodimethylamine
 Concen: 0.363 ng
 RT: 3.576 min Scan# 54
 Delta R.T. 0.000 min
 Lab File: BN016600.D
 Acq: 30 Sep 2021 14:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:	42	Resp:	8381
Ion	Ratio	Lower	Upper
42	100		
74	150.9	121.6	182.4
44	8.0	6.3	9.5



Scan 236 (5.254 min): BN016600.D\data.ms (-228) (#4)
 2-Fluorophenol
 Concen: 0.370 ng
 RT: 5.254 min Scan# 236
 Delta R.T. 0.000 min
 Lab File: BN016600.D
 Acq: 30 Sep 2021 14:55

Tgt Ion:	112	Resp:	29133
Ion	Ratio	Lower	Upper
112	100		
64	53.9	43.1	64.7
63	27.0	21.6	32.4









Phenol-d6
 Concen: 0.371 ng
 RT: 6.821 min Scan# 406
 Delta R.T. 0.000 min
 Lab File: BN016600.D
 Acq: 30 Sep 2021 14:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4

Tgt Ion:	99	Resp:	31261
Ion	Ratio	Lower	Upper
99	100		
42	16.3	13.0	19.6
71	29.4	23.5	35.3



bis(2-Chloroethyl)ether
 Concen: 0.378 ng
 RT: 7.080 min Scan# 434
 Delta R.T. 0.000 min
 Lab File: BN016600.D
 Acq: 30 Sep 2021 14:55




Tgt Ion:	93	Resp:	31548
Ion	Ratio	Lower	Upper
93	100		
63	73.9	59.1	88.7
95	32.8	26.2	39.4




Abundance Scan 733 (10.406 min): BN016600.D\data.ms (-7267-)

Naphthalene-d8
Concen: 0.400 ng
RT: 10.406 min Scan# 733
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4






Tgt Ion:	136	Resp:	137151
Ion Ratio	Lower	Upper	
136	100		
137	10.9	8.7	13.1
54	7.4	5.9	8.9
68	6.3	5.0	7.6





Abundance Scan 605 (8.784 min): BN016600.D\data.ms (-60178)

Nitrobenzene-d5
Concen: 0.340 ng
RT: 8.784 min Scan# 605
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55



Tgt Ion:	82	Resp:	29307
Ion Ratio	Lower	Upper	
82	100		
128	42.0	33.6	50.4
54	47.0	37.6	56.4









Naphthalene
 Concen: 0.368 ng
 RT: 10.457 min Scan# 737
 Delta R.T. 0.000 min
 Lab File: BN016600.D
 Acq: 30 Sep 2021 14:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4

Tgt Ion:	128	Resp:	139454
Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.8	13.2
127	12.9	10.3	15.5




Hexachlorobutadiene
 Concen: 0.368 ng
 RT: 10.749 min Scan# 760
 Delta R.T. 0.000 min
 Lab File: BN016600.D
 Acq: 30 Sep 2021 14:55

Tgt Ion:	225	Resp:	25908
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.8	51.0	76.6






Abundance Scan 932 (12.007 min): BN016600.D\data.ms (-917) (-)




2-Methylnaphthalene-d10
Concen: 0.375 ng
RT: 12.007 min Scan# 932
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:152 Resp: 75242
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4






Abundance Scan 948 (12.078 min): BN016600.D\data.ms (-937) (-)



2-Methylnaphthalene
Concen: 0.379 ng
RT: 12.078 min Scan# 948
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55


Tgt Ion:142 Resp: 89478
Ion Ratio Lower Upper
142 100
141 88.2 70.6 105.8
115 33.2 26.6 39.8






Abundance Scan 1183 (14.268 min): BN016600.D\data.ms (-1167) (-)

Acenaphthene-d10
Concen: 0.400 ng
RT: 14.268 min Scan# 1183
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion:	164	Resp:	70325
Ion Ratio	Lower	Upper	
164	100		
162	102.7	82.2	123.2
160	47.1	37.7	56.5






Abundance Scan 1301 (15.766 min): BN016600.D\data.ms (-1293) (-)


2,4,6-Tribromophenol
Concen: 0.322 ng
RT: 15.766 min Scan# 1301
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55



Tgt Ion:	330	Resp:	10334
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	31.5	25.2	37.8






Abundance Scan 1075 (12.898 min): BN016600.D\data.ms (-1075) (-)




2-Fluorobiphenyl
Concen: 0.363 ng
RT: 12.898 min Scan# 1075
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:	172	Resp:	104790
Ion Ratio	Lower	Upper	
172	100		
171	33.7	27.0	40.4
170	21.8	17.4	26.2






Abundance Scan 1161 (13.989 min): BN016600.D\data.ms (-1161) (-)



Acenaphthylene
Concen: 0.353 ng
RT: 13.989 min Scan# 1161
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55


Tgt Ion:	152	Resp:	110133
Ion Ratio	Lower	Upper	
152	100		
151	19.4	15.5	23.3
153	13.0	10.4	15.6






Abundance Scan 1188 (14.332 min): BN016600.D\data.ms (-1187) (-)

Acenaphthene
Concen: 0.372 ng
RT: 14.332 min Scan# 1188
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion:	154	Resp:	77014
Ion	Ratio	Lower	Upper
154	100		
153	112.7	90.2	135.2
152	55.0	44.0	66.0






Abundance Scan 1266 (15.322 min): BN016600.D\data.ms (-1265) (-)

Fluorene
Concen: 0.365 ng
RT: 15.322 min Scan# 1266
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55




Tgt Ion:	166	Resp:	89872
Ion	Ratio	Lower	Upper
166	100		
165	97.4	77.9	116.9
167	13.4	10.7	16.1






Abundance Scan 1404 (17.029 min): BN016600.D\data.ms (-1388) (-)

Phenanthrene-d10
Concen: 0.400 ng
RT: 17.029 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	6.3	5.0	7.6






Abundance Scan 1273 (15.410 min): BN016600.D\data.ms (-1265) (-)


4,6-Dinitro-2-methylphenol
Concen: 0.251 ng
RT: 15.410 min Scan# 1273
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55



Tgt Ion	Ratio	Lower	Upper
198	100		
182	14.8	11.8	17.8
77	0.0	0.0	0.0





Abundance Scan 1338 (16.226 min): BN016600.D\data.ms (-1333) (-)




4-Bromophenyl-phenylether
Concen: 0.360 ng
RT: 16.226 min Scan# 1338
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	248	Resp:	27940
Ion	Ratio	Lower	Upper
248	100		
250	98.2	78.6	117.8
141	48.2	38.6	57.8



Abundance Scan 1338 (16.226 min): BN016600.D\data.ms (-1330) (-)





Abundance Scan 1347 (16.336 min): BN016600.D\data.ms (-1342) (-)





Hexachlorobenzene
Concen: 0.371 ng
RT: 16.336 min Scan# 1347
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Tgt Ion:	284	Resp:	33596
Ion	Ratio	Lower	Upper
284	100		
142	33.0	26.4	39.6
249	29.3	23.4	35.2



Abundance Scan 1347 (16.336 min): BN016600.D\data.ms (-1339) (-)









#23 (-)
Atrazine
Concen: 0.312 ng
RT: 16.506 min Scan# 1361
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:	200	Resp:	17739
Ion Ratio	Lower	Upper	
200	100		
173	23.2	18.6	27.8
215	49.3	39.4	59.2




#24 (-)
Pentachlorophenol
Concen: 0.324 ng
RT: 16.677 min Scan# 1375
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Tgt Ion:	266	Resp:	9921
Ion Ratio	Lower	Upper	
266	100		
264	62.5	49.8	74.6
268	64.4	51.4	77.2






Abundance Scan 1407 (17.066 min): BN016600.D\data.ms (-1382) (-)




Phenanthrene
Concen: 0.368 ng
RT: 17.066 min Scan# 1407
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:	178	Resp:	139659
Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.4
179	15.3	12.2	18.4






Abundance Scan 1414 (17.151 min): BN016600.D\data.ms (-1406) (-)




Anthracene
Concen: 0.352 ng
RT: 17.151 min Scan# 1414
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Tgt Ion:	178	Resp:	115061
Ion	Ratio	Lower	Upper
178	100		
176	18.1	14.6	21.8
179	15.3	12.2	18.4





Abundance Scan 1697 (19.068 min): BN016600.D\data.ms (-1627) (-)




Fluoranthene-d10
Concen: 0.349 ng
RT: 19.068 min Scan# 1697
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	212	Resp:	120314
Ion Ratio	Lower	Upper	
212	100		
106	13.4	10.6	16.0
104	7.4	5.9	8.9



Abundance Scan 1697 (19.068 min): BN016600.D\data.ms (-1677) (-)





Abundance Scan 1703 (19.097 min): BN016600.D\data.ms (-1628) (-)




Fluoranthene
Concen: 0.361 ng
RT: 19.097 min Scan# 1703
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Tgt Ion:	202	Resp:	150682
Ion Ratio	Lower	Upper	
202	100		
101	11.2	9.0	13.4
203	17.2	13.8	20.6




Abundance Scan 1703 (19.097 min): BN016600.D\data.ms (-1663) (-)






Abundance Scan 2002 (21.238 min): BN016600.D\data.ms (-1983) (-)

Chrysene-d12
Concen: 0.400 ng
RT: 21.238 min Scan# 2002
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion	Ratio	Lower	Upper
240	100		
120	6.2	5.0	7.4
236	25.5	20.4	30.6






Abundance Scan 1776 (19.460 min): BN016600.D\data.ms (-1736) (-)


Pyrene
Concen: 0.404 ng
RT: 19.460 min Scan# 1776
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55



Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.5	24.7
203	17.5	14.0	21.0






Abundance Scan 1820 (19.678 min): BN016600.D\data.ms (-1790) (-)




Terphenyl-d14
Concen: 0.401 ng
RT: 19.678 min Scan# 1820
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.5	6.0	9.0
122	12.6	10.1	15.1







Abundance Scan 2000 (21.217 min): BN016600.D\data.ms (-1981) (-)



Benzo(a)anthracene
Concen: 0.389 ng
RT: 21.217 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.4	21.9	32.9
229	19.6	15.7	23.5









Chrysene
Concen: 0.395 ng
RT: 21.270 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:	228	Resp:	142044
Ion	Ratio	Lower	Upper
228	100		
226	29.6	23.7	35.5
229	19.4	15.5	23.3



Bis(2-ethylhexyl)phthalate
Concen: 0.396 ng
RT: 21.174 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55


Tgt Ion:	149	Resp:	50171
Ion	Ratio	Lower	Upper
149	100		
167	27.8	22.2	33.4
279	4.0	3.2	4.8






Abundance Scan 2428 (23.488 min): BN016600.D\data.ms (-2428) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.488 min Scan# 2428
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.9	19.9	29.9
265	24.0	19.2	28.8






Abundance Scan 3093 (25.764 min): BN016600.D\data.ms (-3093) (-)

Indeno(1,2,3-cd)pyrene
Concen: 0.372 ng
RT: 25.764 min Scan# 3093
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55




Tgt Ion	Ratio	Lower	Upper
276	100		
138	24.5	19.5	29.3
277	24.9	19.9	29.9





Abundance Scan 2230 (22.810 min): BN016600.D\data.ms (-2137) (-)

Benzo(b)fluoranthene
Concen: 0.422 ng
RT: 22.810 min Scan# 2230
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55


Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.8	26.8
125	10.0	8.0	12.0






Abundance Scan 2230 (22.810 min): BN016600.D\data.ms (-2172) (-)




Abundance Scan 2243 (22.854 min): BN016600.D\data.ms (-2238) (-)

Benzo(k)fluoranthene
Concen: 0.411 ng
RT: 22.854 min Scan# 2243
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.8	26.6
125	10.0	8.0	12.0




Abundance Scan 2243 (22.854 min): BN016600.D\data.ms (-2228) (-)





Abundance Scan 2399 (23.388 min): BN016600.D\data.ms (-2341) (-)

Benzo(a)pyrene
Concen: 0.398 ng
RT: 23.388 min Scan# 2399
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55


Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion:	252	Resp:	121790
Ion Ratio	Lower	Upper	
252	100		
253	22.3	17.8	26.8
125	11.2	9.0	13.4






Abundance Scan 2399 (23.388 min): BN016600.D\data.ms (-2341) (-)




Abundance Scan 3100 (25.788 min): BN016600.D\data.ms (-3071) (-)

Dibenzo(a,h)anthracene
Concen: 0.370 ng
RT: 25.788 min Scan# 3100
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55


Tgt Ion:	278	Resp:	117860
Ion Ratio	Lower	Upper	
278	100		
139	16.4	13.1	19.7
279	23.9	19.1	28.7



Abundance Scan 3100 (25.788 min): BN016600.D\data.ms (-3071) (-)




Abundance Scan 3296 (26.459 min): BN016600.D\data.ms (-3270) (-)



Benzo(g,h,i)perylene
Concen: 0.386 ng
RT: 26.459 min Scan# 3296
Delta R.T. 0.000 min
Lab File: BN016600.D
Acq: 30 Sep 2021 14:55

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion:	276	Resp:	133821
Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.0	28.6
138	21.1	16.9	25.3

