

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN100121\  
 Data File : BN016640.D  
 Acq On : 01 Oct 2021 17:45  
 Operator : CG/JU  
 Sample : M3833-14  
 Misc :  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 RE122D3-20210920

Quant Time: Oct 02 00:52:32 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN100121.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Oct 01 13:35:26 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc | Units | Dev(Min)      |
|-------------------------------|--------|------|----------|------|-------|---------------|
| -----                         |        |      |          |      |       |               |
| Internal Standards            |        |      |          |      |       |               |
| 1) 1,4-Dichlorobenzene-d4     | 7.643  | 152  | 28380    | 0.40 | ng    | 0.00          |
| 7) Naphthalene-d8             | 10.407 | 136  | 132552   | 0.40 | ng    | 0.00          |
| 13) Acenaphthene-d10          | 14.269 | 164  | 73104    | 0.40 | ng    | 0.00          |
| 19) Phenanthrene-d10          | 17.018 | 188  | 138583   | 0.40 | ng    | #-0.01        |
| 29) Chrysene-d12              | 21.228 | 240  | 131100   | 0.40 | ng    | #-0.01        |
| 35) Perylene-d12              | 23.481 | 264  | 125874   | 0.40 | ng    | 0.00          |
| System Monitoring Compounds   |        |      |          |      |       |               |
| 4) 2-Fluorophenol             | 5.273  | 112  | 10690    | 0.15 | ng    | 0.02          |
| 5) Phenol-d6                  | 6.831  | 99   | 6719     | 0.08 | ng    | 0.00          |
| 8) Nitrobenzene-d5            | 8.785  | 82   | 28156    | 0.33 | ng    | 0.00          |
| 11) 2-Methylnaphthalene-d10   | 12.003 | 152  | 66684    | 0.34 | ng    | 0.00          |
| 14) 2,4,6-Tribromophenol      | 15.766 | 330  | 12257    | 0.37 | ng    | 0.00          |
| 15) 2-Fluorobiphenyl          | 12.899 | 172  | 90256    | 0.30 | ng    | 0.00          |
| 27) Fluoranthene-d10          | 19.068 | 212  | 150499   | 0.41 | ng    | 0.00          |
| 31) Terphenyl-d14             | 19.679 | 244  | 126372   | 0.44 | ng    | 0.00          |
| Target Compounds              |        |      |          |      |       |               |
| 34) Bis(2-ethylhexyl)phtha... | 21.164 | 149  | 33888    | 0.21 | ng    | Qvalue<br>100 |
| -----                         |        |      |          |      |       |               |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN100121\  
Data File : BN016640.D  
Acq On : 01 Oct 2021 17:45  
Operator : CG/JU  
Sample : M3833-14  
Misc :  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
RE122D3-20210920

Quant Time: Oct 02 00:52:32 2021  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN100121.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Oct 01 13:35:26 2021  
Response via : Initial Calibration

