

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
 Data File : BN016642.D
 Acq On : 01 Oct 2021 18:57
 Operator : CG/JU
 Sample : M3833-03
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW202D1-20210919

Quant Time: Oct 02 00:52:48 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 13:35:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.643	152	31727	0.40	ng	0.00
7) Naphthalene-d8	10.407	136	143581	0.40	ng	0.00
13) Acenaphthene-d10	14.269	164	76282	0.40	ng	0.00
19) Phenanthrene-d10	17.030	188	147966	0.40	ng	0.00
29) Chrysene-d12	21.228	240	138792	0.40	ng	#-0.01
35) Perylene-d12	23.485	264	129788	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.273	112	11142	0.14	ng	0.02
5) Phenol-d6	6.831	99	7326	0.08	ng	0.00
8) Nitrobenzene-d5	8.785	82	29313	0.31	ng	0.00
11) 2-Methylnaphthalene-d10	12.003	152	70630	0.33	ng	0.00
14) 2,4,6-Tribromophenol	15.766	330	14358	0.41	ng	0.00
15) 2-Fluorobiphenyl	12.899	172	99793	0.32	ng	0.00
27) Fluoranthene-d10	19.063	212	151865	0.39	ng	0.00
31) Terphenyl-d14	19.679	244	129935	0.43	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.164	149	49504	0.29	ng	Qvalue # 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
Data File : BN016642.D
Acq On : 01 Oct 2021 18:57
Operator : CG/JU
Sample : M3833-03
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW202D1-20210919

Quant Time: Oct 02 00:52:48 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 01 13:35:26 2021
Response via : Initial Calibration

