

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
 Data File : BN016644.D
 Acq On : 01 Oct 2021 20:10
 Operator : CG/JU
 Sample : M3833-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW201D1-20210919

Manual Integrations
 APPROVED

mohammad
 10/4/2021 11:48:05 AM

Quant Time: Oct 02 00:53:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 13:35:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.642	152	29338	0.400	ng	0.00
7) Naphthalene-d8	10.406	136	134572	0.400	ng	0.00
13) Acenaphthene-d10	14.268	164	73896	0.400	ng	0.00
19) Phenanthrene-d10	17.017	188	142567	0.400	ng	#-0.01
29) Chrysene-d12	21.227	240	137745	0.400	ng	#-0.01
35) Perylene-d12	23.481	264	144287	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.272	112	10499	0.141	ng	0.02
5) Phenol-d6	6.831	99	6815	0.083	ng	0.00
8) Nitrobenzene-d5	8.784	82	26675	0.304	ng	0.00
11) 2-Methylnaphthalene-d10	12.002	152	64452	0.324	ng	0.00
14) 2,4,6-Tribromophenol	15.765	330	11919	0.354	ng	0.00
15) 2-Fluorobiphenyl	12.898	172	89867	0.300	ng	0.00
27) Fluoranthene-d10	19.068	212	161946	0.427	ng	0.00
31) Terphenyl-d14	19.678	244	134247	0.448	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.163	149	46222	0.274	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
Data File : BN016644.D
Acq On : 01 Oct 2021 20:10
Operator : CG/JU
Sample : M3833-05
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW201D1-20210919

Manual Integrations
APPROVED

mohammad
10/4/2021 11:48:05 AM

Quant Time: Oct 02 00:53:02 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 01 13:35:26 2021
Response via : Initial Calibration

