

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
 Data File : BN016646.D
 Acq On : 01 Oct 2021 21:23
 Operator : CG/JU
 Sample : M3833-11
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP01-20210920

Manual Integrations
 APPROVED

mohammad
 10/4/2021 11:48:07 AM

Quant Time: Oct 02 00:53:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 13:35:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.642	152	29088	0.40	ng	0.00
7) Naphthalene-d8	10.407	136	133173	0.40	ng	0.00
13) Acenaphthene-d10	14.269	164	71109	0.40	ng	0.00
19) Phenanthrene-d10	17.018	188	134609	0.40	ng	#-0.01
29) Chrysene-d12	21.238	240	122781	0.40	ng	0.00
35) Perylene-d12	23.484	264	127737	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.273	112	9941	0.13	ng	0.02
5) Phenol-d6	6.831	99	6064	0.07	ng	0.00
8) Nitrobenzene-d5	8.784	82	23410	0.27	ng	0.00
11) 2-Methylnaphthalene-d10	12.003	152	59615	0.30	ng	0.00
14) 2,4,6-Tribromophenol	15.766	330	9554	0.30	ng	0.00
15) 2-Fluorobiphenyl	12.898	172	81420	0.28	ng	0.00
27) Fluoranthene-d10	19.068	212	130559	0.36	ng	0.00
31) Terphenyl-d14	19.678	244	108527	0.41	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.263	88	6705m	0.50	ng	
34) Bis(2-ethylhexyl)phtha...	21.164	149	31239	0.21	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
Data File : BN016646.D
Acq On : 01 Oct 2021 21:23
Operator : CG/JU
Sample : M3833-11
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP01-20210920

Manual Integrations
APPROVED

mohammad
10/4/2021 11:48:07 AM

Quant Time: Oct 02 00:53:17 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 01 13:35:26 2021
Response via : Initial Calibration

