

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
 Data File : BN016663.D
 Acq On : 02 Oct 2021 09:21
 Operator : CG/JU
 Sample : M3829-14
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE120D3-20210917

Manual Integrations
 APPROVED

mohammad
 10/4/2021 11:48:26 AM

Quant Time: Oct 04 03:33:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 13:35:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.642	152	28933	0.40	ng	0.00
7) Naphthalene-d8	10.406	136	132898	0.40	ng	# 0.00
13) Acenaphthene-d10	14.268	164	70382	0.40	ng	0.00
19) Phenanthrene-d10	17.017	188	137081	0.40	ng	#-0.01
29) Chrysene-d12	21.227	240	133629	0.40	ng	#-0.01
35) Perylene-d12	23.481	264	136983	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.272	112	10000	0.14	ng	0.02
5) Phenol-d6	6.821	99	6252	0.08	ng	0.00
8) Nitrobenzene-d5	8.784	82	23361	0.27	ng	0.00
11) 2-Methylnaphthalene-d10	12.002	152	65579	0.33	ng	0.00
14) 2,4,6-Tribromophenol	15.766	330	10642	0.33	ng	0.00
15) 2-Fluorobiphenyl	12.885	172	90880	0.32	ng	-0.01
27) Fluoranthene-d10	19.063	212	158169	0.43	ng	0.00
31) Terphenyl-d14	19.678	244	133577	0.46	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.253	88	1690m	0.13	ng	Qvalue
34) Bis(2-ethylhexyl)phtha...	21.163	149	162949	1.00	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
Data File : BN016663.D
Acq On : 02 Oct 2021 09:21
Operator : CG/JU
Sample : M3829-14
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE120D3-20210917

Manual Integrations
APPROVED

mohammad
10/4/2021 11:48:26 AM

Quant Time: Oct 04 03:33:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 01 13:35:26 2021
Response via : Initial Calibration

