

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
 Data File : BN016678.D
 Acq On : 02 Oct 2021 19:33
 Operator : CG/JU
 Sample : M3829-21
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW202D-20210918

Quant Time: Oct 04 03:35:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 13:35:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.642	152	31036	0.40	ng	0.00
7) Naphthalene-d8	10.407	136	141607	0.40	ng	# 0.00
13) Acenaphthene-d10	14.269	164	74848	0.40	ng	0.00
19) Phenanthrene-d10	17.018	188	146942	0.40	ng	#-0.01
29) Chrysene-d12	21.227	240	138193	0.40	ng	#-0.01
35) Perylene-d12	23.484	264	142624	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.273	112	10243	0.13	ng	0.02
5) Phenol-d6	6.831	99	6318	0.07	ng	0.00
8) Nitrobenzene-d5	8.784	82	21231	0.23	ng	0.00
11) 2-Methylnaphthalene-d10	12.003	152	59229	0.28	ng	0.00
14) 2,4,6-Tribromophenol	15.766	330	10767	0.32	ng	0.00
15) 2-Fluorobiphenyl	12.886	172	81431	0.27	ng	-0.01
27) Fluoranthene-d10	19.063	212	157979	0.40	ng	0.00
31) Terphenyl-d14	19.678	244	132290	0.44	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.253	88	4997	0.35	ng	# 83
34) Bis(2-ethylhexyl)phtha...	21.164	149	36202	0.21	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
Data File : BN016678.D
Acq On : 02 Oct 2021 19:33
Operator : CG/JU
Sample : M3829-21
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW202D-20210918

Quant Time: Oct 04 03:35:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 01 13:35:26 2021
Response via : Initial Calibration

