




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
Data File : BN016687.D
Acq On : 03 Oct 2021 01:34
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 55 Sample Multiplier: 1

Quant Time: Oct 04 05:48:38 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 30 18:14:09 2021
Response via : Initial Calibration



TIC: BN016687.D\data.ms

(70) Pentachlorophenol (C)




16.675min (+ 0.000) 29077.44 ng m

response 339094

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.14
264.00	61.60	62.86
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
Data File : BN016687.D
Acq On : 03 Oct 2021 01:34
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 55 Sample Multiplier: 1

Quant Time: Oct 04 05:48:38 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 30 18:14:09 2021
Response via : Initial Calibration



TIC: BN016687.D\data.ms

(77) Benzidine

19.292min (+ 0.000) 0.00 ng

response 1803928

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.00
183.00	13.20	11.45
0.00	0.00	0.00