

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
 Data File : BN016605.D
 Acq On : 30 Sep 2021 18:33
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN100121

Quant Time: Sep 30 19:05:19 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 30 17:56:53 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	94	0.00
2	1,4-Dioxane	0.400	0.384	4.0	98	0.00
3	n-Nitrosodimethylamine	0.400	0.362	9.5	92	0.00
4 S	2-Fluorophenol	0.400	0.368	8.0	93	0.00
5 S	Phenol-d6	0.400	0.357	10.8	92	0.00
6	bis(2-Chloroethyl)ether	0.400	0.379	5.3	94	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	93	0.00
8 S	Nitrobenzene-d5	0.400	0.328	18.0	93	0.00
9	Naphthalene	0.400	0.373	6.8	93	0.00
10	Hexachlorobutadiene	0.400	0.375	6.3	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.373	6.8	93	0.00
12	2-Methylnaphthalene	0.400	0.378	5.5	94	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.400	0.318	20.5	92	0.00
15 S	2-Fluorobiphenyl	0.400	0.376	6.0	96	0.00
16	Acenaphthylene	0.400	0.352	12.0	92	0.00
17	Acenaphthene	0.400	0.371	7.3	93	0.00
18	Fluorene	0.400	0.374	6.5	95	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	95	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.343	14.2	90	0.00
21	4-Bromophenyl-phenylether	0.400	0.364	9.0	97	0.00
22	Hexachlorobenzene	0.400	0.380	5.0	97	0.00
23	Atrazine	0.400	0.312	22.0	95	0.00
24	Pentachlorophenol	0.400	0.315	21.3	90	0.00
25	Phenanthrene	0.400	0.375	6.3	97	0.00
26	Anthracene	0.400	0.354	11.5	96	0.00
27 SURR	Fluoranthene-d10	0.400	0.366	8.5	97	0.00
28	Fluoranthene	0.400	0.375	6.3	97	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	98	0.00
30	Pyrene	0.400	0.405	-1.3	97	0.00
31 S	Terphenyl-d14	0.400	0.403	-0.8	98	0.00
32	Benzo(a)anthracene	0.400	0.385	3.8	98	0.00
33	Chrysene	0.400	0.413	-3.2	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.339	15.3	91	0.00
35 I	Perylene-d12	0.400	0.400	0.0	107	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.402	-0.5	111	0.00
37	Benzo(b)fluoranthene	0.400	0.381	4.8	101	0.00
38	Benzo(k)fluoranthene	0.400	0.400	0.0	106	0.00
39 C	Benzo(a)pyrene	0.400	0.388	3.0	103	0.00
40	Dibenzo(a,h)anthracene	0.400	0.400	0.0	111	0.00
41	Benzo(g,h,i)perylene	0.400	0.395	1.3	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN100121