

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
 Data File : BN016614.D
 Acq On : 30 Sep 2021 23:58
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 01 01:37:47 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 30 17:56:53 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	105	0.00
2	1,4-Dioxane	0.400	0.390	2.5	110	0.00
3	n-Nitrosodimethylamine	0.400	0.363	9.3	103	0.00
4 S	2-Fluorophenol	0.400	0.367	8.3	103	0.00
5 S	Phenol-d6	0.400	0.357	10.8	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.379	5.3	105	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	103	0.00
8 S	Nitrobenzene-d5	0.400	0.326	18.5	103	0.00
9	Naphthalene	0.400	0.371	7.3	102	0.00
10	Hexachlorobutadiene	0.400	0.376	6.0	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.370	7.5	103	0.00
12	2-Methylnaphthalene	0.400	0.376	6.0	103	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.400	0.321	19.8	102	0.00
15 S	2-Fluorobiphenyl	0.400	0.371	7.3	104	0.00
16	Acenaphthylene	0.400	0.352	12.0	101	0.00
17	Acenaphthene	0.400	0.365	8.8	101	0.00
18	Fluorene	0.400	0.366	8.5	102	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	101	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.327	18.3	88	0.00
21	4-Bromophenyl-phenylether	0.400	0.366	8.5	103	0.00
22	Hexachlorobenzene	0.400	0.388	3.0	106	0.00
23	Atrazine	0.400	0.315	21.3	102	0.00
24	Pentachlorophenol	0.400	0.357	10.8	109	0.00
25	Phenanthrene	0.400	0.377	5.8	103	0.00
26	Anthracene	0.400	0.352	12.0	101	0.00
27 SURR	Fluoranthene-d10	0.400	0.363	9.3	102	0.00
28	Fluoranthene	0.400	0.375	6.3	103	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	105	-0.01
30	Pyrene	0.400	0.405	-1.3	104	0.00
31 S	Terphenyl-d14	0.400	0.401	-0.3	104	0.00
32	Benzo(a)anthracene	0.400	0.368	8.0	100	0.00
33	Chrysene	0.400	0.422	-5.5	109	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.325	18.8	93	-0.01
35 I	Perylene-d12	0.400	0.400	0.0	115	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.405	-1.3	120	0.00
37	Benzo(b)fluoranthene	0.400	0.384	4.0	108	0.00
38	Benzo(k)fluoranthene	0.400	0.391	2.3	111	0.00
39 C	Benzo(a)pyrene	0.400	0.382	4.5	109	0.00
40	Dibenzo(a,h)anthracene	0.400	0.402	-0.5	119	0.00
41	Benzo(g,h,i)perylene	0.400	0.401	-0.3	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4