

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
 Data File : BN016616.D
 Acq On : 01 Oct 2021 02:23
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 01 06:23:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 30 17:56:53 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.184	0.172	6.5	109	0.00
3	n-Nitrosodimethylamine	0.295	0.266	9.8	106	0.00
4 S	2-Fluorophenol	1.019	0.892	12.5	102	0.00
5 S	Phenol-d6	1.115	0.958	14.1	102	0.00
6	bis(2-Chloroethyl)ether	1.085	1.028	5.3	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.261	0.209	19.9	103	0.00
9	Naphthalene	1.087	1.010	7.1	105	0.00
10	Hexachlorobutadiene	0.204	0.194	4.9	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.592	0.549	7.3	106	0.00
12	2-Methylnaphthalene	0.694	0.657	5.3	106	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.182	0.141	22.5	103	0.00
15 S	2-Fluorobiphenyl	1.620	1.495	7.7	108	0.00
16	Acenaphthylene	1.749	1.507	13.8	103	0.00
17	Acenaphthene	1.178	1.072	9.0	105	0.00
18	Fluorene	1.393	1.287	7.6	108	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	108	-0.01
20	4,6-Dinitro-2-methylphenol	0.046	0.024	47.8#	91	0.00
21	4-Bromophenyl-phenylether	0.249	0.225	9.6	109	0.00
22	Hexachlorobenzene	0.288	0.276	4.2	111	-0.01
23	Atrazine	0.181	0.136	24.9	104	0.00
24	Pentachlorophenol	0.095	0.078	17.9	107	0.00
25	Phenanthrene	1.204	1.138	5.5	110	0.00
26	Anthracene	1.039	0.909	12.5	107	0.00
27 SURR	Fluoranthene-d10	1.064	0.955	10.2	108	0.00
28	Fluoranthene	1.299	1.218	6.2	110	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	112	-0.01
30	Pyrene	1.294	1.288	0.5	109	0.00
31 S	Terphenyl-d14	0.870	0.867	0.3	111	0.00
32	Benzo(a)anthracene	1.228	1.112	9.4	105	0.00
33	Chrysene	1.243	1.314	-5.7	116	0.00
34	Bis(2-ethylhexyl)phthalate	0.489	0.353	27.8#	88	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	125	0.00
36	Indeno(1,2,3-cd)pyrene	1.427	1.445	-1.3	130	0.00
37	Benzo(b)fluoranthene	1.285	1.176	8.5	113	0.00
38	Benzo(k)fluoranthene	1.270	1.285	-1.2	125	0.00
39 C	Benzo(a)pyrene	1.145	1.071	6.5	116	0.00
40	Dibenzo(a,h)anthracene	1.151	1.159	-0.7	130	-0.01
41	Benzo(g,h,i)perylene	1.262	1.248	1.1	123	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4