

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
 Data File : BN016633.D
 Acq On : 01 Oct 2021 12:51
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 01 13:35:34 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 01 13:35:26 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	97	0.00
2	1,4-Dioxane	0.400	0.374	6.5	98	0.00
3	n-Nitrosodimethylamine	0.400	0.367	8.3	96	0.00
4 S	2-Fluorophenol	0.400	0.361	9.8	94	0.00
5 S	Phenol-d6	0.400	0.352	12.0	94	0.00
6	bis(2-Chloroethyl)ether	0.400	0.377	5.8	97	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.330	17.5	95	0.00
9	Naphthalene	0.400	0.369	7.8	93	0.00
10	Hexachlorobutadiene	0.400	0.381	4.8	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.370	7.5	94	0.00
12	2-Methylnaphthalene	0.400	0.375	6.3	94	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.400	0.324	19.0	94	0.00
15 S	2-Fluorobiphenyl	0.400	0.375	6.3	96	0.00
16	Acenaphthylene	0.400	0.351	12.3	92	0.00
17	Acenaphthene	0.400	0.372	7.0	94	0.00
18	Fluorene	0.400	0.373	6.8	95	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	96	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.282	29.5#	64	0.00
21	4-Bromophenyl-phenylether	0.400	0.359	10.3	97	0.00
22	Hexachlorobenzene	0.400	0.379	5.3	98	0.00
23	Atrazine	0.400	0.309	22.8	95	0.00
24	Pentachlorophenol	0.400	0.343	14.2	100	0.00
25	Phenanthrene	0.400	0.374	6.5	98	0.00
26	Anthracene	0.400	0.353	11.8	97	0.00
27 SURR	Fluoranthene-d10	0.400	0.364	9.0	97	0.00
28	Fluoranthene	0.400	0.376	6.0	98	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	101	0.00
30	Pyrene	0.400	0.397	0.8	98	0.00
31 S	Terphenyl-d14	0.400	0.397	0.8	100	0.00
32	Benzo(a)anthracene	0.400	0.371	7.3	97	0.00
33	Chrysene	0.400	0.410	-2.5	102	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.306	23.5	85	0.00
35 I	Perylene-d12	0.400	0.400	0.0	108	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.393	1.8	109	0.00
37	Benzo(b)fluoranthene	0.400	0.375	6.3	100	0.00
38	Benzo(k)fluoranthene	0.400	0.397	0.8	106	0.00
39 C	Benzo(a)pyrene	0.400	0.379	5.3	102	0.00
40	Dibenzo(a,h)anthracene	0.400	0.387	3.3	108	0.00
41	Benzo(g,h,i)perylene	0.400	0.396	1.0	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4