



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
 Data File : BN016597.D
 Acq On : 30 Sep 2021 13:05
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN100121.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Sep 30 17:56:53 2021






AutoFind: Scans 2448, 2449, 2450; Background Corrected with Scan 2441

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.6	75461	PASS
68	69	0.00	2	1.8	1466	PASS
69	198	0.00	100	33.9	80935	PASS
70	69	0.00	2	0.8	680	PASS
127	198	10	80	46.7	111528	PASS
197	198	0.00	2	0.5	1080	PASS
198	198	100	100	100.0	238656	PASS
199	198	5	9	7.0	16695	PASS
275	198	10	60	28.0	66877	PASS
365	198	1	100	3.8	9034	PASS
441	198	0.01	100	18.7	44717	PASS
442	442	50	100	100.0	303147	PASS
443	442	15	24	19.0	57579	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
Data File : BN016597.D
Acq On : 30 Sep 2021 13:05
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Sep 30 18:14:23 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 30 18:14:09 2021
Response via : Initial Calibration



TIC: BN016597.D\data.ms

(70) Pentachlorophenol (C)

16.675min (0.000) 29259.20 ng




response 326518

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	63.32
264.00	61.60	62.77
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN100121\
Data File : BN016597.D
Acq On : 30 Sep 2021 13:05
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Sep 30 18:14:23 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 30 18:14:09 2021
Response via : Initial Calibration



TIC: BN016597.D\data.ms

(77) Benzidine

19.292min (0.000) 0.00 ng

response 1488118

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.44
183.00	13.20	12.53
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/1/2021	BNA_N	<u>BN016597.D</u>
Compound Name	Response	Retention Time
DDT	814355	20.539
DDD	3175	20.092
DDE	0	19.592
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3175	817530	0.39